

### CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

Continuing Calib Date: 05/14/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 09:41 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|------------|-----------|-----------|-------|------------|
|                      |            |           | FROM      | TO    |            |
| Aroclor-1248-1 (1)   | 5.66       | 5.67      | 5.57      | 5.77  | 0.01       |
| Aroclor-1248-2 (2)   | 5.93       | 5.94      | 5.84      | 6.04  | 0.01       |
| Aroclor-1248-3 (3)   | 6.13       | 6.14      | 6.04      | 6.24  | 0.01       |
| Aroclor-1248-4 (4)   | 6.53       | 6.54      | 6.44      | 6.64  | 0.01       |
| Aroclor-1248-5 (5)   | 6.57       | 6.58      | 6.48      | 6.68  | 0.01       |
| Tetrachloro-m-xylene | 4.50       | 4.51      | 4.41      | 4.61  | 0.01       |
| Decachlorobiphenyl   | 10.21      | 10.24     | 10.14     | 10.34 | 0.03       |

### CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

Continuing Calib Date: 05/14/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 09:41 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1248-1       | (1) | 4.88       | 4.90      | 4.80      | 5.00 | 0.02       |
| Aroclor-1248-2       | (2) | 5.12       | 5.14      | 5.04      | 5.24 | 0.02       |
| Aroclor-1248-3       | (3) | 5.16       | 5.18      | 5.08      | 5.28 | 0.02       |
| Aroclor-1248-4       | (4) | 5.34       | 5.35      | 5.25      | 5.45 | 0.01       |
| Aroclor-1248-5       | (5) | 5.73       | 5.74      | 5.64      | 5.84 | 0.01       |
| Tetrachloro-m-xylene |     | 3.80       | 3.81      | 3.71      | 3.91 | 0.01       |
| Decachlorobiphenyl   |     | 8.83       | 8.85      | 8.75      | 8.95 | 0.03       |