

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/14/2025 05/14/2025

Client Sample No.: CCAL01 Date Analyzed: 05/16/2025

Lab Sample No.: AR1660CCC500 Data File : PO111128.D Time Analyzed: 09:35

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.772	4.673	4.873	527.580	500.000	5.5
Aroclor-1016-2	4.791	4.692	4.892	535.920	500.000	7.2
Aroclor-1016-3	4.848	4.749	4.949	538.300	500.000	7.7
Aroclor-1016-4	4.968	4.869	5.069	535.840	500.000	7.2
Aroclor-1016-5	5.225	5.126	5.326	537.020	500.000	7.4
Aroclor-1260-1	6.266	6.166	6.366	532.310	500.000	6.5
Aroclor-1260-2	6.455	6.355	6.555	523.540	500.000	4.7
Aroclor-1260-3	6.822	6.722	6.922	524.950	500.000	5.0
Aroclor-1260-4	7.082	6.983	7.183	523.740	500.000	4.7
Aroclor-1260-5	7.324	7.225	7.425	525.040	500.000	5.0
Decachlorobiphenyl	8.723	8.624	8.824	51.510	50.000	3.0
Tetrachloro-m-xylene	3.682	3.582	3.782	54.040	50.000	8.1

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/14/2025 05/14/2025

Client Sample No.: CCAL01 Date Analyzed: 05/16/2025

Lab Sample No.: AR1660CCC500 Data File : PO111128.D Time Analyzed: 09:35

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.758	4.658	4.858	511.230	500.000	2.2
Aroclor-1016-2	4.777	4.677	4.877	513.940	500.000	2.8
Aroclor-1016-3	4.952	4.852	5.052	519.560	500.000	3.9
Aroclor-1016-4	4.994	4.894	5.094	516.210	500.000	3.2
Aroclor-1016-5	5.207	5.107	5.307	515.440	500.000	3.1
Aroclor-1260-1	6.237	6.138	6.338	514.070	500.000	2.8
Aroclor-1260-2	6.426	6.326	6.526	506.270	500.000	1.3
Aroclor-1260-3	6.578	6.479	6.679	485.210	500.000	-3.0
Aroclor-1260-4	7.049	6.949	7.149	502.830	500.000	0.6
Aroclor-1260-5	7.290	7.191	7.391	495.550	500.000	-0.9
Decachlorobiphenyl	8.674	8.574	8.774	49.550	50.000	-0.9
Tetrachloro-m-xylene	3.678	3.579	3.779	52.860	50.000	5.7