

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/14/2025 05/14/2025

Client Sample No.: CCAL02 Date Analyzed: 05/16/2025

Lab Sample No.: AR1660CCC500 Data File : PO111143.D Time Analyzed: 16:37

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.770	4.673	4.873	534.030	500.000	6.8
Aroclor-1016-2	4.789	4.692	4.892	534.550	500.000	6.9
Aroclor-1016-3	4.846	4.749	4.949	545.280	500.000	9.1
Aroclor-1016-4	4.967	4.869	5.069	540.340	500.000	8.1
Aroclor-1016-5	5.224	5.126	5.326	590.050	500.000	18.0
Aroclor-1260-1	6.264	6.166	6.366	533.270	500.000	6.7
Aroclor-1260-2	6.453	6.355	6.555	530.990	500.000	6.2
Aroclor-1260-3	6.820	6.722	6.922	526.430	500.000	5.3
Aroclor-1260-4	7.081	6.983	7.183	525.430	500.000	5.1
Aroclor-1260-5	7.323	7.225	7.425	522.400	500.000	4.5
Decachlorobiphenyl	8.723	8.624	8.824	51.170	50.000	2.3
Tetrachloro-m-xylene	3.680	3.582	3.782	54.280	50.000	8.6

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/14/2025 05/14/2025

Client Sample No.: CCAL02 Date Analyzed: 05/16/2025

Lab Sample No.: AR1660CCC500 Data File : PO111143.D Time Analyzed: 16:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.758	4.658	4.858	513.460	500.000	2.7
Aroclor-1016-2	4.777	4.677	4.877	519.720	500.000	3.9
Aroclor-1016-3	4.952	4.852	5.052	523.560	500.000	4.7
Aroclor-1016-4	4.995	4.894	5.094	496.360	500.000	-0.7
Aroclor-1016-5	5.207	5.107	5.307	552.900	500.000	10.6
Aroclor-1260-1	6.237	6.138	6.338	518.110	500.000	3.6
Aroclor-1260-2	6.425	6.326	6.526	512.700	500.000	2.5
Aroclor-1260-3	6.578	6.479	6.679	489.250	500.000	-2.2
Aroclor-1260-4	7.048	6.949	7.149	508.400	500.000	1.7
Aroclor-1260-5	7.290	7.191	7.391	503.670	500.000	0.7
Decachlorobiphenyl	8.674	8.574	8.774	50.240	50.000	0.5
Tetrachloro-m-xylene	3.679	3.579	3.779	53.300	50.000	6.6