

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/14/2025 05/14/2025

Client Sample No.: CCAL03 Date Analyzed: 05/16/2025

Lab Sample No.: AR1660CCC500 Data File : PO111157.D Time Analyzed: 21:44

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.774	4.673	4.873	530.930	500.000	6.2
Aroclor-1016-2	4.792	4.692	4.892	531.560	500.000	6.3
Aroclor-1016-3	4.849	4.749	4.949	541.180	500.000	8.2
Aroclor-1016-4	4.969	4.869	5.069	536.130	500.000	7.2
Aroclor-1016-5	5.226	5.126	5.326	542.010	500.000	8.4
Aroclor-1260-1	6.266	6.166	6.366	518.080	500.000	3.6
Aroclor-1260-2	6.456	6.355	6.555	512.350	500.000	2.5
Aroclor-1260-3	6.823	6.722	6.922	498.480	500.000	-0.3
Aroclor-1260-4	7.082	6.983	7.183	512.240	500.000	2.4
Aroclor-1260-5	7.325	7.225	7.425	508.410	500.000	1.7
Decachlorobiphenyl	8.725	8.624	8.824	48.060	50.000	-3.9
Tetrachloro-m-xylene	3.683	3.582	3.782	53.720	50.000	7.4

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/14/2025 05/14/2025

Client Sample No.: CCAL03 Date Analyzed: 05/16/2025

Lab Sample No.: AR1660CCC500 Data File : PO111157.D Time Analyzed: 21:44

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.759	4.658	4.858	510.060	500.000	2.0
Aroclor-1016-2	4.778	4.677	4.877	513.910	500.000	2.8
Aroclor-1016-3	4.953	4.852	5.052	519.360	500.000	3.9
Aroclor-1016-4	4.995	4.894	5.094	483.150	500.000	-3.4
Aroclor-1016-5	5.208	5.107	5.307	514.660	500.000	2.9
Aroclor-1260-1	6.239	6.138	6.338	509.330	500.000	1.9
Aroclor-1260-2	6.426	6.326	6.526	503.580	500.000	0.7
Aroclor-1260-3	6.579	6.479	6.679	478.580	500.000	-4.3
Aroclor-1260-4	7.050	6.949	7.149	496.550	500.000	-0.7
Aroclor-1260-5	7.291	7.191	7.391	491.760	500.000	-1.6
Decachlorobiphenyl	8.674	8.574	8.774	48.780	50.000	-2.4
Tetrachloro-m-xylene	3.679	3.579	3.779	52.840	50.000	5.7