

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL05 Date Analyzed: 05/15/2025

Lab Sample No.: AR1660CCC500 Data File : PP072091.D Time Analyzed: 09:15

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.656	5.571	5.771	450.250	500.000	-10.0
Aroclor-1016-2	5.678	5.593	5.793	483.560	500.000	-3.3
Aroclor-1016-3	5.741	5.654	5.854	493.590	500.000	-1.3
Aroclor-1016-4	5.839	5.752	5.952	490.260	500.000	-1.9
Aroclor-1016-5	6.130	6.045	6.245	456.190	500.000	-8.8
Aroclor-1260-1	7.249	7.165	7.365	469.730	500.000	-6.1
Aroclor-1260-2	7.503	7.418	7.618	460.660	500.000	-7.9
Aroclor-1260-3	7.861	7.777	7.977	473.220	500.000	-5.4
Aroclor-1260-4	8.086	8.002	8.202	445.780	500.000	-10.8
Aroclor-1260-5	8.404	8.322	8.522	501.640	500.000	0.3
Decachlorobiphenyl	10.213	10.137	10.337	51.770	50.000	3.5
Tetrachloro-m-xylene	4.504	4.417	4.617	50.390	50.000	0.8

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL05 Date Analyzed: 05/15/2025

Lab Sample No.: AR1660CCC500 Data File : PP072091.D Time Analyzed: 09:15

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.884	4.798	4.998	555.850	500.000	11.2
Aroclor-1016-2	4.902	4.816	5.016	530.340	500.000	6.1
Aroclor-1016-3	5.079	4.993	5.193	490.540	500.000	-1.9
Aroclor-1016-4	5.121	5.036	5.236	494.300	500.000	-1.1
Aroclor-1016-5	5.335	5.250	5.450	532.270	500.000	6.5
Aroclor-1260-1	6.368	6.285	6.485	539.150	500.000	7.8
Aroclor-1260-2	6.556	6.473	6.673	518.460	500.000	3.7
Aroclor-1260-3	6.709	6.627	6.827	506.420	500.000	1.3
Aroclor-1260-4	7.179	7.098	7.298	525.790	500.000	5.2
Aroclor-1260-5	7.421	7.339	7.539	510.340	500.000	2.1
Decachlorobiphenyl	8.823	8.749	8.949	54.320	50.000	8.6
Tetrachloro-m-xylene	3.801	3.712	3.912	52.660	50.000	5.3