

### CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL06 Date Analyzed: 05/15/2025

Lab Sample No.: AR1660CCC500 Data File : PP072097.D Time Analyzed: 11:28

| COMPOUND             | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|--------|-----------|--------|--------------------|-------------------|------|
|                      |        | FROM      | TO     |                    |                   |      |
| Aroclor-1016-1       | 5.656  | 5.571     | 5.771  | 567.640            | 500.000           | 13.5 |
| Aroclor-1016-2       | 5.680  | 5.593     | 5.793  | 504.310            | 500.000           | 0.9  |
| Aroclor-1016-3       | 5.742  | 5.654     | 5.854  | 510.590            | 500.000           | 2.1  |
| Aroclor-1016-4       | 5.839  | 5.752     | 5.952  | 508.830            | 500.000           | 1.8  |
| Aroclor-1016-5       | 6.131  | 6.045     | 6.245  | 488.170            | 500.000           | -2.4 |
| Aroclor-1260-1       | 7.250  | 7.165     | 7.365  | 484.490            | 500.000           | -3.1 |
| Aroclor-1260-2       | 7.504  | 7.418     | 7.618  | 486.730            | 500.000           | -2.7 |
| Aroclor-1260-3       | 7.862  | 7.777     | 7.977  | 495.800            | 500.000           | -0.8 |
| Aroclor-1260-4       | 8.086  | 8.002     | 8.202  | 471.770            | 500.000           | -5.6 |
| Aroclor-1260-5       | 8.406  | 8.322     | 8.522  | 502.310            | 500.000           | 0.5  |
| Decachlorobiphenyl   | 10.215 | 10.137    | 10.337 | 52.810             | 50.000            | 5.6  |
| Tetrachloro-m-xylene | 4.505  | 4.417     | 4.617  | 51.030             | 50.000            | 2.1  |

### CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL06 Date Analyzed: 05/15/2025

Lab Sample No.: AR1660CCC500 Data File : PP072097.D Time Analyzed: 11:28

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|-------|--------------------|-------------------|------|
|                      |       | FROM      | TO    |                    |                   |      |
| Aroclor-1016-1       | 4.884 | 4.798     | 4.998 | 545.360            | 500.000           | 9.1  |
| Aroclor-1016-2       | 4.903 | 4.816     | 5.016 | 531.600            | 500.000           | 6.3  |
| Aroclor-1016-3       | 5.079 | 4.993     | 5.193 | 552.240            | 500.000           | 10.4 |
| Aroclor-1016-4       | 5.120 | 5.036     | 5.236 | 543.190            | 500.000           | 8.6  |
| Aroclor-1016-5       | 5.335 | 5.250     | 5.450 | 550.020            | 500.000           | 10.0 |
| Aroclor-1260-1       | 6.368 | 6.285     | 6.485 | 555.360            | 500.000           | 11.1 |
| Aroclor-1260-2       | 6.557 | 6.473     | 6.673 | 550.400            | 500.000           | 10.1 |
| Aroclor-1260-3       | 6.710 | 6.627     | 6.827 | 540.520            | 500.000           | 8.1  |
| Aroclor-1260-4       | 7.180 | 7.098     | 7.298 | 551.470            | 500.000           | 10.3 |
| Aroclor-1260-5       | 7.421 | 7.339     | 7.539 | 543.640            | 500.000           | 8.7  |
| Decachlorobiphenyl   | 8.826 | 8.749     | 8.949 | 57.150             | 50.000            | 14.3 |
| Tetrachloro-m-xylene | 3.801 | 3.712     | 3.912 | 57.660             | 50.000            | 15.3 |