

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL07 Date Analyzed: 05/15/2025

Lab Sample No.: AR1660CCC500 Data File : PP072112.D Time Analyzed: 16:42

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.659	5.571	5.771	455.210	500.000	-9.0
Aroclor-1016-2	5.681	5.593	5.793	480.350	500.000	-3.9
Aroclor-1016-3	5.744	5.654	5.854	490.340	500.000	-1.9
Aroclor-1016-4	5.841	5.752	5.952	489.550	500.000	-2.1
Aroclor-1016-5	6.134	6.045	6.245	476.750	500.000	-4.7
Aroclor-1260-1	7.252	7.165	7.365	477.340	500.000	-4.5
Aroclor-1260-2	7.506	7.418	7.618	462.870	500.000	-7.4
Aroclor-1260-3	7.864	7.777	7.977	470.710	500.000	-5.9
Aroclor-1260-4	8.089	8.002	8.202	442.830	500.000	-11.4
Aroclor-1260-5	8.408	8.322	8.522	481.350	500.000	-3.7
Decachlorobiphenyl	10.217	10.137	10.337	51.130	50.000	2.3
Tetrachloro-m-xylene	4.507	4.417	4.617	50.710	50.000	1.4

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL07 Date Analyzed: 05/15/2025

Lab Sample No.: AR1660CCC500 Data File : PP072112.D Time Analyzed: 16:42

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.884	4.798	4.998	543.950	500.000	8.8
Aroclor-1016-2	4.902	4.816	5.016	560.600	500.000	12.1
Aroclor-1016-3	5.080	4.993	5.193	509.430	500.000	1.9
Aroclor-1016-4	5.121	5.036	5.236	512.710	500.000	2.5
Aroclor-1016-5	5.335	5.250	5.450	552.120	500.000	10.4
Aroclor-1260-1	6.368	6.285	6.485	541.590	500.000	8.3
Aroclor-1260-2	6.556	6.473	6.673	536.920	500.000	7.4
Aroclor-1260-3	6.709	6.627	6.827	529.210	500.000	5.8
Aroclor-1260-4	7.179	7.098	7.298	534.550	500.000	6.9
Aroclor-1260-5	7.422	7.339	7.539	529.950	500.000	6.0
Decachlorobiphenyl	8.824	8.749	8.949	55.620	50.000	11.2
Tetrachloro-m-xylene	3.801	3.712	3.912	57.680	50.000	15.4