

Matrix:

Final Vol:

SOIL

10000

иL



Report of Analysis

Date Collected: Client: CDM Smith 05/13/25

Project: South River WM Replacement Date Received: 05/13/25

Client Sample ID: TP-29-99 SDG No.: Q2032

Lab Sample ID: Q2032-03 % Solid:

86.8 Decanted: Analytical Method: 8082A

PCB Soil Aliquot Vol: uL Test:

Extraction Type: Injection Volume:

PH: GPC Factor: 1.0

Units:

g

30.08

Prep Method: SW3541B

Sample Wt/Vol:

File ID/Qc Batch: Dilution: Prep Date Date Analyzed Prep Batch ID PO111137.D 05/16/25 09:24 05/16/25 13:57 PB168039

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.50	U	4.50	19.5	ug/kg
11104-28-2	Aroclor-1221	4.60	U	4.60	19.5	ug/kg
11141-16-5	Aroclor-1232	4.30	U	4.30	19.5	ug/kg
53469-21-9	Aroclor-1242	4.60	U	4.60	19.5	ug/kg
12672-29-6	Aroclor-1248	6.80	U	6.80	19.5	ug/kg
11097-69-1	Aroclor-1254	3.70	U	3.70	19.5	ug/kg
37324-23-5	Aroclor-1262	5.80	U	5.80	19.5	ug/kg
11100-14-4	Aroclor-1268	4.10	U	4.10	19.5	ug/kg
11096-82-5	Aroclor-1260	3.70	U	3.70	19.5	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	22.4		32 - 144	112%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.1		32 - 175	81%	SPK: 20

Comments:

U = Not Detected

LOO = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit