



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2032

Analytical Method: 8082A

Client: CDM Smith

Datafile : PP072082.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB168005BS (Column 1)	AR1016	5	4.60	ug/L	92				61	112	
	AR1260	5	4.30	ug/L	86				66	113	
PB168005BS (Column 2)	AR1016	5	4.90	ug/L	98				61	112	
	AR1260	5	4.90	ug/L	98				66	113	