



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2032</u>	Analytical Method:	<u>8082A</u>
Client:	<u>CDM Smith</u>	Datafile :	<u>PP072083.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB168005BSD (Column 1)	AR1016	5	4.50	ug/L	90	2			61	112	20
	AR1260	5	4.30	ug/L	86	0			66	113	20
PB168005BSD (Column 2)	AR1016	5	4.60	ug/L	92	2			61	112	20
	AR1260	5	4.80	ug/L	96	0			66	113	20