

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: CAMP02

Lab Code: CHEM **Case No.:** Q2032 **SAS No.:** Q2032 **SDG NO.:** Q2032

Instrument ID: ECD_L **Date(s) Analyzed:** 05/13/2025 05/13/2025

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	5.89	5.79	5.99	19457800
		2	6.28	6.18	6.38	44722300
		3	7.10	7.00	7.20	132062000
		4	7.19	7.09	7.29	92581300
		5	7.97	7.87	8.07	61074500

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: CAMP02

Lab Code: CHEM **Case No.:** Q2032 **SAS No.:** Q2032 **SDG NO.:** Q2032

Instrument ID: ECD_L **Date(s) Analyzed:** 05/13/2025 05/13/2025

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	5.17	5.07	5.27	41802700
		2	5.85	5.75	5.95	40240100
		3	6.13	6.03	6.23	44424800
		4	6.77	6.67	6.87	129582000
		5	7.21	7.11	7.31	95479000