

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/13/2025 05/13/2025

Client Sample No.: CCAL01 Date Analyzed: 05/14/2025

Lab Sample No.: PSTDCCC050 Data File : PL095667.D Time Analyzed: 16:26

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.746	6.648	6.848	51.350	50.000	2.7
4,4'-DDE	6.237	6.139	6.339	52.400	50.000	4.8
4,4'-DDT	7.061	6.964	7.164	49.610	50.000	-0.8
Aldrin	5.312	5.214	5.414	51.000	50.000	2.0
alpha-BHC	4.042	3.944	4.144	51.070	50.000	2.1
alpha-Chlordane	6.068	5.971	6.171	51.570	50.000	3.1
beta-BHC	4.559	4.460	4.660	49.590	50.000	-0.8
Decachlorobiphenyl	9.112	9.015	9.215	49.600	50.000	-0.8
delta-BHC	4.807	4.709	4.909	50.090	50.000	0.2
Dieldrin	6.389	6.291	6.491	51.500	50.000	3.0
Endosulfan I	6.116	6.019	6.219	51.520	50.000	3.0
Endosulfan II	6.828	6.731	6.931	50.330	50.000	0.7
Endosulfan sulfate	7.190	7.093	7.293	51.440	50.000	2.9
Endrin	6.617	6.519	6.719	48.320	50.000	-3.4
Endrin aldehyde	6.957	6.860	7.060	51.230	50.000	2.5
Endrin ketone	7.672	7.574	7.774	52.320	50.000	4.6
gamma-BHC (Lindane)	4.373	4.275	4.475	50.670	50.000	1.3
gamma-Chlordane	5.987	5.890	6.090	50.850	50.000	1.7
Heptachlor	4.970	4.872	5.072	49.670	50.000	-0.7
Heptachlor epoxide	5.733	5.635	5.835	49.410	50.000	-1.2
Methoxychlor	7.532	7.434	7.634	48.030	50.000	-3.9
Tetrachloro-m-xylene	3.592	3.493	3.693	49.560	50.000	-0.9

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/13/2025 05/13/2025

Client Sample No.: CCAL01 Date Analyzed: 05/14/2025

Lab Sample No.: PSTDCCC050 Data File : PL095667.D Time Analyzed: 16:26

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.943	5.849	6.049	47.700	50.000	-4.6
4,4'-DDE	5.390	5.298	5.498	46.350	50.000	-7.3
4,4'-DDT	6.197	6.103	6.303	44.090	50.000	-11.8
Aldrin	4.386	4.292	4.492	45.940	50.000	-8.1
alpha-BHC	3.415	3.320	3.520	46.660	50.000	-6.7
alpha-Chlordane	5.205	5.112	5.312	45.450	50.000	-9.1
beta-BHC	4.042	3.947	4.147	44.780	50.000	-10.4
Decachlorobiphenyl	8.080	7.987	8.187	46.030	50.000	-7.9
delta-BHC	4.277	4.183	4.383	46.210	50.000	-7.6
Dieldrin	5.526	5.433	5.633	46.570	50.000	-6.9
Endosulfan I	5.261	5.168	5.368	47.830	50.000	-4.3
Endosulfan II	6.091	5.998	6.198	45.590	50.000	-8.8
Endosulfan sulfate	6.493	6.399	6.599	45.770	50.000	-8.5
Endrin	5.801	5.708	5.908	43.950	50.000	-12.1
Endrin aldehyde	6.270	6.176	6.376	47.020	50.000	-6.0
Endrin ketone	6.999	6.905	7.105	47.430	50.000	-5.1
gamma-BHC (Lindane)	3.748	3.653	3.853	46.390	50.000	-7.2
gamma-Chlordane	5.141	5.048	5.248	46.000	50.000	-8.0
Heptachlor	4.101	4.007	4.207	44.550	50.000	-10.9
Heptachlor epoxide	4.889	4.795	4.995	46.010	50.000	-8.0
Methoxychlor	6.767	6.673	6.873	42.910	50.000	-14.2
Tetrachloro-m-xylene	2.905	2.810	3.010	45.950	50.000	-8.1