

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/13/2025 05/13/2025

Client Sample No.: CCAL03 Date Analyzed: 05/16/2025

Lab Sample No.: PSTDCCC050 Data File : PL095715.D Time Analyzed: 10:39

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.747	6.648	6.848	51.360	50.000	2.7
4,4'-DDE	6.237	6.139	6.339	52.530	50.000	5.1
4,4'-DDT	7.062	6.964	7.164	55.510	50.000	11.0
Aldrin	5.313	5.214	5.414	52.440	50.000	4.9
alpha-BHC	4.043	3.944	4.144	52.010	50.000	4.0
alpha-Chlordane	6.069	5.971	6.171	53.040	50.000	6.1
beta-BHC	4.559	4.460	4.660	52.170	50.000	4.3
Decachlorobiphenyl	9.115	9.015	9.215	53.580	50.000	7.2
delta-BHC	4.807	4.709	4.909	52.040	50.000	4.1
Dieldrin	6.390	6.291	6.491	51.780	50.000	3.6
Endosulfan I	6.117	6.019	6.219	51.930	50.000	3.9
Endosulfan II	6.830	6.731	6.931	51.430	50.000	2.9
Endosulfan sulfate	7.192	7.093	7.293	52.320	50.000	4.6
Endrin	6.618	6.519	6.719	54.680	50.000	9.4
Endrin aldehyde	6.958	6.860	7.060	50.860	50.000	1.7
Endrin ketone	7.673	7.574	7.774	51.470	50.000	2.9
gamma-BHC (Lindane)	4.374	4.275	4.475	52.200	50.000	4.4
gamma-Chlordane	5.988	5.890	6.090	52.150	50.000	4.3
Heptachlor	4.971	4.872	5.072	51.580	50.000	3.2
Heptachlor epoxide	5.732	5.635	5.835	52.120	50.000	4.2
Methoxychlor	7.534	7.434	7.634	54.980	50.000	10.0
Tetrachloro-m-xylene	3.592	3.493	3.693	51.410	50.000	2.8

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/13/2025 05/13/2025

Client Sample No.: CCAL03 Date Analyzed: 05/16/2025

Lab Sample No.: PSTDCCC050 Data File : PL095715.D Time Analyzed: 10:39

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.948	5.849	6.049	51.080	50.000	2.2
4,4'-DDE	5.394	5.298	5.498	51.050	50.000	2.1
4,4'-DDT	6.202	6.103	6.303	53.650	50.000	7.3
Aldrin	4.390	4.292	4.492	49.770	50.000	-0.5
alpha-BHC	3.418	3.320	3.520	50.020	50.000	0.0
alpha-Chlordane	5.209	5.112	5.312	49.970	50.000	-0.1
beta-BHC	4.045	3.947	4.147	48.720	50.000	-2.6
Decachlorobiphenyl	8.085	7.987	8.187	49.390	50.000	-1.2
delta-BHC	4.282	4.183	4.383	50.450	50.000	0.9
Dieldrin	5.531	5.433	5.633	52.050	50.000	4.1
Endosulfan I	5.265	5.168	5.368	49.140	50.000	-1.7
Endosulfan II	6.096	5.998	6.198	49.830	50.000	-0.3
Endosulfan sulfate	6.497	6.399	6.599	49.710	50.000	-0.6
Endrin	5.806	5.708	5.908	53.610	50.000	7.2
Endrin aldehyde	6.274	6.176	6.376	50.760	50.000	1.5
Endrin ketone	7.004	6.905	7.105	48.780	50.000	-2.4
gamma-BHC (Lindane)	3.752	3.653	3.853	50.030	50.000	0.1
gamma-Chlordane	5.145	5.048	5.248	50.790	50.000	1.6
Heptachlor	4.105	4.007	4.207	49.890	50.000	-0.2
Heptachlor epoxide	4.892	4.795	4.995	49.850	50.000	-0.3
Methoxychlor	6.771	6.673	6.873	52.580	50.000	5.2
Tetrachloro-m-xylene	2.909	2.810	3.010	48.500	50.000	-3.0