

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/13/2025 05/13/2025

Client Sample No.: CCAL04 Date Analyzed: 05/16/2025

Lab Sample No.: PSTDCCC050 Data File : PL095730.D Time Analyzed: 15:55

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.746	6.648	6.848	51.240	50.000	2.5
4,4'-DDE	6.236	6.139	6.339	52.790	50.000	5.6
4,4'-DDT	7.061	6.964	7.164	54.780	50.000	9.6
Aldrin	5.312	5.214	5.414	52.310	50.000	4.6
alpha-BHC	4.042	3.944	4.144	52.100	50.000	4.2
alpha-Chlordane	6.069	5.971	6.171	53.430	50.000	6.9
beta-BHC	4.558	4.460	4.660	52.410	50.000	4.8
Decachlorobiphenyl	9.115	9.015	9.215	53.160	50.000	6.3
delta-BHC	4.807	4.709	4.909	52.300	50.000	4.6
Dieldrin	6.389	6.291	6.491	52.330	50.000	4.7
Endosulfan I	6.116	6.019	6.219	52.570	50.000	5.1
Endosulfan II	6.829	6.731	6.931	51.230	50.000	2.5
Endosulfan sulfate	7.191	7.093	7.293	52.290	50.000	4.6
Endrin	6.617	6.519	6.719	54.380	50.000	8.8
Endrin aldehyde	6.958	6.860	7.060	51.050	50.000	2.1
Endrin ketone	7.673	7.574	7.774	51.220	50.000	2.4
gamma-BHC (Lindane)	4.372	4.275	4.475	52.470	50.000	4.9
gamma-Chlordane	5.988	5.890	6.090	52.460	50.000	4.9
Heptachlor	4.970	4.872	5.072	51.340	50.000	2.7
Heptachlor epoxide	5.732	5.635	5.835	51.030	50.000	2.1
Methoxychlor	7.533	7.434	7.634	54.180	50.000	8.4
Tetrachloro-m-xylene	3.591	3.493	3.693	51.320	50.000	2.6

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/13/2025 05/13/2025

Client Sample No.: CCAL04 Date Analyzed: 05/16/2025

Lab Sample No.: PSTDCCC050 Data File : PL095730.D Time Analyzed: 15:55

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.947	5.849	6.049	52.070	50.000	4.1
4,4'-DDE	5.393	5.298	5.498	51.350	50.000	2.7
4,4'-DDT	6.201	6.103	6.303	53.770	50.000	7.5
Aldrin	4.389	4.292	4.492	49.970	50.000	-0.1
alpha-BHC	3.417	3.320	3.520	50.260	50.000	0.5
alpha-Chlordane	5.209	5.112	5.312	50.630	50.000	1.3
beta-BHC	4.044	3.947	4.147	49.460	50.000	-1.1
Decachlorobiphenyl	8.085	7.987	8.187	49.700	50.000	-0.6
delta-BHC	4.281	4.183	4.383	51.030	50.000	2.1
Dieldrin	5.530	5.433	5.633	52.390	50.000	4.8
Endosulfan I	5.264	5.168	5.368	49.650	50.000	-0.7
Endosulfan II	6.095	5.998	6.198	50.640	50.000	1.3
Endosulfan sulfate	6.497	6.399	6.599	50.510	50.000	1.0
Endrin	5.805	5.708	5.908	54.200	50.000	8.4
Endrin aldehyde	6.273	6.176	6.376	51.960	50.000	3.9
Endrin ketone	7.003	6.905	7.105	49.710	50.000	-0.6
gamma-BHC (Lindane)	3.751	3.653	3.853	50.330	50.000	0.7
gamma-Chlordane	5.144	5.048	5.248	51.370	50.000	2.7
Heptachlor	4.104	4.007	4.207	49.810	50.000	-0.4
Heptachlor epoxide	4.891	4.795	4.995	50.740	50.000	1.5
Methoxychlor	6.771	6.673	6.873	52.690	50.000	5.4
Tetrachloro-m-xylene	2.908	2.810	3.010	49.040	50.000	-1.9