

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL051325\
 Data File : PL095633.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 May 2025 12:24
 Operator : AR\AJ
 Sample : PSTDICC100
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC100

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 13 13:47:09 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL051325.M
 Quant Title : GC Extractables
 QLast Update : Tue May 13 13:44:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.594	2.912	387.4E6	425.8E6	96.667	95.333
28)	SA Decachlor...	9.117	8.088	326.3E6	383.9E6	97.366	96.824
Target Compounds							
2)	A alpha-BHC	4.044	3.421	588.0E6	672.5E6	98.504	96.200
3)	MA gamma-BHC...	4.375	3.755	551.3E6	634.9E6	98.230	96.543
4)	MA Heptachlor	4.973	4.108	530.7E6	599.1E6	97.765	95.775
5)	MB Aldrin	5.315	4.393	565.7E6	587.7E6	98.257	95.904
6)	B beta-BHC	4.561	4.048	231.7E6	271.3E6	96.742	95.772
7)	B delta-BHC	4.809	4.284	537.0E6	643.8E6	98.546	96.878
8)	B Heptachlo...	5.735	4.895	481.1E6	526.9E6	97.508	95.713
9)	A Endosulfan I	6.119	5.268	476.9E6	503.7E6	97.930	96.980
10)	B gamma-Chl...	5.990	5.148	521.3E6	552.8E6	98.409	97.275
11)	B alpha-Chl...	6.071	5.213	511.5E6	539.0E6	98.219	96.911
12)	B 4,4'-DDE	6.239	5.398	511.5E6	554.1E6	98.471	97.424
13)	MA Dieldrin	6.392	5.534	484.8E6	568.5E6	98.598	96.412
14)	MA Endrin	6.619	5.808	379.2E6	497.4E6	98.352	96.734
15)	B Endosulfa...	6.831	6.099	399.8E6	496.1E6	97.690	96.347
16)	A 4,4'-DDD	6.748	5.951	372.4E6	478.8E6	98.397	96.842
17)	MA 4,4'-DDT	7.064	6.204	373.1E6	492.4E6	99.912	98.672
18)	B Endrin al...	6.959	6.277	289.2E6	361.6E6	98.248	98.122
19)	B Endosulfa...	7.194	6.500	354.7E6	449.1E6	97.938	95.857
20)	A Methoxychlor	7.535	6.774	169.4E6	254.9E6	99.071	97.415
21)	B Endrin ke...	7.675	7.006	368.6E6	547.2E6	98.276	96.836
22)	Mirex	8.156	7.201	285.2E6	392.1E6	96.576	95.765

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL051325\
Data File : PL095633.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 May 2025 12:24
Operator : AR\AJ
Sample : PSTDICC100
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC100

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 13 13:47:09 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL051325.M
Quant Title : GC Extractables
QLast Update : Tue May 13 13:44:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

