

Surrogate Summary

SDG No.: Q2032

Client: CDM Smith

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL095630.D	PIBLK-PL095630.D	Decachlorobiphenyl	1	20	19.1	95		43	140
		Tetrachloro-m-xylene	1	20	15.9	80		77	126
		Decachlorobiphenyl	2	20	18.4	92		43	140
		Tetrachloro-m-xylene	2	20	16.2	81		77	126
I.BLK-PL095666.D	PIBLK-PL095666.D	Decachlorobiphenyl	1	20	20.4	102		43	140
		Tetrachloro-m-xylene	1	20	18.4	92		77	126
		Decachlorobiphenyl	2	20	17.6	88		43	140
		Tetrachloro-m-xylene	2	20	16.3	81		77	126
PB168004BL	PB168004BL	Decachlorobiphenyl	1	20	20.3	101		43	140
		Tetrachloro-m-xylene	1	20	19.5	98		77	126
		Decachlorobiphenyl	2	20	19.3	97		43	140
		Tetrachloro-m-xylene	2	20	18.6	93		77	126
PB168004BS	PB168004BS	Decachlorobiphenyl	1	20	21.1	106		43	140
		Tetrachloro-m-xylene	1	20	20.1	100		77	126
		Decachlorobiphenyl	2	20	20.2	101		43	140
		Tetrachloro-m-xylene	2	20	18.9	95		77	126
PB168004BSD	PB168004BSD	Decachlorobiphenyl	1	20	20.2	101		43	140
		Tetrachloro-m-xylene	1	20	18.9	95		77	126
		Decachlorobiphenyl	2	20	19.6	98		43	140
		Tetrachloro-m-xylene	2	20	18.1	90		77	126
Q2032-11	FB-05132025	Decachlorobiphenyl	1	20	9.36	47		43	140
		Tetrachloro-m-xylene	1	20	20.1	101		77	126
		Decachlorobiphenyl	2	20	8.58	43		43	140
		Tetrachloro-m-xylene	2	20	19.2	96		77	126
I.BLK-PL095682.D	PIBLK-PL095682.D	Decachlorobiphenyl	1	20	21.2	106		43	140
		Tetrachloro-m-xylene	1	20	20.3	102		77	126
		Decachlorobiphenyl	2	20	19.5	97		43	140
		Tetrachloro-m-xylene	2	20	18.8	94		77	126
I.BLK-PL095713.D	PIBLK-PL095713.D	Decachlorobiphenyl	1	20	21.6	108		43	140
		Tetrachloro-m-xylene	1	20	19.0	95		77	126
		Decachlorobiphenyl	2	20	19.4	97		43	140
		Tetrachloro-m-xylene	2	20	17.0	85		77	126
PB168040BL	PB168040BL	Decachlorobiphenyl	1	20	19.3	97		20	144
		Tetrachloro-m-xylene	1	20	16.8	84		19	148
		Decachlorobiphenyl	2	20	17.3	86		20	144
		Tetrachloro-m-xylene	2	20	16.0	80		19	148
PB168040BS	PB168040BS	Decachlorobiphenyl	1	20	19.1	96		20	144
		Tetrachloro-m-xylene	1	20	17.4	87		19	148
		Decachlorobiphenyl	2	20	17.6	88		20	144
		Tetrachloro-m-xylene	2	20	16.9	84		19	148
Q2032-01	TP-11	Decachlorobiphenyl	1	20	11.3	56		20	144

Surrogate Summary

SDG No.: Q2032

Client: CDM Smith

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
Q2032-01	TP-11	Tetrachloro-m-xylene	1	20	13.8	69		19	148
		Decachlorobiphenyl	2	20	9.99	50		20	144
Q2032-02	TP-29	Tetrachloro-m-xylene	2	20	13.8	69		19	148
		Decachlorobiphenyl	1	20	13.2	66		20	144
		Tetrachloro-m-xylene	1	20	17.9	90		19	148
		Decachlorobiphenyl	2	20	12.9	65		20	144
		Tetrachloro-m-xylene	2	20	17.6	88		19	148
Q2032-03	TP-29-99	Decachlorobiphenyl	1	20	15.5	77		20	144
		Tetrachloro-m-xylene	1	20	18.6	93		19	148
		Decachlorobiphenyl	2	20	14.2	71		20	144
		Tetrachloro-m-xylene	2	20	18.0	90		19	148
Q2032-04	TP-24	Decachlorobiphenyl	1	20	14.5	73		20	144
		Tetrachloro-m-xylene	1	20	17.2	86		19	148
		Decachlorobiphenyl	2	20	14.1	71		20	144
		Tetrachloro-m-xylene	2	20	17.4	87		19	148
Q2032-05MS	TP-24MS	Decachlorobiphenyl	1	20	14.9	74		20	144
		Tetrachloro-m-xylene	1	20	17.4	87		19	148
		Decachlorobiphenyl	2	20	14.0	70		20	144
		Tetrachloro-m-xylene	2	20	17.5	88		19	148
Q2032-06MSD	TP-24MSD	Decachlorobiphenyl	1	20	14.9	74		20	144
		Tetrachloro-m-xylene	1	20	17.0	85		19	148
		Decachlorobiphenyl	2	20	13.9	70		20	144
		Tetrachloro-m-xylene	2	20	17.5	87		19	148
		Decachlorobiphenyl	1	20	11.1	56		20	144
Q2032-07	TP-37	Tetrachloro-m-xylene	1	20	15.0	75		19	148
		Decachlorobiphenyl	2	20	11.6	58		20	144
		Tetrachloro-m-xylene	2	20	15.0	75		19	148
		Decachlorobiphenyl	1	20	13.6	68		20	144
Q2032-08	TP-32	Tetrachloro-m-xylene	1	20	18.4	92		19	148
		Decachlorobiphenyl	2	20	14.2	71		20	144
		Tetrachloro-m-xylene	2	20	18.3	91		19	148
		Decachlorobiphenyl	1	20	21.4	107		43	140
I.BLK-PL095729.D	PIBLK-PL095729.D	Tetrachloro-m-xylene	1	20	19.1	95		77	126
		Decachlorobiphenyl	2	20	20.7	103		43	140
		Tetrachloro-m-xylene	2	20	18.5	93		77	126