

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2032

Analytical Method: 8081B

Client: CDM Smith

Datafile : PL095669.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB168004BS (Column 1)	alpha-BHC	0.5	0.51	ug/L	103				85	130	
	beta-BHC	0.5	0.50	ug/L	100				83	126	
	delta-BHC	0.5	0.51	ug/L	102				69	141	
	gamma-BHC (Lindane)	0.5	0.51	ug/L	102				82	129	
	Heptachlor	0.5	0.50	ug/L	101				79	127	
	Aldrin	0.5	0.52	ug/L	104				79	126	
	Heptachlor epoxide	0.5	0.51	ug/L	101				81	124	
	Endosulfan I	0.5	0.53	ug/L	106				85	122	
	Dieldrin	0.5	0.53	ug/L	106				83	125	
	4,4'-DDE	0.5	0.54	ug/L	108				80	127	
	Endrin	0.5	0.50	ug/L	100				81	128	
	Endosulfan II	0.5	0.51	ug/L	103				82	123	
	4,4'-DDD	0.5	0.53	ug/L	105				77	131	
	Endosulfan sulfate	0.5	0.53	ug/L	105				76	129	
	4,4'-DDT	0.5	0.51	ug/L	101				80	133	
	Methoxychlor	0.5	0.48	ug/L	97				78	108	
	Endrin ketone	0.5	0.53	ug/L	106				80	131	
	Endrin aldehyde	0.5	0.53	ug/L	107				82	127	
	alpha-Chlordane	0.5	0.53	ug/L	106				82	125	
	gamma-Chlordane	0.5	0.52	ug/L	104				82	125	
PB168004BS (Column 2)	alpha-BHC	0.5	0.47	ug/L					85	130	
	beta-BHC	0.5	0.45	ug/L					83	126	
	delta-BHC	0.5	0.47	ug/L					69	141	
	gamma-BHC (Lindane)	0.5	0.47	ug/L					82	129	
	Heptachlor	0.5	0.45	ug/L					79	127	
	Aldrin	0.5	0.46	ug/L					79	126	
	Heptachlor epoxide	0.5	0.47	ug/L					81	124	
	Endosulfan I	0.5	0.50	ug/L					85	122	
	Dieldrin	0.5	0.47	ug/L					83	125	
	4,4'-DDE	0.5	0.47	ug/L					80	127	
	Endrin	0.5	0.45	ug/L					81	128	
	Endosulfan II	0.5	0.46	ug/L					82	123	
	4,4'-DDD	0.5	0.49	ug/L					77	131	
	Endosulfan sulfate	0.5	0.47	ug/L					76	129	
	4,4'-DDT	0.5	0.44	ug/L					80	133	
	Methoxychlor	0.5	0.43	ug/L					78	108	
	Endrin ketone	0.5	0.48	ug/L					80	131	
	Endrin aldehyde	0.5	0.49	ug/L					82	127	
	alpha-Chlordane	0.5	0.46	ug/L					82	125	
	gamma-Chlordane	0.5	0.47	ug/L					82	125	