

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

B27-SOIL-SAMPLEMS

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032

SAS No.: Q2032 SDG NO.: Q2032

Lab Sample ID: Q2027-03MS

Date(s) Analyzed: 05/20/2025 05/20/2025

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column: (1): RTX-CLP ID: 0.32 (mm)

GC Column:(2): RTX-CLP2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
2,4-D	1	8.01	7.96	8.06	56.4	1.2
	2	8.64	8.59	8.69	57.1	
2,4,5-TP(Silvex)	1	8.85	8.80	8.90	52.7	3
	2	9.52	9.47	9.57	54.3	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

B27-SOIL-SAMPLEMSD

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032

SAS No.: Q2032 SDG NO.: Q2032

Lab Sample ID: Q2027-03MSD

Date(s) Analyzed: 05/20/2025 05/20/2025

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column: (1): RTX-CLP ID: 0.32 (mm)

GC Column:(2): RTX-CLP2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
2,4-D	1	8.01	7.96	8.06	56.5	0.7
	2	8.64	8.59	8.69	56.9	
2,4,5-TP(Silvex)	1	8.85	8.80	8.90	52.6	2.4
	2	9.52	9.47	9.57	53.9	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB168065BS

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032

SAS No.: Q2032

SDG NO.: Q2032

Lab Sample ID: PB168065BS

Date(s) Analyzed: 05/19/2025 05/19/2025

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column: (1): RTX-CLP ID: 0.32 (mm)

GC Column:(2): RTX-CLP2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
2,4-D	1	8.01	7.96	8.06	5.20	1.9
	2	8.64	8.59	8.69	5.10	
2,4,5-TP(Silvex)	1	8.85	8.80	8.90	5.20	0
	2	9.52	9.47	9.57	5.20	