

**CALIBRATION VERIFICATION SUMMARY**

**Contract:** CAMP02

**Lab Code:** CHEM      **Case No.:** Q2032      **SAS No.:** Q2032      **SDG NO.:** Q2032

**Continuing Calib Date:** 05/20/2025      **Initial Calibration Date(s):** 05/12/2025      05/12/2025

**Continuing Calib Time:** 02:15      **Initial Calibration Time(s):** 12:30      14:06

**GC Column:** RTX-CLP      **ID:** 0.32      (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	6.93	6.93	6.83	7.03	0.00
2,4-D	8.01	8.01	7.91	8.11	0.00
2,4,5-TP(Silvex)	8.85	8.85	8.75	8.95	0.00

**CALIBRATION VERIFICATION SUMMARY**

**Contract:** CAMP02

**Lab Code:** CHEM      **Case No.:** Q2032      **SAS No.:** Q2032      **SDG NO.:** Q2032

**Continuing Calib Date:** 05/20/2025      **Initial Calibration Date(s):** 05/12/2025      05/12/2025

**Continuing Calib Time:** 02:15      **Initial Calibration Time(s):** 12:30      14:06

**GC Column:** RTX-CLP2      **ID:** 0.32      (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.45	7.45	7.35	7.55	0.00
2,4-D	8.64	8.64	8.54	8.74	0.00
2,4,5-TP(Silvex)	9.52	9.52	9.42	9.62	0.00