

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL02 Date Analyzed: 05/20/2025

Lab Sample No.: HSTDCCC750 Data File : PS030300.D Time Analyzed: 02:15

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	8.847	8.750	8.950	740.100	712.500	3.9
2,4-D	8.006	7.910	8.110	727.670	705.000	3.2
2,4-DCAA	6.930	6.834	7.034	766.930	750.000	2.3

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL02 Date Analyzed: 05/20/2025

Lab Sample No.: HSTDCCC750 Data File : PS030300.D Time Analyzed: 02:15

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.516	9.423	9.623	744.460	712.500	4.5
2,4-D	8.638	8.544	8.744	726.120	705.000	3.0
2,4-DCAA	7.449	7.354	7.554	755.260	750.000	0.7