

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

Continuing Calib Date: 05/19/2025 Initial Calibration Date(s): 05/19/2025 05/19/2025

Continuing Calib Time: 18:45 Initial Calibration Time(s): 11:31 12:25

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	9.08	9.08	8.98	9.18	0.00
Tetrachloro-m-xylene	3.55	3.55	3.45	3.65	0.00
gamma-BHC (Lindane)	4.33	4.33	4.23	4.43	0.00
Heptachlor	4.93	4.93	4.83	5.03	0.00
Heptachlor epoxide	5.69	5.69	5.59	5.79	0.00
Endrin	6.58	6.58	6.48	6.68	0.00
Methoxychlor	7.50	7.50	7.40	7.60	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

Continuing Calib Date: 05/19/2025 Initial Calibration Date(s): 05/19/2025 05/19/2025

Continuing Calib Time: 18:45 Initial Calibration Time(s): 11:31 12:25

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	8.08	8.08	7.98	8.18	0.00
Tetrachloro-m-xylene	2.88	2.88	2.78	2.98	0.00
gamma-BHC (Lindane)	3.73	3.73	3.63	3.83	0.00
Heptachlor	4.09	4.09	3.99	4.19	0.00
Heptachlor epoxide	4.88	4.88	4.78	4.98	0.00
Endrin	5.79	5.79	5.69	5.89	0.00
Methoxychlor	6.76	6.76	6.66	6.86	0.00