

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL01 Date Analyzed: 05/19/2025

Lab Sample No.: PSTDCCC050 Data File : PD088614.D Time Analyzed: 18:45

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.078	8.977	9.177	53.900	50.000	7.8
Endrin	6.577	6.477	6.677	51.960	50.000	3.9
gamma-BHC (Lindane)	4.332	4.232	4.432	53.570	50.000	7.1
Heptachlor	4.931	4.832	5.032	51.060	50.000	2.1
Heptachlor epoxide	5.693	5.593	5.793	52.510	50.000	5.0
Methoxychlor	7.496	7.395	7.595	43.320	50.000	-13.4
Tetrachloro-m-xylene	3.551	3.451	3.651	58.460	50.000	16.9

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL01 Date Analyzed: 05/19/2025

Lab Sample No.: PSTDCCC050 Data File : PD088614.D Time Analyzed: 18:45

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.076	7.977	8.177	50.220	50.000	0.4
Endrin	5.792	5.692	5.892	50.570	50.000	1.1
gamma-BHC (Lindane)	3.731	3.631	3.831	51.910	50.000	3.8
Heptachlor	4.085	3.985	4.185	49.920	50.000	-0.2
Heptachlor epoxide	4.875	4.775	4.975	51.370	50.000	2.7
Methoxychlor	6.757	6.658	6.858	45.890	50.000	-8.2
Tetrachloro-m-xylene	2.882	2.782	2.982	57.720	50.000	15.4