

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL02 Date Analyzed: 05/19/2025

Lab Sample No.: PSTDCCC050 Data File : PD088626.D Time Analyzed: 22:43

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.077	8.977	9.177	55.800	50.000	11.6
Endrin	6.577	6.477	6.677	52.910	50.000	5.8
gamma-BHC (Lindane)	4.332	4.232	4.432	54.090	50.000	8.2
Heptachlor	4.931	4.832	5.032	52.230	50.000	4.5
Heptachlor epoxide	5.692	5.593	5.793	53.180	50.000	6.4
Methoxychlor	7.496	7.395	7.595	45.060	50.000	-9.9
Tetrachloro-m-xylene	3.551	3.451	3.651	58.570	50.000	17.1

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL02 Date Analyzed: 05/19/2025

Lab Sample No.: PSTDCCC050 Data File : PD088626.D Time Analyzed: 22:43

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.076	7.977	8.177	53.680	50.000	7.4
Endrin	5.791	5.692	5.892	51.180	50.000	2.4
gamma-BHC (Lindane)	3.731	3.631	3.831	52.150	50.000	4.3
Heptachlor	4.085	3.985	4.185	50.650	50.000	1.3
Heptachlor epoxide	4.875	4.775	4.975	51.470	50.000	2.9
Methoxychlor	6.757	6.658	6.858	47.690	50.000	-4.6
Tetrachloro-m-xylene	2.882	2.782	2.982	57.660	50.000	15.3