

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL03 Date Analyzed: 05/20/2025

Lab Sample No.: PSTDCCC050 Data File : PD088630.D Time Analyzed: 09:23

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.087	8.977	9.177	54.480	50.000	9.0
Endrin	6.585	6.477	6.677	50.820	50.000	1.6
gamma-BHC (Lindane)	4.340	4.232	4.432	50.770	50.000	1.5
Heptachlor	4.940	4.832	5.032	50.890	50.000	1.8
Heptachlor epoxide	5.701	5.593	5.793	50.370	50.000	0.7
Methoxychlor	7.504	7.395	7.595	44.830	50.000	-10.3
Tetrachloro-m-xylene	3.559	3.451	3.651	56.100	50.000	12.2

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL03 Date Analyzed: 05/20/2025

Lab Sample No.: PSTDCCC050 Data File : PD088630.D Time Analyzed: 09:23

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.078	7.977	8.177	52.570	50.000	5.1
Endrin	5.792	5.692	5.892	47.720	50.000	-4.6
gamma-BHC (Lindane)	3.731	3.631	3.831	49.620	50.000	-0.8
Heptachlor	4.085	3.985	4.185	47.730	50.000	-4.5
Heptachlor epoxide	4.875	4.775	4.975	48.770	50.000	-2.5
Methoxychlor	6.759	6.658	6.858	45.580	50.000	-8.8
Tetrachloro-m-xylene	2.881	2.782	2.982	54.650	50.000	9.3