

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL04 Date Analyzed: 05/20/2025

Lab Sample No.: PSTDCCC050 Data File : PD088642.D Time Analyzed: 13:11

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.076	8.977	9.177	51.680	50.000	3.4
Endrin	6.577	6.477	6.677	48.650	50.000	-2.7
gamma-BHC (Lindane)	4.332	4.232	4.432	52.150	50.000	4.3
Heptachlor	4.932	4.832	5.032	50.800	50.000	1.6
Heptachlor epoxide	5.693	5.593	5.793	49.860	50.000	-0.3
Methoxychlor	7.495	7.395	7.595	41.700	50.000	-16.6
Tetrachloro-m-xylene	3.551	3.451	3.651	56.820	50.000	13.6

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL04 Date Analyzed: 05/20/2025

Lab Sample No.: PSTDCCC050 Data File : PD088642.D Time Analyzed: 13:11

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.075	7.977	8.177	48.520	50.000	-3.0
Endrin	5.791	5.692	5.892	48.260	50.000	-3.5
gamma-BHC (Lindane)	3.731	3.631	3.831	50.500	50.000	1.0
Heptachlor	4.084	3.985	4.185	49.120	50.000	-1.8
Heptachlor epoxide	4.874	4.775	4.975	49.530	50.000	-0.9
Methoxychlor	6.756	6.658	6.858	44.770	50.000	-10.5
Tetrachloro-m-xylene	2.882	2.782	2.982	55.910	50.000	11.8