

Surrogate Summary

SDG No.: Q2032

Client: CDM Smith

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PD088583.D	PIBLK-PD088583.D	Decachlorobiphenyl	1	20	18.4	92		43	140
		Tetrachloro-m-xylene	1	20	16.5	83		77	126
		Decachlorobiphenyl	2	20	18.0	90		43	140
		Tetrachloro-m-xylene	2	20	16.9	84		77	126
I.BLK-PD088613.D	PIBLK-PD088613.D	Decachlorobiphenyl	1	20	15.0	75		43	140
		Tetrachloro-m-xylene	1	20	17.3	87		77	126
		Decachlorobiphenyl	2	20	15.6	78		43	140
		Tetrachloro-m-xylene	2	20	18.1	91		77	126
PB168066BL	PB168066BL	Decachlorobiphenyl	1	20	19.6	98		43	140
		Tetrachloro-m-xylene	1	20	18.2	91		77	126
		Decachlorobiphenyl	2	20	19.2	96		43	140
		Tetrachloro-m-xylene	2	20	19.9	100		77	126
PB167994TB	PB167994TB	Decachlorobiphenyl	1	20	20.7	104		43	140
		Tetrachloro-m-xylene	1	20	19.2	96		77	126
		Decachlorobiphenyl	2	20	20.4	102		43	140
		Tetrachloro-m-xylene	2	20	20.9	104		77	126
Q2027-03MS	B27-SOIL-SAMPLEMS	Decachlorobiphenyl	1	20	13.0	65		43	140
		Tetrachloro-m-xylene	1	20	16.6	83		77	126
		Decachlorobiphenyl	2	20	12.3	62		43	140
		Tetrachloro-m-xylene	2	20	18.1	90		77	126
Q2027-03MSD	B27-SOIL-SAMPLEMSD	Decachlorobiphenyl	1	20	13.0	65		43	140
		Tetrachloro-m-xylene	1	20	16.6	83		77	126
		Decachlorobiphenyl	2	20	12.7	64		43	140
		Tetrachloro-m-xylene	2	20	18.6	93		77	126
Q2032-09	COMP-1	Decachlorobiphenyl	1	20	21.6	108		43	140
		Tetrachloro-m-xylene	1	20	20.4	102		77	126
		Decachlorobiphenyl	2	20	20.6	103		43	140
		Tetrachloro-m-xylene	2	20	20.1	101		77	126
I.BLK-PD088625.D	PIBLK-PD088625.D	Decachlorobiphenyl	1	20	19.5	97		43	140
		Tetrachloro-m-xylene	1	20	17.9	89		77	126
		Decachlorobiphenyl	2	20	19.1	95		43	140
		Tetrachloro-m-xylene	2	20	18.5	92		77	126
I.BLK-PD088628.D	PIBLK-PD088628.D	Decachlorobiphenyl	1	20	20.2	101		43	140
		Tetrachloro-m-xylene	1	20	18.4	92		77	126
		Decachlorobiphenyl	2	20	19.7	99		43	140
		Tetrachloro-m-xylene	2	20	18.8	94		77	126
PB168066BS	PB168066BS	Decachlorobiphenyl	1	20	23.8	119		43	140
		Tetrachloro-m-xylene	1	20	21.2	106		77	126
		Decachlorobiphenyl	2	20	23.5	117		43	140
		Tetrachloro-m-xylene	2	20	22.7	113		77	126
I.BLK-PD088641.D	PIBLK-PD088641.D	Decachlorobiphenyl	1	20	18.0	90		43	140

Surrogate Summary

SDG No.: Q2032

Client: CDM Smith

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PD088641.D	PIBLK-PD088641.D	Tetrachloro-m-xylene	1	20	18.2	91		77	126
		Decachlorobiphenyl	2	20	17.7	88		43	140
		Tetrachloro-m-xylene	2	20	18.9	95		77	126