

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2032</u>	Analytical Method:	<u>8081B</u>
Client:	<u>CDM Smith</u>	Datafile :	<u>PD088633.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits		RPD
								Qual	Low	High		
PB168066BS (Column 1)	gamma-BHC (Lindane)	5	5.30	ug/L	106				82	129		
	Heptachlor	5	5.30	ug/L	106				79	127		
	Heptachlor epoxide	5	5.40	ug/L	108				81	124		
	Endrin	5	5.30	ug/L	106				81	128		
	Methoxychlor	5	4.60	ug/L	92				78	108		
PB168066BS (Column 2)	gamma-BHC (Lindane)	5	5.30	ug/L	106				82	129		
	Heptachlor	5	5.10	ug/L	102				79	127		
	Heptachlor epoxide	5	5.30	ug/L	106				81	124		
	Endrin	5	5.20	ug/L	104				81	128		
	Methoxychlor	5	4.60	ug/L	92				78	108		