

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: CAMP02

Lab Code: CHEM

SAS No.: Q2032

SDG NO.: Q2032

Lab File ID: BF142294.D

DFTPP Injection Date: 05/05/2025

Instrument ID: BNA_F

DFTPP Injection Time: 13:24

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	32
68	Less than 2.0% of mass 69	0.5 (1.8) 1
69	Mass 69 relative abundance	25.8
70	Less than 2.0% of mass 69	0.1 (0.5) 1
127	10.0 - 80.0% of mass 198	34.9
197	Less than 2.0% of mass 198	0.2
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	5.2
275	10.0 - 60.0% of mass 198	22.8
365	Greater than 1% of mass 198	3
441	Present, but less than mass 443	15.4
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.6 (19.6) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BF142295.D	05/05/2025	13:54
SSTDICC005	SSTDICC005	BF142296.D	05/05/2025	14:23
SSTDICC010	SSTDICC010	BF142297.D	05/05/2025	14:52
SSTDICC020	SSTDICC020	BF142298.D	05/05/2025	15:21
SSTDICCC040	SSTDICCC040	BF142299.D	05/05/2025	15:50
SSTDICC050	SSTDICC050	BF142300.D	05/05/2025	16:18
SSTDICC060	SSTDICC060	BF142301.D	05/05/2025	16:47
SSTDICC080	SSTDICC080	BF142302.D	05/05/2025	17:15