

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: CAMP02

Lab Code: CHEM

SAS No.: Q2032

SDG NO.: Q2032

Lab File ID: BP024702.D

DFTPP Injection Date: 05/20/2025

Instrument ID: BNA\_P

DFTPP Injection Time: 10:31

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	22.2
68	Less than 2.0% of mass 69	0.3 ( 1.3 ) 1
69	Mass 69 relative abundance	23.8
70	Less than 2.0% of mass 69	0.1 ( 0.6 ) 1
127	10.0 - 80.0% of mass 198	33.7
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	4.7
275	10.0 - 60.0% of mass 198	24
365	Greater than 1% of mass 198	3.9
441	Present, but less than mass 443	15.5
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19 ( 19 ) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BP024703.D	05/20/2025	11:52
PB168048BL	PB168048BL	BP024706.D	05/20/2025	13:55
PB168048BS	PB168048BS	BP024707.D	05/20/2025	14:35
PB168048BSD	PB168048BSD	BP024708.D	05/20/2025	15:16
FB-05132025	Q2032-11	BP024716.D	05/20/2025	20:42