

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/21/2025

Lab File ID: BP024737.D Time Analyzed: 14:48

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	128557	7.646	504282	10.42	338214	14.29
UPPER LIMIT	257114	8.146	1008560	10.916	676428	14.787
LOWER LIMIT	64278.5	7.146	252141	9.916	169107	13.787
EPA SAMPLE NO.						
01 PB168026BS	134273	7.65	511993	10.42	321175	14.29

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.