

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/21/2025

Lab File ID: BP024737.D Time Analyzed: 14:48

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

01

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	643332	17.081	728456	21.51	831576	24.786
UPPER LIMIT	1286660	17.581	1456910	22.01	1663150	25.286
LOWER LIMIT	321666	16.581	364228	21.01	415788	24.286
EPA SAMPLE NO.						
PB168026BS	621819	17.09	697911	21.51	814994	24.79

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.