

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CAMP02

Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG No.: Q2032

Instrument ID: MSVOA_X Calibration Date/Time: 05/15/2025 09:52

Lab File ID: VX046200.D Init. Calib. Date(s): 05/05/2025 05/05/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:35 16:27

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.691	0.611		-11.58	20
1,1-Dichloroethene	0.593	0.532		-10.29	20
2-Butanone	0.543	0.583		7.37	20
Carbon Tetrachloride	0.544	0.510		-6.25	20
Chloroform	1.271	1.299		2.2	20
Benzene	1.417	1.399		-1.27	20
1,2-Dichloroethane	0.612	0.634		3.6	20
Trichloroethene	0.341	0.328		-3.81	20
Tetrachloroethene	0.354	0.326		-7.91	20
Chlorobenzene	1.094	1.065	0.3	-2.65	20
1,2-Dichloroethane-d4	0.932	0.941		0.97	20
Dibromofluoromethane	0.360	0.366		1.67	20
Toluene-d8	1.246	1.215		-2.49	20
4-Bromofluorobenzene	0.478	0.492		2.93	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.