

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2032

Client: CDM Smith

Analytical Method: SW8260-Low

Datafile : VX046206.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0515WBS01	Vinyl chloride	20	19.1	ug/L	96			65	117	
	1,1-Dichloroethene	20	19.0	ug/L	95			74	110	
	2-Butanone	100	110	ug/L	110			65	122	
	Carbon Tetrachloride	20	19.5	ug/L	98			77	113	
	Chloroform	20	21.0	ug/L	105			79	113	
	Benzene	20	20.1	ug/L	101			82	109	
	1,2-Dichloroethane	20	20.6	ug/L	103			80	115	
	Trichloroethene	20	19.3	ug/L	97			77	113	
	Tetrachloroethene	20	19.5	ug/L	98			67	123	
	Chlorobenzene	20	19.5	ug/L	98			82	109	