

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2032

Client: CDM Smith

Analytical Method: SW8260-Low

Datafile : VX046207.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0515WBSD01	Vinyl chloride	20	18.2	ug/L	91	5		65	117	20
	1,1-Dichloroethene	20	17.9	ug/L	90	5		74	110	20
	2-Butanone	100	110	ug/L	110	0		65	122	20
	Carbon Tetrachloride	20	19.2	ug/L	96	2		77	113	20
	Chloroform	20	20.6	ug/L	103	2		79	113	20
	Benzene	20	19.7	ug/L	99	2		82	109	20
	1,2-Dichloroethane	20	20.7	ug/L	104	1		80	115	20
	Trichloroethene	20	18.8	ug/L	94	3		77	113	20
	Tetrachloroethene	20	19.3	ug/L	97	1		67	123	20
	Chlorobenzene	20	19.5	ug/L	98	0		82	109	20