

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02
 Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032
 Lab File ID: VX046200.D Date Analyzed: 05/15/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:52
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	79845	5.54	140483	6.75	127630	10.05
UPPER LIMIT	159690	6.044	280966	7.251	255260	10.549
LOWER LIMIT	39922.5	5.044	70241.5	6.251	63815	9.549
EPA SAMPLE NO.						
COMP-1	60792	5.55	124950	6.76	117007	10.06
VX0515WBL01	65765	5.54	130899	6.76	122283	10.05
VX0515WBS01	89041	5.54	157824	6.76	139362	10.05
VX0515WBSD01	86433	5.54	151750	6.76	132818	10.05

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.