

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN051625

Review By	John Carlone	Review On	5/19/2025 9:34:08 AM
Supervise By	Maresh Dadoda	Supervise On	5/19/2025 2:40:58 PM
SubDirectory	VN051625	HP Acquire Method	HP Processing Method 624N042325W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133946,VP133952
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133947,VP133948 VP133981

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN086660.D	16 May 2025 09:17		JC\MD	Ok
2	VSTDCCC020	VSTDCCC020	VN086661.D	16 May 2025 11:37	pH#Lot#V12668	JC\MD	Ok,M
3	VN0516WBS01	VN0516WBS01	VN086662.D	16 May 2025 12:13		JC\MD	Ok,M
4	VN0516WBSD01	VN0516WBSD01	VN086663.D	16 May 2025 12:46		JC\MD	Ok,M
5	VN0516WBL01	VN0516WBL01	VN086664.D	16 May 2025 13:20		JC\MD	Ok
6	Q2006-01	EFF-WW	VN086665.D	16 May 2025 13:54	vial A pH<2	JC\MD	Ok
7	IBLK	IBLK	VN086666.D	16 May 2025 14:46		JC\MD	Ok
8	VSTDCCC020	VSTDCCC020EC	VN086667.D	16 May 2025 15:09		JC\MD	Ok,M
9	BFB	BFB	VN086668.D	16 May 2025 15:33		JC\MD	Ok
10	VSTDICC001	VSTDICC001	VN086669.D	16 May 2025 16:31		JC\MD	Ok,M
11	VSTDICC005	VSTDICC005	VN086670.D	16 May 2025 16:55		JC\MD	Ok,M
12	VSTDICC010	VSTDICC010	VN086671.D	16 May 2025 17:19		JC\MD	Ok,M
13	VSTDICC020	VSTDICC020	VN086672.D	16 May 2025 17:43		JC\MD	Ok,M
14	VSTDICCC050	VSTDICCC050	VN086673.D	16 May 2025 18:08		JC\MD	Ok,M
15	VSTDICC100	VSTDICC100	VN086674.D	16 May 2025 18:32		JC\MD	Ok,M
16	IBLK	IBLK	VN086675.D	16 May 2025 18:56		JC\MD	Ok
17	VSTDICV050	ICVVN051625	VN086676.D	16 May 2025 19:20		JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN052125

Review By	John Carlone	Review On	5/22/2025 8:01:52 AM
Supervise By	Maresh Dadoda	Supervise On	5/22/2025 8:24:13 AM
SubDirectory	VN052125	HP Acquire Method	HP Processing Method 82N051625W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133982
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133983,VP133984

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN086721.D	21 May 2025 10:37		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN086722.D	21 May 2025 11:10	pH#Lot#V12668	JC\MD	Ok,M
3	VN0521MBL01	VN0521MBL01	VN086723.D	21 May 2025 11:44		JC\MD	Ok
4	VN0521WBL01	VN0521WBL01	VN086724.D	21 May 2025 12:08		JC\MD	Ok
5	VN0521MBS01	VN0521MBS01	VN086725.D	21 May 2025 12:32		JC\MD	Ok,M
6	VN0521WBS01	VN0521WBS01	VN086726.D	21 May 2025 13:05		JC\MD	Ok,M
7	Q2068-01	LAW-25-0075	VN086727.D	21 May 2025 13:29	run straight MEOH	JC\MD	Not Ok
8	VN0521WBSD01	VN0521WBSD01	VN086728.D	21 May 2025 13:53		JC\MD	Ok,M
9	Q2068-01	LAW-25-0075	VN086729.D	21 May 2025 14:17		JC\MD	Ok
10	PB168093TB	PB168093TB	VN086730.D	21 May 2025 14:41		JC\MD	Ok
11	Q2078-05	WC-A4-05-G	VN086731.D	21 May 2025 15:05	vial A pH#5.0 Surrogate fail;Need 20X	JC\MD	Dilution
12	Q2078-09	WC-A1-06A-G	VN086732.D	21 May 2025 15:29	vial A pH#5.0 Surrogate Fail	JC\MD	ReRun
13	Q2078-13	WC-A1-07A-G	VN086733.D	21 May 2025 15:53	vial A pH#5.0	JC\MD	Ok
14	Q2078-17	WC-A4-06-G	VN086734.D	21 May 2025 16:18	vial A pH#5.0	JC\MD	Ok
15	Q2079-01	WC-A2-01-G	VN086735.D	21 May 2025 16:42	vial A pH#5.0 Surrogate Fail	JC\MD	ReRun
16	Q2079-05	WC-A2-02-G	VN086736.D	21 May 2025 17:06	vial A pH#5.0 Surrogate Fail	JC\MD	ReRun
17	Q2079-09	WC-A2-03-G	VN086737.D	21 May 2025 17:30	vial A pH#5.0	JC\MD	Ok

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN052125

Review By	John Carlone	Review On	5/22/2025 8:01:52 AM
Supervise By	Maresh Dadoda	Supervise On	5/22/2025 8:24:13 AM
SubDirectory	VN052125	HP Acquire Method	HP Processing Method 82N051625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133982		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133983,VP133984		

18	Q2084-05	OR-640-COMP-66	VN086738.D	21 May 2025 17:54	vial A pH#5.0	JC\MD	Ok
19	Q2092-02	VNJ-202	VN086739.D	21 May 2025 18:18	vial A pH#5.0	JC\MD	Ok
20	Q2093-02	SOIL-DRUM	VN086740.D	21 May 2025 18:42	vial A pH#5.0	JC\MD	Ok
21	Q2095-04	WCS-TP1	VN086741.D	21 May 2025 19:06	vial A pH#5.0	JC\MD	Ok
22	Q2095-08	WCS-TP2	VN086742.D	21 May 2025 19:31	vial A pH#5.0 RT Shifted	JC\MD	Ok,M
23	Q2078-01	WC-A4-04-G	VN086743.D	21 May 2025 19:54	vial A pH#5.0 Surrogate Fail	JC\MD	ReRun
24	Q2032-01ME	TP-11ME	VN086744.D	21 May 2025 20:18	vial C	JC\MD	Ok
25	Q2075-06ME	SS-MW1-11.5ME	VN086745.D	21 May 2025 20:42	Not Match	JC\MD	Not Ok
26	Q2071-19DL	L2-WC-6-VOC DL	VN086746.D	21 May 2025 21:06		JC\MD	Ok
27	VN0521MBS01	VN0521MBS01	VN086747.D	21 May 2025 21:30	Recovery failed	JC\MD	Not Ok
28	VSTDCCC050	VSTDCCC050EC	VN086748.D	21 May 2025 21:54		JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX050525

Review By	John Carlone	Review On	5/6/2025 9:53:58 AM
Supervise By	Mahesh Dadoda	Supervise On	5/6/2025 12:43:00 PM
SubDirectory	VX050525	HP Acquire Method	HP Processing Method 82X050525W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP133811		
Initial Calibration Stds	VP133832,VP133833,VP133834,VP133835,VP133836,VP133837		
CCC			
Internal Standard/PEM			
ICV/I.BLK	VP133838		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampled	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046038.D	05 May 2025 09:37		JC/MD	Ok
2	VSTDIC001	VSTDIC001	VX046039.D	05 May 2025 10:49	Not used	JC/MD	Not Ok
3	VSTDIC005	VSTDIC005	VX046040.D	05 May 2025 11:12	Not used	JC/MD	Not Ok
4	VSTDIC020	VSTDIC020	VX046041.D	05 May 2025 11:35		JC/MD	Ok,M
5	VSTDIC050	VSTDIC050	VX046042.D	05 May 2025 11:58		JC/MD	Ok,M
6	VSTDIC100	VSTDIC100	VX046043.D	05 May 2025 12:21		JC/MD	Ok,M
7	VSTDIC150	VSTDIC150	VX046044.D	05 May 2025 12:45		JC/MD	Ok,M
8	IBLK	IBLK	VX046045.D	05 May 2025 13:08		JC/MD	Ok
9	VSTDIC005	VSTDIC005	VX046046.D	05 May 2025 16:04		JC/MD	Ok,M
10	VSTDIC001	VSTDIC001	VX046047.D	05 May 2025 16:27		JC/MD	Ok,M
11	VSTDICV050	ICVVX050525	VX046048.D	05 May 2025 16:50		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX051525

Review By	John Carlone	Review On	5/16/2025 9:22:10 AM
Supervise By	Maresh Dadoda	Supervise On	5/16/2025 11:49:40 AM
SubDirectory	VX051525	HP Acquire Method	HP Processing Method 82X050525W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133928
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133929,VP133930

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046199.D	15 May 2025 08:47		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX046200.D	15 May 2025 09:52	pH#Lot#V12668	JC/MD	Ok,M
3	VX0515MBL01	VX0515MBL01	VX046201.D	15 May 2025 10:15		JC/MD	Ok
4	VX0515WBL01	VX0515WBL01	VX046202.D	15 May 2025 10:38		JC/MD	Ok
5	Q2018-06	FB	VX046203.D	15 May 2025 11:01	vial B pH<2 FB	JC/MD	Ok
6	Q2032-11	FB-05132025	VX046204.D	15 May 2025 11:25	vial B pH<2 FB	JC/MD	Ok
7	Q2032-12	TB	VX046205.D	15 May 2025 11:48	vial B pH<2 TB	JC/MD	Ok
8	VX0515WBS01	VX0515WBS01	VX046206.D	15 May 2025 12:11		JC/MD	Ok,M
9	VX0515WBSD01	VX0515WBSD01	VX046207.D	15 May 2025 12:38		JC/MD	Ok,M
10	Q2050-02	BP-VPB-182-GW-60-62	VX046208.D	15 May 2025 13:01	vial A pH<2	JC/MD	Ok
11	Q2050-01	BP-VPB-182-TB-20250	VX046209.D	15 May 2025 13:24	vial A pH<2 TB	JC/MD	Ok
12	Q2050-08	BP-VPB-182-EB-20250	VX046210.D	15 May 2025 13:47	vial A pH<2 EB	JC/MD	Ok
13	Q2050-03	BP-VPB-182-GW-100-1	VX046211.D	15 May 2025 14:11	vial A pH<2	JC/MD	Ok
14	Q2050-04	BP-VPB-182-GW-150-1	VX046212.D	15 May 2025 14:34	vial A pH<2	JC/MD	Ok
15	Q2050-05	BP-VPB-182-GW-205-2	VX046213.D	15 May 2025 14:58	vial A pH<2	JC/MD	Ok
16	Q2050-06	BP-VPB-182-GW-220-2	VX046214.D	15 May 2025 15:21	vial A pH<2	JC/MD	Ok
17	Q2050-07	BP-VPB-182-DUP-2025	VX046215.D	15 May 2025 15:44	vial A pH<2	JC/MD	Ok
18	Q2050-09	BP-VPB-182-GW-240-2	VX046216.D	15 May 2025 16:08	vial A pH<2	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX051525

Review By	John Carlone	Review On	5/16/2025 9:22:10 AM
Supervise By	Maresh Dadoda	Supervise On	5/16/2025 11:49:40 AM
SubDirectory	VX051525	HP Acquire Method	HP Processing Method 82X050525W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133928		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133929,VP133930		

19	Q2019-04	MH-K	VX046217.D	15 May 2025 16:31	vial A pH#5.0	JC/MD	Ok
20	Q2034-16	L3-WC-4	VX046218.D	15 May 2025 16:54	vial A pH#5.0	JC/MD	Ok
21	Q2034-20	L3-WC-5	VX046219.D	15 May 2025 17:18	vial A pH#5.0	JC/MD	Ok
22	Q2032-09	COMP-1	VX046220.D	15 May 2025 17:41	vial A pH#5.0	JC/MD	Ok
23	Q2034-04	L3-WC-1	VX046221.D	15 May 2025 18:04	vial A pH#5.0	JC/MD	Ok
24	Q2034-08	L3-WC-2	VX046222.D	15 May 2025 18:28	vial A pH#5.0	JC/MD	Ok
25	Q2034-12	L3-WC-3	VX046223.D	15 May 2025 18:51	vial A pH#5.0	JC/MD	Ok
26	Q2038-02	72-11991	VX046224.D	15 May 2025 19:15	vial A pH#5.0	JC/MD	Ok
27	Q2043-05	3139	VX046225.D	15 May 2025 19:38	vial A pH<2 sample oily,smelly	JC/MD	Ok
28	VSTDCCC050	VSTDCCC050EC	VX046226.D	15 May 2025 20:01		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY051525

Review By	Mahesh Dadoda	Review On	5/16/2025 11:49:17 AM				
Supervise By	Semsettin Yesilyurt	Supervise On	5/16/2025 11:51:38 AM				
SubDirectory	VY051525	HP Acquire Method	MSVOA_Y	HP Processing Method	82y051525s.m		
STD. NAME		STD REF.#					
Tune/Reschk		VP133916					
Initial Calibration Stds		VP133918,VP133919,VP133920,VP133921,VP133922,VP133923					
CCC		VP133925					
Internal Standard/PEM		VP131783					
ICV/I.BLK		VP133924					
Surrogate Standard							
MS/MSD Standard							
LCS Standard							

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022252.D	15 May 2025 07:52		SY/MD	Ok
2	VSTDICC005	VSTDICC005	VY022253.D	15 May 2025 09:46		SY/MD	Ok,M
3	VSTDICC010	VSTDICC010	VY022254.D	15 May 2025 10:16		SY/MD	Ok,M
4	VSTDICC020	VSTDICC020	VY022255.D	15 May 2025 10:39		SY/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VY022256.D	15 May 2025 11:02		SY/MD	Ok,M
6	VSTDICC100	VSTDICC100	VY022257.D	15 May 2025 11:24		SY/MD	Ok,M
7	VSTDICC150	VSTDICC150	VY022258.D	15 May 2025 11:47		SY/MD	Ok,M
8	VIBLK	VIBLK	VY022259.D	15 May 2025 12:34		SY/MD	Ok
9	VSTDICV050	ICVVY051525	VY022260.D	15 May 2025 13:20		SY/MD	Ok,M
10	VY0515SBL01	VY0515SBL01	VY022261.D	15 May 2025 13:44		SY/MD	Ok
11	VY0515SBS01	VY0515SBS01	VY022262.D	15 May 2025 14:07		SY/MD	Ok,M
12	VY0515SBSD01	VY0515SBSD01	VY022263.D	15 May 2025 14:29		SY/MD	Ok,M
13	Q1984-01	OU4-PCS-TC-33-05072	VY022264.D	15 May 2025 15:12	vial-A	SY/MD	Ok
14	Q1984-03	OU4-PCS-TC-34-05072	VY022265.D	15 May 2025 15:35	vial-A	SY/MD	Ok
15	Q1984-05	OU4-PCS-TC-35-05072	VY022266.D	15 May 2025 15:59	vial-A	SY/MD	Ok
16	Q1984-07	OU4-TS-24-050725	VY022267.D	15 May 2025 16:22	vial-A Surrogate fail	SY/MD	ReRun
17	Q1984-09	OU4-TS-25-050725	VY022268.D	15 May 2025 16:46	vial-A Surrogate fail	SY/MD	Not Ok
18	Q1984-11	OU4-TS-26-050725	VY022269.D	15 May 2025 17:09	vial-A Surrogate fail	SY/MD	Not Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY051525

Review By	Mahesh Dadoda	Review On	5/16/2025 11:49:17 AM		
Supervise By	Semsettin Yesilyurt	Supervise On	5/16/2025 11:51:38 AM		
SubDirectory	VY051525	HP Acquire Method	MSVOA_Y	HP Processing Method	82y051525s.m
STD. NAME		STD REF.#			
Tune/Reschk		VP133916			
Initial Calibration Stds		VP133918,VP133919,VP133920,VP133921,VP133922,VP133923			
CCC		VP133925			
Internal Standard/PEM		VP131783			
ICV/I.BLK		VP133924			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q1984-13	OU4-TS-27-050725	VY022270.D	15 May 2025 17:33	vial-A Surrogate fail	SY/MD	Not Ok
20	Q1984-15	OU4-TS-28-050725	VY022271.D	15 May 2025 17:56	vial-A Surrogate fail	SY/MD	ReRun
21	Q1982-04	TP-4	VY022272.D	15 May 2025 18:19	vial-A	SY/MD	Ok
22	Q1982-05	TP-5	VY022273.D	15 May 2025 18:43	vial-A	SY/MD	Ok
23	Q1982-06	TP-6	VY022274.D	15 May 2025 19:06	vial-A	SY/MD	Ok
24	VSTDCCC050	VSTDCCC050EC	VY022275.D	15 May 2025 19:52		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY051625

Review By	Amit Patel	Review On	5/19/2025 1:35:43 PM
Supervise By	Maresh Dadoda	Supervise On	5/19/2025 2:39:18 PM
SubDirectory	VY051625	HP Acquire Method	MSVOA_Y HP Processing Method 82y051525s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133936
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133942,VP133943 VP133934

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022276.D	16 May 2025 07:54		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY022277.D	16 May 2025 08:24		SY/MD	Ok,M
3	VY0516SBL01	VY0516SBL01	VY022278.D	16 May 2025 09:37		SY/MD	Ok
4	VY0516SBS01	VY0516SBS01	VY022279.D	16 May 2025 10:09		SY/MD	Ok,M
5	VY0516SBSD01	VY0516SBSD01	VY022280.D	16 May 2025 10:32		SY/MD	Ok,M
6	Q1984-07RE	OU4-TS-24-050725RE	VY022281.D	16 May 2025 11:17	vial-B Surrogate fail	SY/MD	Confirms
7	Q1984-09	OU4-TS-25-050725	VY022282.D	16 May 2025 11:41	vial-B	SY/MD	Ok
8	Q1984-11	OU4-TS-26-050725	VY022283.D	16 May 2025 12:04	vial-B	SY/MD	Ok
9	Q1984-13	OU4-TS-27-050725	VY022284.D	16 May 2025 12:29	vial-B	SY/MD	Ok
10	Q1984-15RE	OU4-TS-28-050725RE	VY022285.D	16 May 2025 12:54	vial-B Surrogate fail	SY/MD	Confirms
11	Q1982-01	TP-1	VY022286.D	16 May 2025 13:17	vial-B	SY/MD	Ok
12	Q1982-02	TP-2B	VY022287.D	16 May 2025 13:41	vial-B	SY/MD	Ok
13	Q1982-03	TP-3	VY022288.D	16 May 2025 14:04	Vial A	SY/MD	Ok
14	Q1982-07	TP-8	VY022289.D	16 May 2025 14:28	Vial A	SY/MD	Ok
15	Q1982-08	TP-9	VY022290.D	16 May 2025 14:51	Vial A	SY/MD	Ok
16	Q2010-03	TP-14	VY022291.D	16 May 2025 15:15	vial-B	SY/MD	Ok
17	Q2010-04	TP-17	VY022292.D	16 May 2025 15:38	vial-B	SY/MD	Ok
18	Q2003-02	VOC	VY022293.D	16 May 2025 16:02	vial-A Internal standard fail	SY/MD	ReRun

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY051625

Review By	Amit Patel	Review On	5/19/2025 1:35:43 PM		
Supervise By	Mahesh Dadoda	Supervise On	5/19/2025 2:39:18 PM		
SubDirectory	VY051625	HP Acquire Method	MSVOA_Y	HP Processing Method	82y051525s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP133936			
CCC		VP133942,VP133943			
Internal Standard/PEM		VP133934			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q2032-03	TP-29-99	VY022294.D	16 May 2025 16:25	vial-A	SY/MD	Ok
20	Q2032-04	TP-24	VY022295.D	16 May 2025 16:48	vial-A	SY/MD	Ok
21	Q2032-07	TP-37	VY022296.D	16 May 2025 17:12	vial-A	SY/MD	Ok
22	Q2032-08	TP-32	VY022297.D	16 May 2025 17:35	vial-A	SY/MD	Ok
23	Q1984-19	OU4-TB01-050725	VY022298.D	16 May 2025 17:59	vial-A	SY/MD	Ok
24	Q2032-05MS	TP-24MS	VY022299.D	16 May 2025 18:21	vial-A Internal standard fail;Recovery Fail	SY/MD	Not Ok
25	Q2032-06MSD	TP-24MSD	VY022300.D	16 May 2025 18:44	vial-A Internal standard fail;Surrogate fail	SY/MD	Not Ok
26	VSTDCCC050	VSTDCCC050EC	VY022301.D	16 May 2025 19:29		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY052025

Review By	Maresh Dadoda	Review On	5/21/2025 3:43:30 PM
Supervise By	Semsettin Yesilyurt	Supervise On	5/23/2025 2:22:55 AM
SubDirectory	VY052025	HP Acquire Method	MSVOA_Y HP Processing Method 82y051525s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133969
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133970,VP133971 VP133934

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022328.D	20 May 2025 08:09		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY022329.D	20 May 2025 08:40		SY/MD	Ok,M
3	VY0520SBL01	VY0520SBL01	VY022330.D	20 May 2025 09:48		SY/MD	Ok
4	VY0520SBS01	VY0520SBS01	VY022331.D	20 May 2025 10:21		SY/MD	Ok,M
5	VY0520SBSD01	VY0520SBSD01	VY022332.D	20 May 2025 10:43		SY/MD	Ok,M
6	Q2062-19	L1-WC-5-VOC	VY022333.D	20 May 2025 11:28	vial-B	SY/MD	Ok
7	Q2062-23	L1-WC-6-VOC	VY022334.D	20 May 2025 11:51	vial-B	SY/MD	Ok
8	Q2071-11	L1-WC-9-VOC	VY022335.D	20 May 2025 12:15	vial-B	SY/MD	Ok
9	Q2056-02	GVD12	VY022336.D	20 May 2025 12:38	vial-B	SY/MD	Ok
10	Q2056-03	GVD13	VY022337.D	20 May 2025 13:01	vial-B	SY/MD	Ok
11	Q2056-04	GVD14	VY022338.D	20 May 2025 13:25	vial-A	SY/MD	Ok
12	Q2056-05	GVD15	VY022339.D	20 May 2025 13:48	vial-A Internal standard fail	SY/MD	Not Ok
13	Q2080-02	PL-HRH-VOC-01	VY022340.D	20 May 2025 14:12	vial-A	SY/MD	Ok
14	Q2080-08	PL-HRH-VOC-02	VY022341.D	20 May 2025 14:35	vial-A	SY/MD	Ok
15	Q2032-02	TP-29	VY022342.D	20 May 2025 14:59	vial-A	SY/MD	Ok
16	Q2075-01	SS-10	VY022343.D	20 May 2025 15:22	vial-A	SY/MD	Ok
17	Q2075-02	SS-910	VY022344.D	20 May 2025 15:45	vial-A	SY/MD	Ok
18	Q2075-03	SS-11	VY022345.D	20 May 2025 16:09	vial-A	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY052025

Review By	Mahesh Dadoda	Review On	5/21/2025 3:43:30 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	5/23/2025 2:22:55 AM		
SubDirectory	VY052025	HP Acquire Method	MSVOA_Y	HP Processing Method	82y051525s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP133969			
CCC		VP133970,VP133971			
Internal Standard/PEM		VP133934			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q2084-02	OR-640-VOC-66	VY022346.D	20 May 2025 16:32	vial-A	SY/MD	Ok
20	Q2032-01	TP-11	VY022347.D	20 May 2025 16:56	vial-A Surrogate fail;Need MeOH	SY/MD	Dilution
21	Q2032-05MS	TP-24MS	VY022348.D	20 May 2025 17:19	vial-B	SY/MD	Ok,M
22	Q2032-06MSD	TP-24MSD	VY022349.D	20 May 2025 17:41	vial-B Internal standard fail	SY/MD	Ok,M
23	Q2075-04MS	SS-11MS	VY022350.D	20 May 2025 18:04	vial-A	SY/MD	Ok,M
24	Q2075-05MSD	SS-11MSD	VY022351.D	20 May 2025 18:26	vial-A	SY/MD	Ok,M
25	VSTDCCC050	VSTDCCC050EC	VY022352.D	20 May 2025 19:34	ENDCCC failed for com.#06	SY/MD	Ok,M

M : Manual Integration