

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02
 Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032
 Lab File ID: VN086722.D Date Analyzed: 05/21/2025
 Instrument ID: MSVOA_N Time Analyzed: 11:10
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	182299	8.22	331408	9.10	299657	11.87
UPPER LIMIT	364598	8.724	662816	9.6	599314	12.365
LOWER LIMIT	91149.5	7.724	165704	8.6	149829	11.365
EPA SAMPLE NO.						
TP-11ME	228680	8.22	425329	9.09	398917	11.87
VN0521MBL01	158233	8.22	315437	9.10	297842	11.87
VN0521MBS01	187953	8.22	345557	9.10	296696	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.