

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02
Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032
Lab File ID: VY022277.D Date Analyzed: 05/16/2025
Instrument ID: MSVOA_Y Time Analyzed: 08:24
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	212993	7.71	356583	8.62	305829	11.42
UPPER LIMIT	425986	8.207	713166	9.115	611658	11.92
LOWER LIMIT	106497	7.207	178292	8.115	152915	10.92
EPA SAMPLE NO.						
TP-29-99	265101	7.71	475372	8.61	384945	11.41
TP-24	265542	7.71	477294	8.62	397008	11.41
TP-37	264744	7.71	474666	8.61	382466	11.41
TP-32	262305	7.71	463101	8.61	367604	11.41
VY0516SBL01	314347	7.71	555500	8.62	426459	11.42
VY0516SBS01	205076	7.71	351708	8.62	300126	11.42

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.