

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02  
 Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032  
 Lab File ID: VN086722.D Date Analyzed: 05/21/2025  
 Instrument ID: MSVOA\_N Time Analyzed: 11:10  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

|                | IS4<br>AREA # | RT #   |  |  |  |  |
|----------------|---------------|--------|--|--|--|--|
| 12 HOUR STD    | 144979        | 13.788 |  |  |  |  |
| UPPER LIMIT    | 289958        | 14.288 |  |  |  |  |
| LOWER LIMIT    | 72489.5       | 13.288 |  |  |  |  |
| EPA SAMPLE NO. |               |        |  |  |  |  |
| TP-11ME        | 172654        | 13.79  |  |  |  |  |
| VN0521MBL01    | 126400        | 13.79  |  |  |  |  |
| VN0521MBS01    | 133846        | 13.79  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.