

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02
 Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032
 Lab File ID: VX046200.D Date Analyzed: 05/15/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:52
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	62876	12.018				
UPPER LIMIT	125752	12.518				
LOWER LIMIT	31438	11.518				
EPA SAMPLE NO.						
FB-05132025	46318	12.02				
TB	46824	12.02				
VX0515WBL01	51360	12.02				
VX0515WBS01	65777	12.02				
VX0515WBSD01	63406	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.