

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: CAMP02
Lab Code: CHEM Case No.: Q2032 SAS No.: Q2032 SDG NO.: Q2032
Lab File ID: VY022276.D BFB Injection Date: 05/16/2025
Instrument ID: MSVOA_Y BFB Injection Time: 07:54
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	25.1
75	30.0 - 60.0% of mass 95	58.3
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.9
173	Less than 2.0% of mass 174	0.9 (1) 1
174	50.0 - 100.0% of mass 95	83.3
175	5.0 - 9.0% of mass 174	6.2 (7.4) 1
176	95.0 - 101.0% of mass 174	81.9 (98.3) 1
177	5.0 - 9.0% of mass 176	5.6 (6.9) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VY022277.D	05/16/2025	08:24
VY0516SBL01	VY0516SBL01	VY022278.D	05/16/2025	09:37
VY0516SBS01	VY0516SBS01	VY022279.D	05/16/2025	10:09
TP-29-99	Q2032-03	VY022294.D	05/16/2025	16:25
TP-24	Q2032-04	VY022295.D	05/16/2025	16:48
TP-37	Q2032-07	VY022296.D	05/16/2025	17:12
TP-32	Q2032-08	VY022297.D	05/16/2025	17:35