

**Analysis Method:** 9040C  
**Parameter:** pH  
**Run Number:** LB135781

**Analyst By :** jignesh

**Supervisor Review By :** Iwona

**Slope :** 99.2

**pH Meter ID :** WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER, COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.05].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	05/15/2025	08:15
2	CAL2	1	Water	NA	NA	20.2	7.00	05/15/2025	08:16
3	CAL3	1	Water	NA	NA	20.2	10.02	05/15/2025	08:18
4	ICV	1	Water	NA	NA	20.2	7.02	05/15/2025	08:20
5	CCV1	1	Water	NA	NA	20.1	2.01	05/15/2025	08:22
6	Q2033-01	1	Water	NA	NA	21.4	12.32	05/15/2025	08:30
7	Q2033-01DUP	1	Water	NA	NA	21.5	12.33	05/15/2025	08:31
8	Q2043-05	1	Water	NA	NA	21.7	7.01	05/15/2025	08:45
9	Q2043-05DUP	1	Water	NA	NA	21.8	7.02	05/15/2025	08:46
10	Q2044-02	1	Water	NA	NA	21.7	7.80	05/15/2025	08:55
11	Q2044-02DUP	1	Water	NA	NA	21.8	7.81	05/15/2025	08:56
12	CCV2	1	Water	NA	NA	20.2	12.02	05/15/2025	09:00

# WORKLIST(Hardcopy Internal Chain)

LB 135781

WorkList Name : PH W Q2044

WorkList ID : 189525

Department : Wet-Chemistry

Date : 05-15-2025 07:59:19

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2033-01	D TW-WTS-08	Water	pH	Cool 4 deg C	ENTA05	L41	05/12/2025	9040C
Q2043-05	C 3139	Water	pH	Cool 4 deg C	PSEG03	L21	05/14/2025	9040C
Q2044-02	A 3859	Water	pH	Cool 4 deg C	PSEG03	L31	05/14/2025	9040C

Date/Time 05-15-25 08:10

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 05-15-25 11:30

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]