

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q2033

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135762

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1						
Bromide	mg/L	10.2	10	102	90-110	04/22/2025
Chloride	mg/L	3.1	3	103	90-110	04/22/2025
Fluoride	mg/L	2	2	100	90-110	04/22/2025
Nitrite	mg/L	3.1	3	103	90-110	04/22/2025
Nitrate	mg/L	2.6	2.5	104	90-110	04/22/2025
Sulfate	mg/L	15.1	15	101	90-110	04/22/2025
Orthophosphate as P	mg/L	5.2	5	104	90-110	04/22/2025
Sample ID: CCV1						
Bromide	mg/L	10.1	10	101	90-110	05/14/2025
Chloride	mg/L	3.1	3	103	90-110	05/14/2025
Fluoride	mg/L	2	2	100	90-110	05/14/2025
Nitrite	mg/L	3	3	100	90-110	05/14/2025
Nitrate	mg/L	2.5	2.5	100	90-110	05/14/2025
Sulfate	mg/L	14.6	15	97	90-110	05/14/2025
Orthophosphate as P	mg/L	4.9	5	98	90-110	05/14/2025
Sample ID: CCV2						
Bromide	mg/L	10	10	100	90-110	05/14/2025
Chloride	mg/L	3.1	3	103	90-110	05/14/2025
Fluoride	mg/L	2	2	100	90-110	05/14/2025
Nitrite	mg/L	3	3	100	90-110	05/14/2025
Nitrate	mg/L	2.5	2.5	100	90-110	05/14/2025
Sulfate	mg/L	14.7	15	98	90-110	05/14/2025
Orthophosphate as P	mg/L	5.1	5	102	90-110	05/14/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q2033

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135773

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.497	0.5	99	90-110	05/14/2025
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.497	0.5	99	90-110	05/14/2025
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.498	0.5	100	90-110	05/14/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q2033

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135781

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
pH		pH	7.02	7	100	90-110	05/15/2025
Sample ID:	CCV1						
pH		pH	2.01	2.00	101	90-110	05/15/2025
Sample ID:	CCV2						
pH		pH	12.02	12.00	100	90-110	05/15/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q2033

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135789

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV						
Flash Point	o F	82.1	81	101	78-84	05/15/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q2033

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135830

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 TKN	mg/L	4.9	5	98	90-110	05/19/2025
Sample ID: CCV1 TKN	mg/L	5	5	100	90-110	05/19/2025
Sample ID: CCV2 TKN	mg/L	5.1	5	102	90-110	05/19/2025
Sample ID: CCV3 TKN	mg/L	5.1	5	102	90-110	05/19/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q2033

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135830

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------