

Initial and Continuing Calibration Blank Summary

Client: ENTACT

SDG No.: Q2033

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135762

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	04/22/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	04/22/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	04/22/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	04/22/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	04/22/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	04/22/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	04/22/2025
Sample ID: CCB1							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	05/14/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	05/14/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	05/14/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	05/14/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	05/14/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	05/14/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	05/14/2025
Sample ID: CCB2							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	05/14/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	05/14/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	05/14/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	05/14/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	05/14/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	05/14/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	05/14/2025

Initial and Continuing Calibration Blank Summary

Client: ENTACT

SDG No.: Q2033

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135773

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0029	0.01	05/14/2025
Sample ID: CCB1 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0029	0.01	05/14/2025
Sample ID: CCB2 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0029	0.01	05/14/2025

Initial and Continuing Calibration Blank Summary

Client: ENTACT

SDG No.: Q2033

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135830

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 TKN	mg/L	< 0.2500	0.2500	U	0.11	0.5	05/19/2025
Sample ID: CCB1 TKN	mg/L	< 0.2500	0.2500	U	0.11	0.5	05/19/2025
Sample ID: CCB2 TKN	mg/L	< 0.2500	0.2500	U	0.11	0.5	05/19/2025
Sample ID: CCB3 TKN	mg/L	< 0.2500	0.2500	U	0.11	0.5	05/19/2025

Initial and Continuing Calibration Blank Summary

Client: ENTACT

SDG No.: Q2033

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135830

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
---------	-------	--------	----------------------	--------------	-----	-----	------------------