



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: ENTACT **SDG No.:** Q2033
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q2033 **SAS No.:** Q2033
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Cadmium	6.33	6.0	106	65 - 135	P	05/19/2025	12:48	LB135823
	Copper	22.7	20.0	114	65 - 135	P	05/19/2025	12:48	LB135823
	Lead	12.8	12.0	107	65 - 135	P	05/19/2025	12:48	LB135823
	Nickel	40.0	40.0	100	65 - 135	P	05/19/2025	12:48	LB135823
	Zinc	42.3	40.0	106	65 - 135	P	05/19/2025	12:48	LB135823
CRA	Mercury	0.19	0.2	97	70 - 130	CV	05/20/2025	09:47	LB135832