

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: ENTACT **Level:** LOW **SDG No.:** Q2033
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q2033 **SAS No.:** Q2033
Matrix: Water **Sample ID:** Q2033-01 **Client ID:** TW-WTS-08DUP
Percent Solids for Sample: NA **Duplicate ID** Q2033-01DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cadmium	ug/L	20	3.00	U	3.00	U			P
Copper	ug/L	20	19.1		19.5		2		P
Lead	ug/L	20	4.28	J	4.38	J	2		P
Nickel	ug/L	20	18.1	J	18.9	J	4		P
Zinc	ug/L	20	8.68	J	20.0	U	200.0		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: ENTACT **Level:** LOW **SDG No.:** Q2033
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q2033 **SAS No.:** Q2033
Matrix: Water **Sample ID:** Q2033-01MS **Client ID:** TW-WTS-08MSD
Percent Solids for Sample: NA **Duplicate ID** Q2033-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cadmium	ug/L	20	99.9		101		1		P
Copper	ug/L	20	160		162		1		P
Lead	ug/L	20	481		484		1		P
Nickel	ug/L	20	261		263		1		P
Zinc	ug/L	20	107		108		1		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: ENTACT **Level:** LOW **SDG No.:** Q2033
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q2033 **SAS No.:** Q2033
Matrix: Water **Sample ID:** Q2064-01 **Client ID:** RW8-SP100-70-20250514DUP
Percent Solids for Sample: NA **Duplicate ID** Q2064-01DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	0.20	U	0.20	U			CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: ENTACT **Level:** LOW **SDG No.:** Q2033
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q2033 **SAS No.:** Q2033
Matrix: Water **Sample ID:** Q2064-01MS **Client ID:** RW8-SP100-70-20250514MSD
Percent Solids for Sample: NA **Duplicate ID** Q2064-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	3.25		3.71		13		CV