

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT **SDG No.:** Q2033
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q2033 **SAS No.:** Q2033

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168036BS							
Cadmium	ug/L	100	96.1		96	88 - 113	P
Copper	ug/L	150	147		98	86 - 114	P
Lead	ug/L	500	469		94	86 - 113	P
Nickel	ug/L	250	236		94	88 - 113	P
Zinc	ug/L	100	97.0		97	87 - 115	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT **SDG No.:** Q2033
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q2033 **SAS No.:** Q2033

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168071BS Mercury	ug/L	4.0	3.61		90	82 - 119	CV