

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2033 SAS No.: Q2033 SDG NO.: Q2033

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL01 Date Analyzed: 05/15/2025

Lab Sample No.: AR1660CCC500 Data File : PP072076.D Time Analyzed: 04:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.656	5.571	5.771	435.980	500.000	-12.8
Aroclor-1016-2	5.678	5.593	5.793	497.580	500.000	-0.5
Aroclor-1016-3	5.741	5.654	5.854	498.750	500.000	-0.3
Aroclor-1016-4	5.838	5.752	5.952	490.730	500.000	-1.9
Aroclor-1016-5	6.129	6.045	6.245	483.490	500.000	-3.3
Aroclor-1260-1	7.249	7.165	7.365	482.510	500.000	-3.5
Aroclor-1260-2	7.503	7.418	7.618	481.610	500.000	-3.7
Aroclor-1260-3	7.861	7.777	7.977	497.180	500.000	-0.6
Aroclor-1260-4	8.086	8.002	8.202	475.890	500.000	-4.8
Aroclor-1260-5	8.405	8.322	8.522	506.250	500.000	1.3
Decachlorobiphenyl	10.213	10.137	10.337	52.590	50.000	5.2
Tetrachloro-m-xylene	4.504	4.417	4.617	49.550	50.000	-0.9

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2033 SAS No.: Q2033 SDG NO.: Q2033

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL01 Date Analyzed: 05/15/2025

Lab Sample No.: AR1660CCC500 Data File : PP072076.D Time Analyzed: 04:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.884	4.798	4.998	550.030	500.000	10.0
Aroclor-1016-2	4.902	4.816	5.016	517.740	500.000	3.5
Aroclor-1016-3	5.079	4.993	5.193	530.110	500.000	6.0
Aroclor-1016-4	5.121	5.036	5.236	495.050	500.000	-1.0
Aroclor-1016-5	5.335	5.250	5.450	533.270	500.000	6.7
Aroclor-1260-1	6.368	6.285	6.485	547.160	500.000	9.4
Aroclor-1260-2	6.557	6.473	6.673	549.140	500.000	9.8
Aroclor-1260-3	6.710	6.627	6.827	544.890	500.000	9.0
Aroclor-1260-4	7.180	7.098	7.298	562.890	500.000	12.6
Aroclor-1260-5	7.422	7.339	7.539	557.850	500.000	11.6
Decachlorobiphenyl	8.826	8.749	8.949	56.210	50.000	12.4
Tetrachloro-m-xylene	3.801	3.712	3.912	52.980	50.000	6.0