

Cover Page

Order ID : Q2033

Project ID : 540 Degraw St, Brooklyn, NY - E9309

Client : ENTACT

Lab Sample Number

Q2033-01

Client Sample Number

TW-WTS-08

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 5/19/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q2033

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: PRATIK PATEL

Date: 05/19/2025

LAB CHRONICLE

OrderID:	Q2033	OrderDate:	5/13/2025 4:02:00 PM
Client:	ENTACT	Project:	540 Degraw St, Brooklyn, NY - E9309
Contact:	Jarod Stanfield	Location:	L41,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2033-01	TW-WTS-08	WATER	PCB	8082A	05/12/25	05/14/25	05/15/25	05/13/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q2033

Order ID: Q2033

Client: ENTACT

Project ID: 540 Degraw St, Brooklyn, NY - E9309

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units

Client ID :

Total Concentration: 0.000



QC SUMMARY

Surrogate Summary

SDG No.: Q2033

Client: ENTACT

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP071388.D	PIBLK-PP071388.D	Tetrachloro-m-xylene	1	20	17.1	85		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	17.6	88		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	17.2	86		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	17.6	88		70 (60)	130 (140)
I.BLK-PP072080.D	PIBLK-PP072080.D	Tetrachloro-m-xylene	1	20	16.7	83		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	18.4	92		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	19.4	97		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	20.9	104		70 (60)	130 (140)
PB168005BL	PB168005BL	Tetrachloro-m-xylene	1	20	18.9	94		30 (16)	150 (158)
		Decachlorobiphenyl	1	20	19.6	98		30 (10)	150 (173)
		Tetrachloro-m-xylene	2	20	21.0	105		30 (16)	150 (158)
		Decachlorobiphenyl	2	20	21.9	109		30 (10)	150 (173)
PB168005BS	PB168005BS	Tetrachloro-m-xylene	1	20	19.4	97		30 (16)	150 (158)
		Decachlorobiphenyl	1	20	20.7	103		30 (10)	150 (173)
		Tetrachloro-m-xylene	2	20	21.2	106		30 (16)	150 (158)
		Decachlorobiphenyl	2	20	23.0	115		30 (10)	150 (173)
PB168005BSD	PB168005BSD	Tetrachloro-m-xylene	1	20	19.2	96		30 (16)	150 (158)
		Decachlorobiphenyl	1	20	20.5	103		30 (10)	150 (173)
		Tetrachloro-m-xylene	2	20	21.0	105		30 (16)	150 (158)
		Decachlorobiphenyl	2	20	23.4	117		30 (10)	150 (173)
Q2033-01	TW-WTS-08	Tetrachloro-m-xylene	1	20	18.2	91		30 (16)	150 (158)
		Decachlorobiphenyl	1	20	13.3	66		30 (10)	150 (173)
		Tetrachloro-m-xylene	2	20	21.8	109		30 (16)	150 (158)
		Decachlorobiphenyl	2	20	15.4	77		30 (10)	150 (173)
I.BLK-PP072095.D	PIBLK-PP072095.D	Tetrachloro-m-xylene	1	20	16.9	85		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	17.9	90		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	19.2	96		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	21.0	105		70 (60)	130 (140)

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2033</u>	Analytical Method:	<u>8082A</u>
Client:	<u>ENTACT</u>	Datafile :	<u>PP072082.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB168005BS (Column 1)	AR1016	5	4.60	ug/L	92				40 (61)	140 (112)	
	AR1260	5	4.30	ug/L	86				40 (66)	140 (113)	
PB168005BS (Column 2)	AR1016	5	4.90	ug/L	98				40 (61)	140 (112)	
	AR1260	5	4.90	ug/L	98				40 (66)	140 (113)	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2033

Analytical Method: 8082A

Client: ENTACT

Datafile : PP072083.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB168005BSD (Column 1)	AR1016	5	4.50	ug/L	90	2			40 (61)	140 (112)	20 (20)
	AR1260	5	4.30	ug/L	86	0			40 (66)	140 (113)	20 (20)
PB168005BSD (Column 2)	AR1016	5	4.60	ug/L	92	2			40 (61)	140 (112)	20 (20)
	AR1260	5	4.80	ug/L	96	0			40 (66)	140 (113)	20 (20)

4C

PESTICIDE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB168005BL

Lab Name: CHEMTECH

Contract: ENTA05

Lab Code: CHEM Case No.: Q2033

SAS No.: Q2033 SDG NO.: Q2033

Lab Sample ID: PB168005BL

Lab File ID: PP072081.D

Matrix: (soil/water) WATER

Extraction: (Type) SEPF

Sulfur Cleanup: (Y/N) N

Date Extracted: 05/14/2025

Date Analyzed (1): 05/15/2025

Date Analyzed (2): 05/15/2025

Time Analyzed (1): 05:43

Time Analyzed (2): 05:43

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column (1): ZB-MR1 ID: 0.32 (mm)

GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
PB168005BS	PB168005BS	PP072082.D	05/15/2025	05/15/2025
PB168005BSD	PB168005BSD	PP072083.D	05/15/2025	05/15/2025
TW-WTS-08	Q2033-01	PP072090.D	05/15/2025	05/15/2025

COMMENTS: _____



SAMPLE DATA

Report of Analysis

Client:	ENTACT	Date Collected:	05/12/25
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	05/13/25
Client Sample ID:	TW-WTS-08	SDG No.:	Q2033
Lab Sample ID:	Q2033-01	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	970	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP072090.D	1	05/14/25 09:08	05/15/25 08:10	PB168005

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.10	U	0.10	0.52	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.52	ug/L
11141-16-5	Aroclor-1232	0.099	U	0.099	0.52	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.52	ug/L
12672-29-6	Aroclor-1248	0.073	U	0.073	0.52	ug/L
11097-69-1	Aroclor-1254	0.097	U	0.097	0.52	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.52	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.52	ug/L
11096-82-5	Aroclor-1260	0.084	U	0.084	0.52	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.8		30 (16) - 150 (158)	109%	SPK: 20
2051-24-3	Decachlorobiphenyl	15.4		30 (10) - 150 (173)	77%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051425\
 Data File : PP072090.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 May 2025 08:10
 Operator : YP\AJ
 Sample : Q2033-01 (Sig #1); Q2023-01 (Sig #2)
 Misc :
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 TW-WTS-08

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 15 09:01:46 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.505	3.803	36166069	30691128	18.241	21.788m
2) SA Decachlor...	10.213	8.825	19180170	13536551	13.246	15.434m

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051425\
Data File : PP072090.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 May 2025 08:10
Operator : YP\AJ
Sample : Q2033-01 (Sig #1); Q2023-01 (Sig #2)
Misc :
ALS Vial : 46 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
TW-WTS-08

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 15 09:01:46 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.505 min
Delta R.T.: -0.012 min
Response: 36166069
Conc: 18.24 ng/ml

Instrument :
ECD_P
ClientSampleId :
TW-WTS-08




#1 Tetrachloro-m-xylene

R.T.: 3.803 min
Delta R.T.: -0.010 min
Response: 30691128
Conc: 21.79 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.213 min
Delta R.T.: -0.024 min
Response: 19180170
Conc: 13.25 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.825 min
Delta R.T.: -0.024 min
Response: 13536551
Conc: 15.43 ng/ml m



CALIBRATION SUMMARY

RETENTION TIMES OF INITIAL CALIBRATION

Contract: ENTA05
Lab Code: CHEM **Case No.:** Q2033 **SAS No.:** Q2033 **SDG NO.:** Q2033
Instrument ID: ECD_P **Calibration Date(s):** 04/22/2025 04/22/2025
Calibration Times: 10:29 17:49

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:	RT 1000 = <u>PP071389.D</u>	RT 750 = <u>PP071390.D</u>
RT 500 = <u>PP071391.D</u>	RT 250 = <u>PP071392.D</u>	RT 050 = <u>PP071393.D</u>

COMPOUND		RT 1000	RT 750	RT 500	RT 250	RT 050	MEAN RT	RT WINDOW	
								FROM	TO
Aroclor-1016-1	(1)	5.66	5.67	5.67	5.67	5.67	5.67	5.57	5.77
Aroclor-1016-2	(2)	5.69	5.69	5.69	5.69	5.69	5.69	5.59	5.79
Aroclor-1016-3	(3)	5.75	5.75	5.75	5.75	5.75	5.75	5.65	5.85
Aroclor-1016-4	(4)	5.85	5.85	5.85	5.85	5.85	5.85	5.75	5.95
Aroclor-1016-5	(5)	6.14	6.14	6.15	6.14	6.14	6.14	6.04	6.24
Aroclor-1260-1	(1)	7.26	7.26	7.27	7.26	7.26	7.26	7.16	7.36
Aroclor-1260-2	(2)	7.51	7.51	7.52	7.52	7.52	7.52	7.42	7.62
Aroclor-1260-3	(3)	7.87	7.87	7.88	7.87	7.88	7.87	7.77	7.97
Aroclor-1260-4	(4)	8.10	8.10	8.10	8.10	8.10	8.10	8.00	8.20
Aroclor-1260-5	(5)	8.42	8.42	8.42	8.42	8.42	8.42	8.32	8.52
Decachlorobiphenyl		10.23	10.23	10.24	10.23	10.24	10.23	10.13	10.33
Tetrachloro-m-xylene		4.51	4.51	4.52	4.51	4.52	4.51	4.41	4.61
Aroclor-1242-1	(1)	5.67	5.67	5.67	5.67	5.67	5.67	5.57	5.77
Aroclor-1242-2	(2)	5.69	5.69	5.69	5.69	5.69	5.69	5.59	5.79
Aroclor-1242-3	(3)	5.75	5.75	5.75	5.75	5.75	5.75	5.65	5.85
Aroclor-1242-4	(4)	5.85	5.85	5.85	5.85	5.85	5.85	5.75	5.95
Aroclor-1242-5	(5)	6.58	6.58	6.58	6.58	6.58	6.58	6.48	6.68
Decachlorobiphenyl		10.23	10.24	10.24	10.23	10.24	10.24	10.14	10.34
Tetrachloro-m-xylene		4.51	4.52	4.51	4.51	4.52	4.51	4.41	4.61
Aroclor-1248-1	(1)	5.67	5.67	5.67	5.67	5.67	5.67	5.57	5.77
Aroclor-1248-2	(2)	5.94	5.94	5.94	5.94	5.94	5.94	5.84	6.04
Aroclor-1248-3	(3)	6.14	6.14	6.14	6.14	6.14	6.14	6.04	6.24
Aroclor-1248-4	(4)	6.54	6.54	6.54	6.54	6.54	6.54	6.44	6.64
Aroclor-1248-5	(5)	6.58	6.58	6.58	6.58	6.58	6.58	6.48	6.68
Decachlorobiphenyl		10.24	10.24	10.24	10.24	10.24	10.24	10.14	10.34
Tetrachloro-m-xylene		4.52	4.52	4.51	4.51	4.51	4.51	4.41	4.61
Aroclor-1254-1	(1)	6.52	6.52	6.52	6.52	6.52	6.52	6.42	6.62
Aroclor-1254-2	(2)	6.74	6.74	6.74	6.73	6.74	6.74	6.64	6.84
Aroclor-1254-3	(3)	7.10	7.10	7.10	7.10	7.10	7.10	7.00	7.20
Aroclor-1254-4	(4)	7.38	7.38	7.38	7.38	7.38	7.38	7.28	7.48
Aroclor-1254-5	(5)	7.80	7.80	7.80	7.80	7.80	7.80	7.70	7.90
Decachlorobiphenyl		10.24	10.24	10.24	10.23	10.24	10.24	10.14	10.34
Tetrachloro-m-xylene		4.52	4.52	4.52	4.51	4.51	4.51	4.41	4.61
Aroclor-1268-1	(1)	8.73	8.73	8.73	8.74	8.73	8.73	8.63	8.83
Aroclor-1268-2	(2)	8.83	8.83	8.83	8.83	8.83	8.83	8.73	8.93
Aroclor-1268-3	(3)	9.06	9.06	9.06	9.06	9.06	9.06	8.96	9.16
Aroclor-1268-4	(4)	9.48	9.48	9.48	9.48	9.48	9.48	9.38	9.58
Aroclor-1268-5	(5)	9.89	9.89	9.89	9.90	9.89	9.90	9.80	10.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	10.24	10.23	10.24	10.24	10.24	10.24	10.14	10.34
Tetrachloro-m-xylene	4.51	4.52	4.52	4.52	4.51	4.52	4.42	4.62

RETENTION TIMES OF INITIAL CALIBRATION

Contract:	<u>ENTA05</u>						
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q2033</u>	SAS No.:	<u>Q2033</u>	SDG NO.:	<u>Q2033</u>
Instrument ID:	<u>ECD_P</u>	Calibration Date(s):		<u>04/22/2025</u>	<u>04/22/2025</u>		
		Calibration Times:		<u>10:29</u>	<u>17:49</u>		

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID: RT 1000 = PP071389.D RT 750 = PP071390.D
RT 500 = PP071391.D RT 250 = PP071392.D RT 050 = PP071393.D

[illegible]

[illegible]

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: ENTA05
Lab Code: CHEM **Case No.:** Q2033 **SAS No.:** Q2033 **SDG NO.:** Q2033
Instrument ID: ECD_P
Calibration Date(s): 04/22/2025 04/22/2025
Calibration Times: 10:29 17:49

GC Column: ZB-MR1 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PP071389.D</u>		CF 750 = <u>PP071390.D</u>			
CF 500 = <u>PP071391.D</u>		CF 250 = <u>PP071392.D</u>		CF 050 = <u>PP071393.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	64709894	67652695	70083532	71128656	61674340	67049823
Aroclor-1016-2	(2)	98977911	102364897	107369074	108813892	95615840	102628323
Aroclor-1016-3	(3)	59905708	62127143	64296308	65849928	56056440	61647105
Aroclor-1016-4	(4)	50563747	52419789	53619076	54941428	44946700	51298148
Aroclor-1016-5	(5)	47772963	49344036	50670668	50888564	42515680	48238382
Aroclor-1260-1	(1)	90879847	94857525	98814986	99659744	95910940	96024608
Aroclor-1260-2	(2)	137163459	142386411	147498372	149887700	138450980	143077384
Aroclor-1260-3	(3)	110350751	114445481	117702804	126515524	91505860	112104084
Aroclor-1260-4	(4)	105570732	110780085	111211068	131260612	105501660	112864831
Aroclor-1260-5	(5)	220536830	229070229	228919678	255108180	199231240	226573231
Decachlorobiphenyl		1409816860	1467002307	1522583160	1539948480	1300700600	1448010281
Tetrachloro-m-xylene		1947183160	2005072840	2066603680	2100352680	1793966200	1982635712
Aroclor-1242-1	(1)	53870259	55684409	58494980	58676784	47130380	54771362
Aroclor-1242-2	(2)	81881116	85641493	90236470	89744200	77399340	84980524
Aroclor-1242-3	(3)	49423643	51294363	53309060	53836868	47752740	51123335
Aroclor-1242-4	(4)	41638031	42855355	45034468	44656444	36858500	42208560
Aroclor-1242-5	(5)	44955701	46392219	48330700	47956460	43817460	46290508
Decachlorobiphenyl		1368547670	1421241013	1488838040	1503303480	1179092600	1392204561
Tetrachloro-m-xylene		1868945350	1926494133	2017272340	2030390160	1613146400	1891249677
Aroclor-1248-1	(1)	42447397	43370463	44490672	45736348	38649560	42938888
Aroclor-1248-2	(2)	57932834	60215229	62023042	63285912	58705480	60432499
Aroclor-1248-3	(3)	65030901	68053485	69372090	70318056	61147600	66784426
Aroclor-1248-4	(4)	80777360	84441085	86910244	89142916	81782860	84610893
Aroclor-1248-5	(5)	76029464	78502729	80776474	83552224	78380620	79448302
Decachlorobiphenyl		1379130770	1440268240	1485619680	1521197800	1358343200	1436911938
Tetrachloro-m-xylene		1879191440	1925518227	1989640420	2037603040	1811949400	1928780505
Aroclor-1254-1	(1)	78574372	81246615	85577334	88598188	75759400	81951182
Aroclor-1254-2	(2)	121624435	125816099	132146328	137115708	118678260	127076166
Aroclor-1254-3	(3)	123432630	127799839	132981654	137623252	117036460	127774767
Aroclor-1254-4	(4)	112512019	116259037	121737436	124392856	112665620	117513394
Aroclor-1254-5	(5)	106399089	105922557	113785120	116153044	96391680	107730298
Decachlorobiphenyl		1410101810	1469202520	1526389740	1549706840	1328725400	1456825262
Tetrachloro-m-xylene		1888241120	1936990333	2012368600	2071565920	1722515200	1926336235
Aroclor-1268-1	(1)	299594266	309097848	319744804	331826088	282063800	308465361



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	251109850	259047004	267960354	275856908	228123960	256419615	7
Aroclor-1268-3	(3)	211915080	218354113	225284280	231505696	216470880	220706010	4
Aroclor-1268-4	(4)	93231632	96431025	99007040	98059984	82736300	93893196	7
Aroclor-1268-5	(5)	590082071	603740247	621062448	631335032	536707900	596585540	6
Decachlorobiphenyl		2486623660	2556709627	2669414960	2694667240	2263272000	2534137497	7
Tetrachloro-m-xylene		1883977950	1931567613	1982008260	2006991440	1590966000	1879102253	9

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: ENTA05
Lab Code: CHEM **Case No.:** Q2033 **SAS No.:** Q2033 **SDG NO.:** Q2033
Instrument ID: ECD_P **Calibration Date(s):** 04/22/2025 04/22/2025
Calibration Times: 10:29 17:49

GC Column: ZB-MR2 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PP071389.D</u>		CF 750 = <u>PP071390.D</u>			
CF 500 = <u>PP071391.D</u>		CF 250 = <u>PP071392.D</u>		CF 050 = <u>PP071393.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	48121828	49414308	53806128	56180596	60145020	53533576
Aroclor-1016-2	(2)	69736709	72096436	77213744	78520924	84250740	76363711
Aroclor-1016-3	(3)	38517048	40455577	43831736	45082616	43165080	42210411
Aroclor-1016-4	(4)	31206636	32814844	36123020	37095092	35190100	34485938
Aroclor-1016-5	(5)	39458687	41985771	45850608	46345748	48595640	44447291
Aroclor-1260-1	(1)	63219917	66123699	73727666	76350368	82074520	72299234
Aroclor-1260-2	(2)	77338708	77548395	89921718	91492956	101005300	87461415
Aroclor-1260-3	(3)	72062678	70944365	81698234	79344004	87888600	78387576
Aroclor-1260-4	(4)	57333188	56779915	65331750	68083052	68025140	63110609
Aroclor-1260-5	(5)	143809415	139161576	155842678	157512280	157560640	150777318
Decachlorobiphenyl		831147600	875337973	902319620	934998560	841460400	877052831
Tetrachloro-m-xylene		1380433570	1378340613	1415112020	1448341640	1420974200	1408640409
Aroclor-1242-1	(1)	42717527	44660211	47919806	49576672	46343800	46243603
Aroclor-1242-2	(2)	62346958	64901092	68295906	70279172	64290540	66022734
Aroclor-1242-3	(3)	34614031	35703872	38469574	40739776	32714120	36448275
Aroclor-1242-4	(4)	33518120	34659665	37548334	39683016	33143760	35710579
Aroclor-1242-5	(5)	41214500	42688541	44971456	47258656	41407120	43508055
Decachlorobiphenyl		815615080	851726400	897002660	917481280	766770800	849719244
Tetrachloro-m-xylene		1295159620	1325412680	1503876340	1432488040	1389045200	1389196376
Aroclor-1248-1	(1)	30996932	35945815	36538014	39256084	37796080	36106585
Aroclor-1248-2	(2)	43624561	47309389	48616518	53505512	51756020	48962400
Aroclor-1248-3	(3)	45791659	49899927	50774174	56002176	52959580	51085503
Aroclor-1248-4	(4)	53871444	58562581	58680784	64889280	67282380	60657294
Aroclor-1248-5	(5)	52150607	56290108	57608714	65271236	60133720	58290877
Decachlorobiphenyl		826375100	860756013	900400040	940375840	921771600	889935719
Tetrachloro-m-xylene		1324712230	1514156880	1548203640	1434858960	1416930600	1447772462
Aroclor-1254-1	(1)	81690839	87361264	89755404	97285536	100908960	91400401
Aroclor-1254-2	(2)	70045476	75469849	77317914	83154376	87573280	78712179
Aroclor-1254-3	(3)	109725744	114103376	120096818	124380164	124235440	118508308
Aroclor-1254-4	(4)	70169900	74068720	77810162	81877428	81901880	77165618
Aroclor-1254-5	(5)	92229111	96109599	104690794	107384400	106924600	101467701
Decachlorobiphenyl		846307500	875335867	916104900	955261880	868223800	892246789
Tetrachloro-m-xylene		1390553300	1501013893	1450901500	1518608160	1468503000	1465915971
Aroclor-1268-1	(1)	181860271	188460193	203483780	205414332	201904640	196224643



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	154801411	159518377	172755008	173272408	170561360	166181713	5
Aroclor-1268-3	(3)	128491160	133089205	143083692	144773184	135469120	136981272	5
Aroclor-1268-4	(4)	56964280	59695839	63521398	64177792	55793320	60030526	6
Aroclor-1268-5	(5)	367021219	377695863	385175068	387479012	342861840	372046600	5
Decachlorobiphenyl		1499051200	1546040093	1602790460	1652739760	1464953600	1553115023	5
Tetrachloro-m-xylene		1476470990	1445952813	1464214380	1457737600	1300882000	1429051557	5



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: ENTA05

Lab Code: CHEM Case No.: Q2033 SAS No.: Q2033 SDG NO.: Q2033

Instrument ID: ECD_P Date(s) Analyzed: 04/22/2025 04/22/2025

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.72	4.62	4.82	24116800
		2	4.80	4.70	4.90	18016900
		3	4.88	4.78	4.98	56611200
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.88	4.78	4.98	44912000
		2	5.40	5.30	5.50	23013000
		3	5.69	5.59	5.79	47217600
		4	5.85	5.75	5.95	23179400
		5	5.94	5.84	6.04	16145700
Aroclor-1262	500	1	8.10	8.00	8.20	144491000
		2	8.42	8.32	8.52	274186000
		3	8.74	8.64	8.84	182878000
		4	8.82	8.72	8.92	134211000
		5	9.48	9.38	9.58	88162000



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: ENTA05

Lab Code: CHEM Case No.: Q2033 SAS No.: Q2033 SDG NO.: Q2033

Instrument ID: ECD_P Date(s) Analyzed: 04/22/2025 04/22/2025

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.02	3.92	4.12	21995000
		2	4.11	4.01	4.21	16561700
		3	4.19	4.09	4.29	48477000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.19	4.09	4.29	38317800
		2	4.92	4.82	5.02	38187400
		3	5.09	4.99	5.19	21175200
		4	5.18	5.08	5.28	18540400
		5	5.35	5.25	5.45	19901900
Aroclor-1262	500	1	6.94	6.84	7.04	113612000
		2	7.20	7.10	7.30	92339400
		3	7.72	7.62	7.82	75694400
		4	7.79	7.69	7.89	124826000
		5	8.29	8.19	8.39	55493200

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071389.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 10:29
 Operator : YP\AJ
 Sample : AR1660ICC1000
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 11:18:22 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 11:14:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.511	3.811	194.7E6	138.0E6	97.025	98.760
2) SA Decachlor...	10.230	8.849	141.0E6	83114760	96.154	95.894
Target Compounds						
3) L1 AR-1016-1	5.664	4.897	64709894	48121828	960.134	944.232
4) L1 AR-1016-2	5.686	4.915	98977911	69736709	959.335	949.119
5) L1 AR-1016-3	5.749	5.092	59905708	38517048	964.650	935.461
6) L1 AR-1016-4	5.846	5.134	50563747	31206636	970.673	926.980
7) L1 AR-1016-5	6.139	5.349	47772963	39458687	970.565	925.074
31) L7 AR-1260-1	7.259	6.384	90879847	63219917	958.169	923.272
32) L7 AR-1260-2	7.512	6.572	137.2E6	77338708	963.694	924.770
33) L7 AR-1260-3	7.872	6.726	110.4E6	72062678	967.762	937.334
34) L7 AR-1260-4	8.095	7.197	105.6E6	57333188	973.982	934.793
35) L7 AR-1260-5	8.416	7.439	220.5E6	143.8E6	981.349	959.843



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071389.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 10:29
Operator : YP\AJ
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 11:18:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 11:14:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.511 min
Delta R.T.: 0.000 min
Response: 194718316
Conc: 97.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.811 min
Delta R.T.: 0.000 min
Response: 138043357
Conc: 98.76 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.230 min
Delta R.T.: 0.000 min
Response: 140981686
Conc: 96.15 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.849 min
Delta R.T.: 0.000 min
Response: 83114760
Conc: 95.89 ng/ml



#3 AR-1016-1


R.T.: 5.664 min
Delta R.T.: 0.000 min
Response: 64709894
Conc: 960.13 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#3 AR-1016-1

R.T.: 4.897 min
Delta R.T.: 0.000 min
Response: 48121828
Conc: 944.23 ng/ml




#4 AR-1016-2

R.T.: 5.686 min
Delta R.T.: 0.000 min
Response: 98977911
Conc: 959.33 ng/ml



#4 AR-1016-2


R.T.: 4.915 min
Delta R.T.: 0.000 min
Response: 69736709
Conc: 949.12 ng/ml



#5 AR-1016-3


R.T.: 5.749 min
Delta R.T.: 0.000 min
Response: 59905708
Conc: 964.65 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#5 AR-1016-3

R.T.: 5.092 min
Delta R.T.: 0.000 min
Response: 38517048
Conc: 935.46 ng/ml




#6 AR-1016-4

R.T.: 5.846 min
Delta R.T.: 0.000 min
Response: 50563747
Conc: 970.67 ng/ml



#6 AR-1016-4


R.T.: 5.134 min
Delta R.T.: 0.000 min
Response: 31206636
Conc: 926.98 ng/ml



#7 AR-1016-5


R.T.: 6.139 min
Delta R.T.: 0.000 min
Response: 47772963
Conc: 970.56 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#7 AR-1016-5

R.T.: 5.349 min
Delta R.T.: 0.000 min
Response: 39458687
Conc: 925.07 ng/ml




#31 AR-1260-1

R.T.: 7.259 min
Delta R.T.: 0.000 min
Response: 90879847
Conc: 958.17 ng/ml



#31 AR-1260-1


R.T.: 6.384 min
Delta R.T.: 0.000 min
Response: 63219917
Conc: 923.27 ng/ml



#32 AR-1260-2


R.T.: 7.512 min
Delta R.T.: 0.000 min
Response: 137163459
Conc: 963.69 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#32 AR-1260-2

R.T.: 6.572 min
Delta R.T.: 0.000 min
Response: 77338708
Conc: 924.77 ng/ml




#33 AR-1260-3

R.T.: 7.872 min
Delta R.T.: 0.000 min
Response: 110350751
Conc: 967.76 ng/ml



#33 AR-1260-3


R.T.: 6.726 min
Delta R.T.: 0.000 min
Response: 72062678
Conc: 937.33 ng/ml



#34 AR-1260-4


R.T.: 8.095 min
Delta R.T.: 0.000 min
Response: 105570732
Conc: 973.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#34 AR-1260-4

R.T.: 7.197 min
Delta R.T.: 0.000 min
Response: 57333188
Conc: 934.79 ng/ml



#35 AR-1260-5

R.T.: 8.416 min
Delta R.T.: 0.000 min
Response: 220536830
Conc: 981.35 ng/ml



#35 AR-1260-5

R.T.: 7.439 min
Delta R.T.: 0.000 min
Response: 143809415
Conc: 959.84 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071390.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 10:45
 Operator : YP\AJ
 Sample : AR1660ICC750
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 11:20:57 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 11:14:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.811	150.4E6	103.4E6	74.955	74.302
2) SA Decachlor...	10.233	8.848	110.0E6	65650348	75.027	75.495
Target Compounds						
3) L1 AR-1016-1	5.668	4.897	50739521	37060731	751.897	734.641
4) L1 AR-1016-2	5.690	4.915	76773673	54072327	746.071	740.558
5) L1 AR-1016-3	5.752	5.092	46595357	30341683	750.210	741.220
6) L1 AR-1016-4	5.849	5.134	39314842	24611133	753.145	737.269
7) L1 AR-1016-5	6.142	5.349	37008027	31489328	751.240	742.118
31) L7 AR-1260-1	7.261	6.383	71143144	49592774	750.053	732.641
32) L7 AR-1260-2	7.514	6.572	106.8E6	58161296	750.195	712.735
33) L7 AR-1260-3	7.873	6.725	85834111	53208274	751.834	710.374
34) L7 AR-1260-4	8.097	7.197	83085064	42584936	760.941	711.945
35) L7 AR-1260-5	8.418	7.438	171.8E6	104.4E6	759.599	713.546



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071390.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 10:45
Operator : YP\AJ
Sample : AR1660ICC750
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 11:20:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 11:14:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.514 min
Delta R.T.: 0.000 min
Response: 150380463
Conc: 74.95 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.811 min
Delta R.T.: 0.000 min
Response: 103375546
Conc: 74.30 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.233 min
Delta R.T.: 0.000 min
Response: 110025173
Conc: 75.03 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.848 min
Delta R.T.: 0.000 min
Response: 65650348
Conc: 75.49 ng/ml



#3 AR-1016-1


R.T.: 5.668 min
Delta R.T.: 0.000 min
Response: 50739521
Conc: 751.90 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#3 AR-1016-1

R.T.: 4.897 min
Delta R.T.: 0.000 min
Response: 37060731
Conc: 734.64 ng/ml




#4 AR-1016-2

R.T.: 5.690 min
Delta R.T.: 0.000 min
Response: 76773673
Conc: 746.07 ng/ml



#4 AR-1016-2


R.T.: 4.915 min
Delta R.T.: 0.000 min
Response: 54072327
Conc: 740.56 ng/ml



#5 AR-1016-3


R.T.: 5.752 min
Delta R.T.: 0.000 min
Response: 46595357
Conc: 750.21 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#5 AR-1016-3

R.T.: 5.092 min
Delta R.T.: 0.000 min
Response: 30341683
Conc: 741.22 ng/ml




#6 AR-1016-4

R.T.: 5.849 min
Delta R.T.: 0.000 min
Response: 39314842
Conc: 753.15 ng/ml



#6 AR-1016-4


R.T.: 5.134 min
Delta R.T.: 0.000 min
Response: 24611133
Conc: 737.27 ng/ml



#7 AR-1016-5


R.T.: 6.142 min
Delta R.T.: 0.000 min
Response: 37008027
Conc: 751.24 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#7 AR-1016-5

R.T.: 5.349 min
Delta R.T.: 0.000 min
Response: 31489328
Conc: 742.12 ng/ml




#31 AR-1260-1

R.T.: 7.261 min
Delta R.T.: 0.000 min
Response: 71143144
Conc: 750.05 ng/ml



#31 AR-1260-1


R.T.: 6.383 min
Delta R.T.: 0.000 min
Response: 49592774
Conc: 732.64 ng/ml



#32 AR-1260-2


R.T.: 7.514 min
Delta R.T.: 0.000 min
Response: 106789808
Conc: 750.19 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#32 AR-1260-2

R.T.: 6.572 min
Delta R.T.: 0.000 min
Response: 58161296
Conc: 712.74 ng/ml




#33 AR-1260-3

R.T.: 7.873 min
Delta R.T.: 0.000 min
Response: 85834111
Conc: 751.83 ng/ml



#33 AR-1260-3


R.T.: 6.725 min
Delta R.T.: 0.000 min
Response: 53208274
Conc: 710.37 ng/ml



#34 AR-1260-4


R.T.: 8.097 min
Delta R.T.: 0.000 min
Response: 83085064
Conc: 760.94 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#34 AR-1260-4

R.T.: 7.197 min
Delta R.T.: 0.000 min
Response: 42584936
Conc: 711.94 ng/ml



#35 AR-1260-5

R.T.: 8.418 min
Delta R.T.: 0.000 min
Response: 171802672
Conc: 759.60 ng/ml



#35 AR-1260-5

R.T.: 7.438 min
Delta R.T.: 0.000 min
Response: 104371182
Conc: 713.55 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071391.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 11:02
 Operator : YP\AJ
 Sample : AR1660ICC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 11:15:13 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 11:14:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.517	3.812	103.3E6	70755601	50.000	50.000
2) SA Decachlor...	10.237	8.849	76129158	45115981	50.000	50.000
Target Compounds						
3) L1 AR-1016-1	5.671	4.898	35041766	26903064	500.000	500.000
4) L1 AR-1016-2	5.693	4.916	53684537	38606872	500.000	500.000
5) L1 AR-1016-3	5.754	5.093	32148154	21915868	500.000	500.000
6) L1 AR-1016-4	5.852	5.136	26809538	18061510	500.000	500.000
7) L1 AR-1016-5	6.145	5.350	25335334	22925304	500.000	500.000
31) L7 AR-1260-1	7.265	6.385	49407493	36863833	500.000	500.000
32) L7 AR-1260-2	7.518	6.573	73749186	44960859	500.000	500.000
33) L7 AR-1260-3	7.877	6.727	58851402	40849117	500.000	500.000
34) L7 AR-1260-4	8.102	7.198	55605534	32665875	500.000	500.000
35) L7 AR-1260-5	8.422	7.439	114.5E6	77921339	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071391.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 11:02
Operator : YP\AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 11:15:13 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 11:14:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.517 min
Delta R.T.: 0.000 min
Response: 103330184
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.812 min
Delta R.T.: 0.000 min
Response: 70755601
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.237 min
Delta R.T.: 0.000 min
Response: 76129158
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.849 min
Delta R.T.: 0.000 min
Response: 45115981
Conc: 50.00 ng/ml



#3 AR-1016-1


R.T.: 5.671 min
Delta R.T.: 0.000 min
Response: 35041766
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500




#3 AR-1016-1

R.T.: 4.898 min
Delta R.T.: 0.000 min
Response: 26903064
Conc: 500.00 ng/ml




#4 AR-1016-2

R.T.: 5.693 min
Delta R.T.: 0.000 min
Response: 53684537
Conc: 500.00 ng/ml



#4 AR-1016-2


R.T.: 4.916 min
Delta R.T.: 0.000 min
Response: 38606872
Conc: 500.00 ng/ml



#5 AR-1016-3


R.T.: 5.754 min
Delta R.T.: 0.000 min
Response: 32148154
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500




#5 AR-1016-3

R.T.: 5.093 min
Delta R.T.: 0.000 min
Response: 21915868
Conc: 500.00 ng/ml




#6 AR-1016-4

R.T.: 5.852 min
Delta R.T.: 0.000 min
Response: 26809538
Conc: 500.00 ng/ml



#6 AR-1016-4


R.T.: 5.136 min
Delta R.T.: 0.000 min
Response: 18061510
Conc: 500.00 ng/ml



#7 AR-1016-5


R.T.: 6.145 min
Delta R.T.: 0.000 min
Response: 25335334
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500




#7 AR-1016-5

R.T.: 5.350 min
Delta R.T.: 0.000 min
Response: 22925304
Conc: 500.00 ng/ml




#31 AR-1260-1

R.T.: 7.265 min
Delta R.T.: 0.000 min
Response: 49407493
Conc: 500.00 ng/ml



#31 AR-1260-1


R.T.: 6.385 min
Delta R.T.: 0.000 min
Response: 36863833
Conc: 500.00 ng/ml



#32 AR-1260-2


R.T.: 7.518 min
Delta R.T.: 0.000 min
Response: 73749186
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500




#32 AR-1260-2

R.T.: 6.573 min
Delta R.T.: 0.000 min
Response: 44960859
Conc: 500.00 ng/ml




#33 AR-1260-3

R.T.: 7.877 min
Delta R.T.: 0.000 min
Response: 58851402
Conc: 500.00 ng/ml



#33 AR-1260-3


R.T.: 6.727 min
Delta R.T.: 0.000 min
Response: 40849117
Conc: 500.00 ng/ml



#34 AR-1260-4


R.T.: 8.102 min
Delta R.T.: 0.000 min
Response: 55605534
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500




#34 AR-1260-4

R.T.: 7.198 min
Delta R.T.: 0.000 min
Response: 32665875
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 8.422 min
Delta R.T.: 0.000 min
Response: 114459839
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 7.439 min
Delta R.T.: 0.000 min
Response: 77921339
Conc: 500.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071392.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 11:18
 Operator : YP\AJ
 Sample : AR1660ICC250
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 11:30:03 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 11:29:52 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.813	52508817	36208541	25.869	25.761
2) SA Decachlor...	10.233	8.850	38498712	23374964	25.928	26.384
Target Compounds						
3) L1 AR-1016-1	5.668	4.899	17782164	14045149	259.997	270.720
4) L1 AR-1016-2	5.689	4.917	27203473	19630231	260.616	263.876
5) L1 AR-1016-3	5.752	5.094	16462482	11270654	261.124	268.530
6) L1 AR-1016-4	5.849	5.136	13735357	9273773	259.716	270.294
7) L1 AR-1016-5	6.142	5.351	12722141	11586437	256.138	266.906
31) L7 AR-1260-1	7.261	6.386	24914936	19087592	259.387	273.244
32) L7 AR-1260-2	7.515	6.574	37471925	22873239	259.800	272.056
33) L7 AR-1260-3	7.873	6.727	31628881	19836001	269.748	260.958
34) L7 AR-1260-4	8.098	7.199	32815153	17020763	286.081	275.052
35) L7 AR-1260-5	8.417	7.441	63777045	39378070	273.242	264.138



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071392.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 11:18
Operator : YP\AJ
Sample : AR1660ICC250
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 11:30:03 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 11:29:52 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.514 min
Delta R.T.: 0.000 min
Response: 52508817
Conc: 25.87 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.813 min
Delta R.T.: 0.000 min
Response: 36208541
Conc: 25.76 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.233 min
Delta R.T.: 0.000 min
Response: 38498712
Conc: 25.93 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.850 min
Delta R.T.: 0.000 min
Response: 23374964
Conc: 26.38 ng/ml



#3 AR-1016-1


R.T.: 5.668 min
Delta R.T.: 0.000 min
Response: 17782164
Conc: 260.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#3 AR-1016-1

R.T.: 4.899 min
Delta R.T.: 0.000 min
Response: 14045149
Conc: 270.72 ng/ml




#4 AR-1016-2

R.T.: 5.689 min
Delta R.T.: 0.000 min
Response: 27203473
Conc: 260.62 ng/ml



#4 AR-1016-2


R.T.: 4.917 min
Delta R.T.: 0.000 min
Response: 19630231
Conc: 263.88 ng/ml



#5 AR-1016-3


R.T.: 5.752 min
Delta R.T.: 0.000 min
Response: 16462482
Conc: 261.12 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#5 AR-1016-3

R.T.: 5.094 min
Delta R.T.: 0.000 min
Response: 11270654
Conc: 268.53 ng/ml




#6 AR-1016-4

R.T.: 5.849 min
Delta R.T.: 0.000 min
Response: 13735357
Conc: 259.72 ng/ml



#6 AR-1016-4


R.T.: 5.136 min
Delta R.T.: 0.000 min
Response: 9273773
Conc: 270.29 ng/ml



#7 AR-1016-5


R.T.: 6.142 min
Delta R.T.: 0.000 min
Response: 12722141
Conc: 256.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#7 AR-1016-5

R.T.: 5.351 min
Delta R.T.: 0.000 min
Response: 11586437
Conc: 266.91 ng/ml




#31 AR-1260-1

R.T.: 7.261 min
Delta R.T.: 0.000 min
Response: 24914936
Conc: 259.39 ng/ml



#31 AR-1260-1


R.T.: 6.386 min
Delta R.T.: 0.000 min
Response: 19087592
Conc: 273.24 ng/ml



#32 AR-1260-2


R.T.: 7.515 min
Delta R.T.: 0.000 min
Response: 37471925
Conc: 259.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#32 AR-1260-2

R.T.: 6.574 min
Delta R.T.: 0.000 min
Response: 22873239
Conc: 272.06 ng/ml




#33 AR-1260-3

R.T.: 7.873 min
Delta R.T.: 0.000 min
Response: 31628881
Conc: 269.75 ng/ml



#33 AR-1260-3


R.T.: 6.727 min
Delta R.T.: 0.000 min
Response: 19836001
Conc: 260.96 ng/ml



#34 AR-1260-4


R.T.: 8.098 min
Delta R.T.: 0.000 min
Response: 32815153
Conc: 286.08 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#34 AR-1260-4

R.T.: 7.199 min
Delta R.T.: 0.000 min
Response: 17020763
Conc: 275.05 ng/ml



#35 AR-1260-5

R.T.: 8.417 min
Delta R.T.: 0.000 min
Response: 63777045
Conc: 273.24 ng/ml



#35 AR-1260-5

R.T.: 7.441 min
Delta R.T.: 0.000 min
Response: 39378070
Conc: 264.14 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071393.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 11:34
 Operator : YP\AJ
 Sample : AR1660ICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/23/2025
 Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 11:47:04 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 11:46:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.812	8969831	7104871	4.524	5.044
2) SA Decachlor...	10.235	8.850	6503503	4207302	4.491	4.797
Target Compounds						
3) L1 AR-1016-1	5.669	4.898	3083717	3007251	45.991	56.175
4) L1 AR-1016-2	5.691	4.916	4780792	4212537	46.584	55.164
5) L1 AR-1016-3	5.753	5.094	2802822	2158254	45.466	51.131
6) L1 AR-1016-4	5.851	5.135	2247335	1759505	43.809	51.021
7) L1 AR-1016-5	6.144	5.349	2125784	2429782	44.068	54.667
31) L7 AR-1260-1	7.262	6.384	4795547	4103726	49.941	56.760
32) L7 AR-1260-2	7.516	6.573	6922549	5050265	48.383	57.743
33) L7 AR-1260-3	7.875	6.726	4575293	4394430	40.810m	56.060 #
34) L7 AR-1260-4	8.098	7.197	5275083	3401257	46.807m	53.894
35) L7 AR-1260-5	8.417	7.439	9961562	7878032	43.897m	52.249

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071393.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 11:34
Operator : YP\AJ
Sample : AR1660ICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1660ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 11:47:04 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 11:46:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm





#1 Tetrachloro-m-xylene

R.T.: 4.515 min
Delta R.T.: 0.000 min
Response: 8969831
Conc: 4.52 ng/ml

Instrument :


ECD_P

ClientSampleId :

AR1660ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#1 Tetrachloro-m-xylene

R.T.: 3.812 min
Delta R.T.: 0.000 min
Response: 7104871
Conc: 5.04 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.235 min
Delta R.T.: 0.000 min
Response: 6503503
Conc: 4.49 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.850 min
Delta R.T.: 0.000 min
Response: 4207302
Conc: 4.80 ng/ml




#3 AR-1016-1

R.T.: 5.669 min
Delta R.T.: 0.000 min
Response: 3083717
Conc: 45.99 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#3 AR-1016-1

R.T.: 4.898 min
Delta R.T.: 0.000 min
Response: 3007251
Conc: 56.18 ng/ml




#4 AR-1016-2

R.T.: 5.691 min
Delta R.T.: 0.000 min
Response: 4780792
Conc: 46.58 ng/ml



#4 AR-1016-2

R.T.: 4.916 min
Delta R.T.: 0.000 min
Response: 4212537
Conc: 55.16 ng/ml




#5 AR-1016-3

R.T.: 5.753 min
Delta R.T.: 0.000 min
Response: 2802822
Conc: 45.47 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#5 AR-1016-3

R.T.: 5.094 min
Delta R.T.: 0.000 min
Response: 2158254
Conc: 51.13 ng/ml




#6 AR-1016-4

R.T.: 5.851 min
Delta R.T.: 0.000 min
Response: 2247335
Conc: 43.81 ng/ml



#6 AR-1016-4

R.T.: 5.135 min
Delta R.T.: 0.000 min
Response: 1759505
Conc: 51.02 ng/ml



#7 AR-1016-5

R.T.: 6.144 min
Delta R.T.: 0.000 min
Response: 2125784
Conc: 44.07 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025

#7 AR-1016-5


R.T.: 5.349 min
Delta R.T.: 0.000 min
Response: 2429782
Conc: 54.67 ng/ml

#31 AR-1260-1

R.T.: 7.262 min
Delta R.T.: 0.000 min
Response: 4795547
Conc: 49.94 ng/ml

#31 AR-1260-1

R.T.: 6.384 min
Delta R.T.: 0.000 min
Response: 4103726
Conc: 56.76 ng/ml




#32 AR-1260-2

R.T.: 7.516 min
Delta R.T.: 0.000 min
Response: 6922549
Conc: 48.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#32 AR-1260-2

R.T.: 6.573 min
Delta R.T.: 0.000 min
Response: 5050265
Conc: 57.74 ng/ml




#33 AR-1260-3

R.T.: 7.875 min
Delta R.T.: 0.000 min
Response: 4575293
Conc: 40.81 ng/ml m



#33 AR-1260-3

R.T.: 6.726 min
Delta R.T.: 0.000 min
Response: 4394430
Conc: 56.06 ng/ml




#34 AR-1260-4

R.T.: 8.098 min
Delta R.T.: 0.000 min
Response: 5275083
Conc: 46.81 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#34 AR-1260-4

R.T.: 7.197 min
Delta R.T.: 0.000 min
Response: 3401257
Conc: 53.89 ng/ml



#35 AR-1260-5

R.T.: 8.417 min
Delta R.T.: 0.000 min
Response: 9961562
Conc: 43.90 ng/ml m



#35 AR-1260-5

R.T.: 7.439 min
Delta R.T.: 0.000 min
Response: 7878032
Conc: 52.25 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071394.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 11:51
 Operator : YP\AJ
 Sample : AR1221ICC500
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1221ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 12:26:52 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 12:26:37 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.513	3.812	102.0E6	70886063	50.000	50.000
2) SA Decachlor...	10.233	8.849	73129811	43684216	50.000	50.000
Target Compounds						
8) L2 AR-1221-1	4.715	4.022	12058427	10997465	500.000	500.000
9) L2 AR-1221-2	4.800	4.109	9008471	8280856	500.000	500.000
10) L2 AR-1221-3	4.876	4.185	28305615	24238496	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071394.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 11:51
Operator : YP\AJ
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 12:26:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 12:26:37 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.513 min
Delta R.T.: 0.000 min
Response: 101953072
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.812 min
Delta R.T.: 0.000 min
Response: 70886063
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.233 min
Delta R.T.: 0.000 min
Response: 73129811
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.849 min
Delta R.T.: 0.000 min
Response: 43684216
Conc: 50.00 ng/ml



#8 AR-1221-1


R.T.: 4.715 min
Delta R.T.: 0.000 min
Response: 12058427
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500




#8 AR-1221-1

R.T.: 4.022 min
Delta R.T.: 0.000 min
Response: 10997465
Conc: 500.00 ng/ml




#9 AR-1221-2

R.T.: 4.800 min
Delta R.T.: 0.000 min
Response: 9008471
Conc: 500.00 ng/ml



#9 AR-1221-2


R.T.: 4.109 min
Delta R.T.: 0.000 min
Response: 8280856
Conc: 500.00 ng/ml



#10 AR-1221-3

R.T.: 4.876 min
Delta R.T.: 0.000 min
Response: 28305615
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500



#10 AR-1221-3

R.T.: 4.185 min
Delta R.T.: 0.000 min
Response: 24238496
Conc: 500.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071395.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 12:07
 Operator : YP\AJ
 Sample : AR1232ICC500
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1232ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 12:29:45 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 12:26:37 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.813	97587735	67524433	50.000	50.000
2) SA Decachlor...	10.233	8.851	71248479	42941843	50.000	50.000
Target Compounds						
11) L3 AR-1232-1	4.878	4.186	22456038	19158895	500.000	500.000
12) L3 AR-1232-2	5.403	4.917	11506547	19093695	500.000	500.000
13) L3 AR-1232-3	5.690	5.094	23608778	10587643	500.000	500.000
14) L3 AR-1232-4	5.850	5.178	11589674	9270206	500.000	500.000
15) L3 AR-1232-5	5.940	5.351	8072861	9950926	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071395.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 12:07
Operator : YP\AJ
Sample : AR1232ICC500
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 12:29:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 12:26:37 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.515 min
Delta R.T.: 0.000 min
Response: 97587735
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.813 min
Delta R.T.: 0.000 min
Response: 67524433
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.233 min
Delta R.T.: 0.000 min
Response: 71248479
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.851 min
Delta R.T.: 0.000 min
Response: 42941843
Conc: 50.00 ng/ml



#11 AR-1232-1


R.T.: 4.878 min
Delta R.T.: 0.000 min
Response: 22456038
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500




#11 AR-1232-1

R.T.: 4.186 min
Delta R.T.: 0.000 min
Response: 19158895
Conc: 500.00 ng/ml




#12 AR-1232-2

R.T.: 5.403 min
Delta R.T.: 0.000 min
Response: 11506547
Conc: 500.00 ng/ml



#12 AR-1232-2


R.T.: 4.917 min
Delta R.T.: 0.000 min
Response: 19093695
Conc: 500.00 ng/ml



#13 AR-1232-3


R.T.: 5.690 min
Delta R.T.: 0.000 min
Response: 23608778
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500




#13 AR-1232-3

R.T.: 5.094 min
Delta R.T.: 0.000 min
Response: 10587643
Conc: 500.00 ng/ml




#14 AR-1232-4

R.T.: 5.850 min
Delta R.T.: 0.000 min
Response: 11589674
Conc: 500.00 ng/ml



#14 AR-1232-4


R.T.: 5.178 min
Delta R.T.: 0.000 min
Response: 9270206
Conc: 500.00 ng/ml



#15 AR-1232-5

R.T.: 5.940 min
Delta R.T.: 0.000 min
Response: 8072861
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500



#15 AR-1232-5

R.T.: 5.351 min
Delta R.T.: 0.000 min
Response: 9950926
Conc: 500.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071396.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 12:23
 Operator : YP\AJ
 Sample : AR1242ICC1000
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 13:20:33 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 13:13:15 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.513	3.812	186.9E6	129.5E6	96.183	92.543
2) SA Decachlor...	10.233	8.849	136.9E6	81561508	95.790	95.248
Target Compounds						
16) L4 AR-1242-1	5.668	4.897	53870259	42717527	958.842	942.603
17) L4 AR-1242-2	5.689	4.916	81881116	62346958	951.456	954.464
18) L4 AR-1242-3	5.751	5.093	49423643	34614031	962.179	947.245
19) L4 AR-1242-4	5.848	5.177	41638031	33518120	960.813	943.289
20) L4 AR-1242-5	6.580	5.701	44955701	41214500	963.821	956.409



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071396.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 12:23
Operator : YP\AJ
Sample : AR1242ICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 13:20:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 13:13:15 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.513 min
Delta R.T.: 0.000 min
Response: 186894535
Conc: 96.18 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.812 min
Delta R.T.: 0.000 min
Response: 129515962
Conc: 92.54 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.233 min
Delta R.T.: 0.000 min
Response: 136854767
Conc: 95.79 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.849 min
Delta R.T.: 0.000 min
Response: 81561508
Conc: 95.25 ng/ml



#16 AR-1242-1


R.T.: 5.668 min
Delta R.T.: 0.000 min
Response: 53870259
Conc: 958.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000




#16 AR-1242-1

R.T.: 4.897 min
Delta R.T.: 0.000 min
Response: 42717527
Conc: 942.60 ng/ml




#17 AR-1242-2

R.T.: 5.689 min
Delta R.T.: 0.000 min
Response: 81881116
Conc: 951.46 ng/ml



#17 AR-1242-2


R.T.: 4.916 min
Delta R.T.: 0.000 min
Response: 62346958
Conc: 954.46 ng/ml



#18 AR-1242-3


R.T.: 5.751 min
Delta R.T.: 0.000 min
Response: 49423643
Conc: 962.18 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000




#18 AR-1242-3

R.T.: 5.093 min
Delta R.T.: 0.000 min
Response: 34614031
Conc: 947.24 ng/ml




#19 AR-1242-4

R.T.: 5.848 min
Delta R.T.: 0.000 min
Response: 41638031
Conc: 960.81 ng/ml



#19 AR-1242-4


R.T.: 5.177 min
Delta R.T.: 0.000 min
Response: 33518120
Conc: 943.29 ng/ml



#20 AR-1242-5

R.T.: 6.580 min
Delta R.T.: 0.000 min
Response: 44955701
Conc: 963.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000



#20 AR-1242-5

R.T.: 5.701 min
Delta R.T.: 0.000 min
Response: 41214500
Conc: 956.41 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071397.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 12:39
 Operator : YP\AJ
 Sample : AR1242ICC750
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 13:23:22 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 13:13:15 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.813	144.5E6	99405951	74.571	72.305
2) SA Decachlor...	10.236	8.850	106.6E6	63879480	74.739	74.732
Target Compounds						
16) L4 AR-1242-1	5.669	4.899	41763307	33495158	745.553	742.700
17) L4 AR-1242-2	5.691	4.917	64231120	48675819	747.572	746.776
18) L4 AR-1242-3	5.753	5.094	38470772	26777904	749.299	738.446
19) L4 AR-1242-4	5.850	5.179	32141516	25994749	744.431	737.606
20) L4 AR-1242-5	6.581	5.703	34794164	32016406	747.305	745.293



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071397.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 12:39
Operator : YP\AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 13:23:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 13:13:15 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.515 min
Delta R.T.: 0.000 min
Response: 144487060
Conc: 74.57 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.813 min
Delta R.T.: 0.000 min
Response: 99405951
Conc: 72.30 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.236 min
Delta R.T.: 0.000 min
Response: 106593076
Conc: 74.74 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.850 min
Delta R.T.: 0.000 min
Response: 63879480
Conc: 74.73 ng/ml



#16 AR-1242-1


R.T.: 5.669 min
Delta R.T.: 0.000 min
Response: 41763307
Conc: 745.55 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750




#16 AR-1242-1

R.T.: 4.899 min
Delta R.T.: 0.000 min
Response: 33495158
Conc: 742.70 ng/ml




#17 AR-1242-2

R.T.: 5.691 min
Delta R.T.: 0.000 min
Response: 64231120
Conc: 747.57 ng/ml



#17 AR-1242-2


R.T.: 4.917 min
Delta R.T.: 0.000 min
Response: 48675819
Conc: 746.78 ng/ml



#18 AR-1242-3


R.T.: 5.753 min
Delta R.T.: 0.000 min
Response: 38470772
Conc: 749.30 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750




#18 AR-1242-3

R.T.: 5.094 min
Delta R.T.: 0.000 min
Response: 26777904
Conc: 738.45 ng/ml




#19 AR-1242-4

R.T.: 5.850 min
Delta R.T.: 0.000 min
Response: 32141516
Conc: 744.43 ng/ml



#19 AR-1242-4


R.T.: 5.179 min
Delta R.T.: 0.000 min
Response: 25994749
Conc: 737.61 ng/ml



#20 AR-1242-5

R.T.: 6.581 min
Delta R.T.: 0.000 min
Response: 34794164
Conc: 747.30 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750



#20 AR-1242-5

R.T.: 5.703 min
Delta R.T.: 0.000 min
Response: 32016406
Conc: 745.29 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071398.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 12:56
 Operator : YP\AJ
 Sample : AR1242ICC500
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 13:13:34 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 13:13:15 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.812	100.9E6	75193817	50.000	50.000
2) SA Decachlor...	10.235	8.850	74441902	44850133	50.000	50.000
Target Compounds						
16) L4 AR-1242-1	5.669	4.898	29247490	23959903	500.000	500.000
17) L4 AR-1242-2	5.690	4.916	45118235	34147953	500.000	500.000
18) L4 AR-1242-3	5.752	5.093	26654530	19234787	500.000	500.000
19) L4 AR-1242-4	5.849	5.178	22517234	18774167	500.000	500.000
20) L4 AR-1242-5	6.581	5.701	24165350	22485728	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071398.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 12:56
Operator : YP\AJ
Sample : AR1242ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 13:13:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 13:13:15 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.514 min
Delta R.T.: 0.000 min
Response: 100863617
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.812 min
Delta R.T.: 0.000 min
Response: 75193817
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.235 min
Delta R.T.: 0.000 min
Response: 74441902
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.850 min
Delta R.T.: 0.000 min
Response: 44850133
Conc: 50.00 ng/ml



#16 AR-1242-1


R.T.: 5.669 min
Delta R.T.: 0.000 min
Response: 29247490
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500




#16 AR-1242-1

R.T.: 4.898 min
Delta R.T.: 0.000 min
Response: 23959903
Conc: 500.00 ng/ml




#17 AR-1242-2

R.T.: 5.690 min
Delta R.T.: 0.000 min
Response: 45118235
Conc: 500.00 ng/ml



#17 AR-1242-2


R.T.: 4.916 min
Delta R.T.: 0.000 min
Response: 34147953
Conc: 500.00 ng/ml



#18 AR-1242-3


R.T.: 5.752 min
Delta R.T.: 0.000 min
Response: 26654530
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500




#18 AR-1242-3

R.T.: 5.093 min
Delta R.T.: 0.000 min
Response: 19234787
Conc: 500.00 ng/ml




#19 AR-1242-4

R.T.: 5.849 min
Delta R.T.: 0.000 min
Response: 22517234
Conc: 500.00 ng/ml



#19 AR-1242-4


R.T.: 5.178 min
Delta R.T.: 0.000 min
Response: 18774167
Conc: 500.00 ng/ml



#20 AR-1242-5

R.T.: 6.581 min
Delta R.T.: 0.000 min
Response: 24165350
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500



#20 AR-1242-5

R.T.: 5.701 min
Delta R.T.: 0.000 min
Response: 22485728
Conc: 500.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071399.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 13:12
 Operator : YP\AJ
 Sample : AR1242ICC250
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 13:34:40 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 13:34:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.813	50759754	35812201	25.888	25.778
2) SA Decachlor...	10.233	8.850	37582587	22937032	26.000	26.351
Target Compounds						
16) L4 AR-1242-1	5.668	4.899	14669196	12394168	258.800	268.164
17) L4 AR-1242-2	5.689	4.917	22436050	17569793	258.254	264.383
18) L4 AR-1242-3	5.751	5.094	13459217	10184944	259.001	272.457
19) L4 AR-1242-4	5.849	5.178	11164111	9920754	256.375	272.906
20) L4 AR-1242-5	6.580	5.702	11989115	11814664	255.584	268.312



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071399.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 13:12
Operator : YP\AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 13:34:40 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 13:34:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.514 min
Delta R.T.: 0.000 min
Response: 50759754
Conc: 25.89 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.813 min
Delta R.T.: 0.000 min
Response: 35812201
Conc: 25.78 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.233 min
Delta R.T.: 0.000 min
Response: 37582587
Conc: 26.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.850 min
Delta R.T.: 0.000 min
Response: 22937032
Conc: 26.35 ng/ml



#16 AR-1242-1


R.T.: 5.668 min
Delta R.T.: 0.000 min
Response: 14669196
Conc: 258.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250




#16 AR-1242-1

R.T.: 4.899 min
Delta R.T.: 0.000 min
Response: 12394168
Conc: 268.16 ng/ml




#17 AR-1242-2

R.T.: 5.689 min
Delta R.T.: 0.000 min
Response: 22436050
Conc: 258.25 ng/ml



#17 AR-1242-2


R.T.: 4.917 min
Delta R.T.: 0.000 min
Response: 17569793
Conc: 264.38 ng/ml



#18 AR-1242-3


R.T.: 5.751 min
Delta R.T.: 0.000 min
Response: 13459217
Conc: 259.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250




#18 AR-1242-3

R.T.: 5.094 min
Delta R.T.: 0.000 min
Response: 10184944
Conc: 272.46 ng/ml




#19 AR-1242-4

R.T.: 5.849 min
Delta R.T.: 0.000 min
Response: 11164111
Conc: 256.37 ng/ml



#19 AR-1242-4


R.T.: 5.178 min
Delta R.T.: 0.000 min
Response: 9920754
Conc: 272.91 ng/ml



#20 AR-1242-5

R.T.: 6.580 min
Delta R.T.: 0.000 min
Response: 11989115
Conc: 255.58 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250



#20 AR-1242-5

R.T.: 5.702 min
Delta R.T.: 0.000 min
Response: 11814664
Conc: 268.31 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071400.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 13:28
 Operator : YP\AJ
 Sample : AR1242ICC050
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 13:44:30 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 13:44:19 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.812	8065732	6945226	4.265	4.999
2) SA Decachlor...	10.237	8.850	5895463	3833854	4.235	4.512
Target Compounds						
16) L4 AR-1242-1	5.670	4.899	2356519	2317190	43.025	50.108
17) L4 AR-1242-2	5.691	4.917	3869967	3214527	45.539	48.688
18) L4 AR-1242-3	5.754	5.094	2387637	1635706	46.703	44.877
19) L4 AR-1242-4	5.851	5.179	1842925	1657188	43.662	46.406
20) L4 AR-1242-5	6.582	5.702	2190873	2070356	47.329	47.586



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071400.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 13:28
Operator : YP\AJ
Sample : AR1242ICCO50
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICCO50

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 13:44:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 13:44:19 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.515 min
Delta R.T.: 0.000 min
Response: 8065732
Conc: 4.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050




#1 Tetrachloro-m-xylene

R.T.: 3.812 min
Delta R.T.: 0.000 min
Response: 6945226
Conc: 5.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.237 min
Delta R.T.: 0.000 min
Response: 5895463
Conc: 4.23 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.850 min
Delta R.T.: 0.000 min
Response: 3833854
Conc: 4.51 ng/ml



#16 AR-1242-1


R.T.: 5.670 min
Delta R.T.: 0.000 min
Response: 2356519
Conc: 43.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050




#16 AR-1242-1

R.T.: 4.899 min
Delta R.T.: 0.000 min
Response: 2317190
Conc: 50.11 ng/ml




#17 AR-1242-2

R.T.: 5.691 min
Delta R.T.: 0.000 min
Response: 3869967
Conc: 45.54 ng/ml



#17 AR-1242-2


R.T.: 4.917 min
Delta R.T.: 0.000 min
Response: 3214527
Conc: 48.69 ng/ml



#18 AR-1242-3


R.T.: 5.754 min
Delta R.T.: 0.000 min
Response: 2387637
Conc: 46.70 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050




#18 AR-1242-3

R.T.: 5.094 min
Delta R.T.: 0.000 min
Response: 1635706
Conc: 44.88 ng/ml




#19 AR-1242-4

R.T.: 5.851 min
Delta R.T.: 0.000 min
Response: 1842925
Conc: 43.66 ng/ml



#19 AR-1242-4


R.T.: 5.179 min
Delta R.T.: 0.000 min
Response: 1657188
Conc: 46.41 ng/ml



#20 AR-1242-5

R.T.: 6.582 min
Delta R.T.: 0.000 min
Response: 2190873
Conc: 47.33 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050



#20 AR-1242-5

R.T.: 5.702 min
Delta R.T.: 0.000 min
Response: 2070356
Conc: 47.59 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071401.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 13:45
 Operator : YP\AJ
 Sample : AR1248ICC1000
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 14:30:57 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 14:27:59 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.813	187.9E6	132.5E6	97.145	92.221
2) SA Decachlor...	10.237	8.850	137.9E6	82637510	96.283	95.713
Target Compounds						
21) L5 AR-1248-1	5.669	4.898	42447397	30996932	976.497	917.952
22) L5 AR-1248-2	5.941	5.136	57932834	43624561	965.902	945.881
23) L5 AR-1248-3	6.144	5.178	65030901	45791659	967.700	948.403
24) L5 AR-1248-4	6.543	5.350	80777360	53871444	963.427	957.270
25) L5 AR-1248-5	6.582	5.743	76029464	52150607	969.727	950.272



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071401.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 13:45
Operator : YP\AJ
Sample : AR1248ICC1000
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 14:30:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 14:27:59 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.515 min
Delta R.T.: 0.000 min
Response: 187919144
Conc: 97.15 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.813 min
Delta R.T.: 0.000 min
Response: 132471223
Conc: 92.22 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.237 min
Delta R.T.: 0.000 min
Response: 137913077
Conc: 96.28 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.850 min
Delta R.T.: 0.000 min
Response: 82637510
Conc: 95.71 ng/ml



#21 AR-1248-1


R.T.: 5.669 min
Delta R.T.: 0.000 min
Response: 42447397
Conc: 976.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000




#21 AR-1248-1

R.T.: 4.898 min
Delta R.T.: 0.000 min
Response: 30996932
Conc: 917.95 ng/ml




#22 AR-1248-2

R.T.: 5.941 min
Delta R.T.: 0.000 min
Response: 57932834
Conc: 965.90 ng/ml



#22 AR-1248-2


R.T.: 5.136 min
Delta R.T.: 0.000 min
Response: 43624561
Conc: 945.88 ng/ml



#23 AR-1248-3


R.T.: 6.144 min
Delta R.T.: 0.000 min
Response: 65030901
Conc: 967.70 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000




#23 AR-1248-3

R.T.: 5.178 min
Delta R.T.: 0.000 min
Response: 45791659
Conc: 948.40 ng/ml




#24 AR-1248-4

R.T.: 6.543 min
Delta R.T.: 0.000 min
Response: 80777360
Conc: 963.43 ng/ml



#24 AR-1248-4


R.T.: 5.350 min
Delta R.T.: 0.000 min
Response: 53871444
Conc: 957.27 ng/ml



#25 AR-1248-5

R.T.: 6.582 min
Delta R.T.: 0.000 min
Response: 76029464
Conc: 969.73 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000



#25 AR-1248-5

R.T.: 5.743 min
Delta R.T.: 0.000 min
Response: 52150607
Conc: 950.27 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071402.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 14:01
 Operator : YP\AJ
 Sample : AR1248ICC750
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 14:33:31 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 14:27:59 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.516	3.813	144.4E6	113.6E6	74.770	77.657
2) SA Decachlor...	10.238	8.850	108.0E6	64556701	75.275	74.847
Target Compounds						
21) L5 AR-1248-1	5.670	4.898	32527847	26959361	748.865	781.576
22) L5 AR-1248-2	5.941	5.135	45161422	35482042	751.976	762.779
23) L5 AR-1248-3	6.144	5.177	51040114	37424945	756.312	766.560
24) L5 AR-1248-4	6.543	5.350	63330814	43921936	753.553	770.043
25) L5 AR-1248-5	6.583	5.743	58877047	42217581	750.636	762.741



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071402.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 14:01
Operator : YP\AJ
Sample : AR1248ICC750
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 14:33:31 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 14:27:59 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.516 min
Delta R.T.: 0.000 min
Response: 144413867
Conc: 74.77 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.813 min
Delta R.T.: 0.000 min
Response: 113561766
Conc: 77.66 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.238 min
Delta R.T.: 0.000 min
Response: 108020118
Conc: 75.28 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.850 min
Delta R.T.: 0.000 min
Response: 64556701
Conc: 74.85 ng/ml



#21 AR-1248-1


R.T.: 5.670 min
Delta R.T.: 0.000 min
Response: 32527847
Conc: 748.87 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750




#21 AR-1248-1

R.T.: 4.898 min
Delta R.T.: 0.000 min
Response: 26959361
Conc: 781.58 ng/ml




#22 AR-1248-2

R.T.: 5.941 min
Delta R.T.: 0.000 min
Response: 45161422
Conc: 751.98 ng/ml



#22 AR-1248-2


R.T.: 5.135 min
Delta R.T.: 0.000 min
Response: 35482042
Conc: 762.78 ng/ml



#23 AR-1248-3


R.T.: 6.144 min
Delta R.T.: 0.000 min
Response: 51040114
Conc: 756.31 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750




#23 AR-1248-3

R.T.: 5.177 min
Delta R.T.: 0.000 min
Response: 37424945
Conc: 766.56 ng/ml




#24 AR-1248-4

R.T.: 6.543 min
Delta R.T.: 0.000 min
Response: 63330814
Conc: 753.55 ng/ml



#24 AR-1248-4


R.T.: 5.350 min
Delta R.T.: 0.000 min
Response: 43921936
Conc: 770.04 ng/ml



#25 AR-1248-5

R.T.: 6.583 min
Delta R.T.: 0.000 min
Response: 58877047
Conc: 750.64 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750



#25 AR-1248-5

R.T.: 5.743 min
Delta R.T.: 0.000 min
Response: 42217581
Conc: 762.74 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071403.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 14:17
 Operator : YP\AJ
 Sample : AR1248ICC500
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 14:28:25 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 14:27:59 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.813	99482021	77410182	50.000	50.000
2) SA Decachlor...	10.237	8.850	74280984	45020002	50.000	50.000
Target Compounds						
21) L5 AR-1248-1	5.668	4.898	22245336	18269007	500.000	500.000
22) L5 AR-1248-2	5.940	5.136	31011521	24308259	500.000	500.000
23) L5 AR-1248-3	6.142	5.178	34686045	25387087	500.000	500.000
24) L5 AR-1248-4	6.542	5.350	43455122	29340392	500.000	500.000
25) L5 AR-1248-5	6.581	5.743	40388237	28804357	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071403.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 14:17
Operator : YP\AJ
Sample : AR1248ICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 14:28:25 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 14:27:59 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.514 min
Delta R.T.: 0.000 min
Response: 99482021
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.813 min
Delta R.T.: 0.000 min
Response: 77410182
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.237 min
Delta R.T.: 0.000 min
Response: 74280984
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.850 min
Delta R.T.: 0.000 min
Response: 45020002
Conc: 50.00 ng/ml



#21 AR-1248-1


R.T.: 5.668 min
Delta R.T.: 0.000 min
Response: 22245336
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500




#21 AR-1248-1

R.T.: 4.898 min
Delta R.T.: 0.000 min
Response: 18269007
Conc: 500.00 ng/ml




#22 AR-1248-2

R.T.: 5.940 min
Delta R.T.: 0.000 min
Response: 31011521
Conc: 500.00 ng/ml



#22 AR-1248-2


R.T.: 5.136 min
Delta R.T.: 0.000 min
Response: 24308259
Conc: 500.00 ng/ml



#23 AR-1248-3


R.T.: 6.142 min
Delta R.T.: 0.000 min
Response: 34686045
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500




#23 AR-1248-3

R.T.: 5.178 min
Delta R.T.: 0.000 min
Response: 25387087
Conc: 500.00 ng/ml




#24 AR-1248-4

R.T.: 6.542 min
Delta R.T.: 0.000 min
Response: 43455122
Conc: 500.00 ng/ml



#24 AR-1248-4


R.T.: 5.350 min
Delta R.T.: 0.000 min
Response: 29340392
Conc: 500.00 ng/ml



#25 AR-1248-5

R.T.: 6.581 min
Delta R.T.: 0.000 min
Response: 40388237
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500



#25 AR-1248-5

R.T.: 5.743 min
Delta R.T.: 0.000 min
Response: 28804357
Conc: 500.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071404.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 14:33
 Operator : YP\AJ
 Sample : AR1248ICC250
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 14:46:35 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 14:46:22 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.812	50940076	35871474	26.017	24.646
2) SA Decachlor...	10.236	8.850	38029945	23509396	26.110	26.655
Target Compounds						
21) L5 AR-1248-1	5.668	4.898	11434087	9814021	259.799	275.024
22) L5 AR-1248-2	5.940	5.135	15821478	13376378	259.947	277.150
23) L5 AR-1248-3	6.143	5.178	17579514	14000544	257.788	276.598
24) L5 AR-1248-4	6.542	5.350	22285729	16222320	261.208	274.950
25) L5 AR-1248-5	6.581	5.743	20888056	16317809	262.033	282.168



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071404.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 14:33
Operator : YP\AJ
Sample : AR1248ICC250
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 14:46:35 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 14:46:22 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.514 min
Delta R.T.: 0.000 min
Response: 50940076
Conc: 26.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.812 min
Delta R.T.: 0.000 min
Response: 35871474
Conc: 24.65 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.236 min
Delta R.T.: 0.000 min
Response: 38029945
Conc: 26.11 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.850 min
Delta R.T.: 0.000 min
Response: 23509396
Conc: 26.66 ng/ml



#21 AR-1248-1


R.T.: 5.668 min
Delta R.T.: 0.000 min
Response: 11434087
Conc: 259.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250




#21 AR-1248-1

R.T.: 4.898 min
Delta R.T.: 0.000 min
Response: 9814021
Conc: 275.02 ng/ml




#22 AR-1248-2

R.T.: 5.940 min
Delta R.T.: 0.000 min
Response: 15821478
Conc: 259.95 ng/ml



#22 AR-1248-2


R.T.: 5.135 min
Delta R.T.: 0.000 min
Response: 13376378
Conc: 277.15 ng/ml



#23 AR-1248-3


R.T.: 6.143 min
Delta R.T.: 0.000 min
Response: 17579514
Conc: 257.79 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250




#23 AR-1248-3

R.T.: 5.178 min
Delta R.T.: 0.000 min
Response: 14000544
Conc: 276.60 ng/ml




#24 AR-1248-4

R.T.: 6.542 min
Delta R.T.: 0.000 min
Response: 22285729
Conc: 261.21 ng/ml



#24 AR-1248-4


R.T.: 5.350 min
Delta R.T.: 0.000 min
Response: 16222320
Conc: 274.95 ng/ml



#25 AR-1248-5

R.T.: 6.581 min
Delta R.T.: 0.000 min
Response: 20888056
Conc: 262.03 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250



#25 AR-1248-5

R.T.: 5.743 min
Delta R.T.: 0.000 min
Response: 16317809
Conc: 282.17 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071405.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 14:50
 Operator : YP\AJ
 Sample : AR1248ICC050
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 15:04:57 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 15:04:44 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.513	3.813	9059747	7084653	4.697	4.893
2) SA Decachlor...	10.236	8.850	6791716	4608858	4.727	5.179
Target Compounds						
21) L5 AR-1248-1	5.667	4.899	1932478	1889804	45.005	52.340
22) L5 AR-1248-2	5.939	5.136	2935274	2587801	48.571	52.853
23) L5 AR-1248-3	6.142	5.178	3057380	2647979	45.780	51.834
24) L5 AR-1248-4	6.541	5.350	4089143	3364119	48.329	55.461
25) L5 AR-1248-5	6.581	5.744	3919031	3006686	49.328	51.581



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071405.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 14:50
Operator : YP\AJ
Sample : AR1248ICC050
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 15:04:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 15:04:44 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.513 min
Delta R.T.: 0.000 min
Response: 9059747
Conc: 4.70 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC050




#1 Tetrachloro-m-xylene

R.T.: 3.813 min
Delta R.T.: 0.000 min
Response: 7084653
Conc: 4.89 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.236 min
Delta R.T.: 0.000 min
Response: 6791716
Conc: 4.73 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.850 min
Delta R.T.: 0.000 min
Response: 4608858
Conc: 5.18 ng/ml



#21 AR-1248-1


R.T.: 5.667 min
Delta R.T.: 0.000 min
Response: 1932478
Conc: 45.01 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC050




#21 AR-1248-1

R.T.: 4.899 min
Delta R.T.: 0.000 min
Response: 1889804
Conc: 52.34 ng/ml




#22 AR-1248-2

R.T.: 5.939 min
Delta R.T.: 0.000 min
Response: 2935274
Conc: 48.57 ng/ml



#22 AR-1248-2


R.T.: 5.136 min
Delta R.T.: 0.000 min
Response: 2587801
Conc: 52.85 ng/ml



#23 AR-1248-3


R.T.: 6.142 min
Delta R.T.: 0.000 min
Response: 3057380
Conc: 45.78 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC050




#23 AR-1248-3

R.T.: 5.178 min
Delta R.T.: 0.000 min
Response: 2647979
Conc: 51.83 ng/ml




#24 AR-1248-4

R.T.: 6.541 min
Delta R.T.: 0.000 min
Response: 4089143
Conc: 48.33 ng/ml



#24 AR-1248-4


R.T.: 5.350 min
Delta R.T.: 0.000 min
Response: 3364119
Conc: 55.46 ng/ml



#25 AR-1248-5

R.T.: 6.581 min
Delta R.T.: 0.000 min
Response: 3919031
Conc: 49.33 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC050



#25 AR-1248-5

R.T.: 5.744 min
Delta R.T.: 0.000 min
Response: 3006686
Conc: 51.58 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071406.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 15:06
 Operator : YP\AJ
 Sample : AR1254ICC1000
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 15:51:39 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 15:48:40 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.812	188.8E6	139.1E6	96.818	97.876
2) SA Decachlor...	10.236	8.850	141.0E6	84630750	96.040	96.040
Target Compounds						
26) L6 AR-1254-1	6.519	5.702	78574372	81690839	957.338	952.962
27) L6 AR-1254-2	6.735	5.851	121.6E6	70045476	958.538	950.650
28) L6 AR-1254-3	7.099	6.254	123.4E6	109.7E6	962.759	954.874
29) L6 AR-1254-4	7.381	6.483	112.5E6	70169900	960.617	948.370
30) L6 AR-1254-5	7.798	6.901	106.4E6	92229111	966.455	936.717



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071406.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 15:06
Operator : YP\AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 15:51:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 15:48:40 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.515 min
Delta R.T.: 0.000 min
Response: 188824112
Conc: 96.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.812 min
Delta R.T.: 0.000 min
Response: 139055330
Conc: 97.88 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.236 min
Delta R.T.: 0.000 min
Response: 141010181
Conc: 96.04 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.850 min
Delta R.T.: 0.000 min
Response: 84630750
Conc: 96.04 ng/ml



#26 AR-1254-1


R.T.: 6.519 min
Delta R.T.: 0.000 min
Response: 78574372
Conc: 957.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000




#26 AR-1254-1

R.T.: 5.702 min
Delta R.T.: 0.000 min
Response: 81690839
Conc: 952.96 ng/ml




#27 AR-1254-2

R.T.: 6.735 min
Delta R.T.: 0.000 min
Response: 121624435
Conc: 958.54 ng/ml



#27 AR-1254-2


R.T.: 5.851 min
Delta R.T.: 0.000 min
Response: 70045476
Conc: 950.65 ng/ml



#28 AR-1254-3


R.T.: 7.099 min
Delta R.T.: 0.000 min
Response: 123432630
Conc: 962.76 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000




#28 AR-1254-3

R.T.: 6.254 min
Delta R.T.: 0.000 min
Response: 109725744
Conc: 954.87 ng/ml




#29 AR-1254-4

R.T.: 7.381 min
Delta R.T.: 0.000 min
Response: 112512019
Conc: 960.62 ng/ml



#29 AR-1254-4


R.T.: 6.483 min
Delta R.T.: 0.000 min
Response: 70169900
Conc: 948.37 ng/ml



#30 AR-1254-5

R.T.: 7.798 min
Delta R.T.: 0.000 min
Response: 106399089
Conc: 966.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000



#30 AR-1254-5

R.T.: 6.901 min
Delta R.T.: 0.000 min
Response: 92229111
Conc: 936.72 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071407.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 15:22
 Operator : YP\AJ
 Sample : AR1254ICC750
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC750

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/23/2025
 Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 15:54:33 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 15:48:40 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.812	145.3E6	112.6E6	74.658	77.773
2) SA Decachlor...	10.236	8.850	110.2E6	65650190	75.033	74.666
Target Compounds						
26) L6 AR-1254-1	6.519	5.703	60934961	65520948	744.931	759.494
27) L6 AR-1254-2	6.735	5.851	94362074	56602387	745.775	762.037
28) L6 AR-1254-3	7.098	6.255	95849879	85577532	748.410	746.476
29) L6 AR-1254-4	7.381	6.483	87194278	55551540	746.295	750.532
30) L6 AR-1254-5	7.797	6.901	79441918	72082199	721.000m	737.969

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071407.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 15:22
Operator : YP\AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254ICC750


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 15:54:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 15:48:40 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.515 min
Delta R.T.: 0.000 min
Response: 145274275
Conc: 74.66 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#1 Tetrachloro-m-xylene

R.T.: 3.812 min
Delta R.T.: 0.000 min
Response: 112576042
Conc: 77.77 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.236 min
Delta R.T.: 0.000 min
Response: 110190189
Conc: 75.03 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.850 min
Delta R.T.: 0.000 min
Response: 65650190
Conc: 74.67 ng/ml




#26 AR-1254-1

R.T.: 6.519 min
Delta R.T.: 0.000 min
Response: 60934961
Conc: 744.93 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#26 AR-1254-1

R.T.: 5.703 min
Delta R.T.: 0.000 min
Response: 65520948
Conc: 759.49 ng/ml




#27 AR-1254-2

R.T.: 6.735 min
Delta R.T.: 0.000 min
Response: 94362074
Conc: 745.77 ng/ml



#27 AR-1254-2

R.T.: 5.851 min
Delta R.T.: 0.000 min
Response: 56602387
Conc: 762.04 ng/ml




#28 AR-1254-3

R.T.: 7.098 min
Delta R.T.: 0.000 min
Response: 95849879
Conc: 748.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#28 AR-1254-3

R.T.: 6.255 min
Delta R.T.: 0.000 min
Response: 85577532
Conc: 746.48 ng/ml




#29 AR-1254-4

R.T.: 7.381 min
Delta R.T.: 0.000 min
Response: 87194278
Conc: 746.30 ng/ml



#29 AR-1254-4

R.T.: 6.483 min
Delta R.T.: 0.000 min
Response: 55551540
Conc: 750.53 ng/ml




#30 AR-1254-5

R.T.: 7.797 min
Delta R.T.: 0.000 min
Response: 79441918
Conc: 721.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025



#30 AR-1254-5

R.T.: 6.901 min
Delta R.T.: 0.000 min
Response: 72082199
Conc: 737.97 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071408.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 15:38
 Operator : YP\AJ
 Sample : AR1254ICC500
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 15:48:55 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 15:48:40 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.812	100.6E6	72545075	50.000	50.000
2) SA Decachlor...	10.237	8.849	76319487	45805245	50.000	50.000
Target Compounds						
26) L6 AR-1254-1	6.519	5.702	42788667	44877702	500.000	500.000
27) L6 AR-1254-2	6.735	5.851	66073164	38658957	500.000	500.000
28) L6 AR-1254-3	7.098	6.254	66490827	60048409	500.000	500.000
29) L6 AR-1254-4	7.381	6.483	60868718	38905081	500.000	500.000
30) L6 AR-1254-5	7.797	6.901	56892560	52345397	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071408.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 15:38
Operator : YP\AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 15:48:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 15:48:40 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.515 min
Delta R.T.: 0.000 min
Response: 100618430
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.812 min
Delta R.T.: 0.000 min
Response: 72545075
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.237 min
Delta R.T.: 0.000 min
Response: 76319487
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.849 min
Delta R.T.: 0.000 min
Response: 45805245
Conc: 50.00 ng/ml



#26 AR-1254-1


R.T.: 6.519 min
Delta R.T.: 0.000 min
Response: 42788667
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500




#26 AR-1254-1

R.T.: 5.702 min
Delta R.T.: 0.000 min
Response: 44877702
Conc: 500.00 ng/ml




#27 AR-1254-2

R.T.: 6.735 min
Delta R.T.: 0.000 min
Response: 66073164
Conc: 500.00 ng/ml



#27 AR-1254-2


R.T.: 5.851 min
Delta R.T.: 0.000 min
Response: 38658957
Conc: 500.00 ng/ml



#28 AR-1254-3


R.T.: 7.098 min
Delta R.T.: 0.000 min
Response: 66490827
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500




#28 AR-1254-3

R.T.: 6.254 min
Delta R.T.: 0.000 min
Response: 60048409
Conc: 500.00 ng/ml




#29 AR-1254-4

R.T.: 7.381 min
Delta R.T.: 0.000 min
Response: 60868718
Conc: 500.00 ng/ml



#29 AR-1254-4


R.T.: 6.483 min
Delta R.T.: 0.000 min
Response: 38905081
Conc: 500.00 ng/ml



#30 AR-1254-5

R.T.: 7.797 min
Delta R.T.: 0.000 min
Response: 56892560
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500



#30 AR-1254-5

R.T.: 6.901 min
Delta R.T.: 0.000 min
Response: 52345397
Conc: 500.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071409.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 15:55
 Operator : YP\AJ
 Sample : AR1254ICC250
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 16:07:19 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 16:07:08 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.513	3.813	51789148	37965204	26.192	25.910
2) SA Decachlor...	10.234	8.850	38742671	23881547	26.022	26.587
Target Compounds						
26) L6 AR-1254-1	6.517	5.703	22149547	24321384	265.267	273.203
27) L6 AR-1254-2	6.733	5.851	34278927	20788594	265.367	271.757
28) L6 AR-1254-3	7.096	6.255	34405813	31095041	263.728	265.596
29) L6 AR-1254-4	7.379	6.483	31098214	20469357	261.934	269.399
30) L6 AR-1254-5	7.796	6.901	29038261	26846100	262.635	268.183



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071409.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 15:55
Operator : YP\AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 16:07:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 16:07:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.513 min
Delta R.T.: 0.000 min
Response: 51789148
Conc: 26.19 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.813 min
Delta R.T.: 0.000 min
Response: 37965204
Conc: 25.91 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.234 min
Delta R.T.: 0.000 min
Response: 38742671
Conc: 26.02 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.850 min
Delta R.T.: 0.000 min
Response: 23881547
Conc: 26.59 ng/ml



#26 AR-1254-1


R.T.: 6.517 min
Delta R.T.: 0.000 min
Response: 22149547
Conc: 265.27 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250




#26 AR-1254-1

R.T.: 5.703 min
Delta R.T.: 0.000 min
Response: 24321384
Conc: 273.20 ng/ml




#27 AR-1254-2

R.T.: 6.733 min
Delta R.T.: 0.000 min
Response: 34278927
Conc: 265.37 ng/ml



#27 AR-1254-2


R.T.: 5.851 min
Delta R.T.: 0.000 min
Response: 20788594
Conc: 271.76 ng/ml



#28 AR-1254-3


R.T.: 7.096 min
Delta R.T.: 0.000 min
Response: 34405813
Conc: 263.73 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250




#28 AR-1254-3

R.T.: 6.255 min
Delta R.T.: 0.000 min
Response: 31095041
Conc: 265.60 ng/ml




#29 AR-1254-4

R.T.: 7.379 min
Delta R.T.: 0.000 min
Response: 31098214
Conc: 261.93 ng/ml



#29 AR-1254-4


R.T.: 6.483 min
Delta R.T.: 0.000 min
Response: 20469357
Conc: 269.40 ng/ml



#30 AR-1254-5

R.T.: 7.796 min
Delta R.T.: 0.000 min
Response: 29038261
Conc: 262.64 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250



#30 AR-1254-5

R.T.: 6.901 min
Delta R.T.: 0.000 min
Response: 26846100
Conc: 268.18 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071410.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 16:11
 Operator : YP\AJ
 Sample : AR1254ICC050
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/23/2025
 Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 16:54:51 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 16:54:36 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.813	8612576	7342515	4.471m	5.009
2) SA Decachlor...	10.236	8.850	6643627	4341119	4.560	4.865
Target Compounds						
26) L6 AR-1254-1	6.520	5.703	3787970	5045448	46.222	55.202
27) L6 AR-1254-2	6.735	5.852	5933913	4378664	46.696	55.629
28) L6 AR-1254-3	7.099	6.255	5851823	6211772	45.798	52.416
29) L6 AR-1254-4	7.382	6.484	5633281	4095094	47.937	53.069
30) L6 AR-1254-5	7.798	6.901	4819584	5346230	44.737	52.689

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071410.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 16:11
Operator : YP\AJ
Sample : AR1254ICC050
Misc :
ALS Vial : 24 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 16:54:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 16:54:36 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.514 min
Delta R.T.: 0.000 min
Response: 8612576
Conc: 4.47 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#1 Tetrachloro-m-xylene

R.T.: 3.813 min
Delta R.T.: 0.000 min
Response: 7342515
Conc: 5.01 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.236 min
Delta R.T.: 0.000 min
Response: 6643627
Conc: 4.56 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.850 min
Delta R.T.: 0.000 min
Response: 4341119
Conc: 4.87 ng/ml




#26 AR-1254-1

R.T.: 6.520 min
Delta R.T.: 0.000 min
Response: 3787970
Conc: 46.22 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#26 AR-1254-1

R.T.: 5.703 min
Delta R.T.: 0.000 min
Response: 5045448
Conc: 55.20 ng/ml




#27 AR-1254-2

R.T.: 6.735 min
Delta R.T.: 0.000 min
Response: 5933913
Conc: 46.70 ng/ml



#27 AR-1254-2

R.T.: 5.852 min
Delta R.T.: 0.000 min
Response: 4378664
Conc: 55.63 ng/ml




#28 AR-1254-3

R.T.: 7.099 min
Delta R.T.: 0.000 min
Response: 5851823
Conc: 45.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#28 AR-1254-3

R.T.: 6.255 min
Delta R.T.: 0.000 min
Response: 6211772
Conc: 52.42 ng/ml




#29 AR-1254-4

R.T.: 7.382 min
Delta R.T.: 0.000 min
Response: 5633281
Conc: 47.94 ng/ml



#29 AR-1254-4

R.T.: 6.484 min
Delta R.T.: 0.000 min
Response: 4095094
Conc: 53.07 ng/ml




#30 AR-1254-5

R.T.: 7.798 min
Delta R.T.: 0.000 min
Response: 4819584
Conc: 44.74 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025



#30 AR-1254-5

R.T.: 6.901 min
Delta R.T.: 0.000 min
Response: 5346230
Conc: 52.69 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071411.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 16:27
 Operator : YP\AJ
 Sample : AR1262ICC500
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1262ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 17:08:30 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 17:08:14 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.813	98220466	71170259	50.000	50.000
2) SA Decachlor...	10.235	8.850	74643123	44870081	50.000	50.000
Target Compounds						
36) L8 AR-1262-1	8.100	6.939	72245642	56806132	500.000	500.000
37) L8 AR-1262-2	8.420	7.198	137.1E6	46169678	500.000	500.000
38) L8 AR-1262-3	8.737	7.721	91438860	37847231	500.000	500.000
39) L8 AR-1262-4	8.822	7.786	67105487	62412938	500.000	500.000
40) L8 AR-1262-5	9.478	8.286	44080983	27746588	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071411.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 16:27
Operator : YP\AJ
Sample : AR1262ICC500
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1262ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 17:08:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 17:08:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.515 min
Delta R.T.: 0.000 min
Response: 98220466
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1262ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.813 min
Delta R.T.: 0.000 min
Response: 71170259
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.235 min
Delta R.T.: 0.000 min
Response: 74643123
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.850 min
Delta R.T.: 0.000 min
Response: 44870081
Conc: 50.00 ng/ml



#36 AR-1262-1


R.T.: 8.100 min
Delta R.T.: 0.000 min
Response: 72245642
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1262ICC500




#36 AR-1262-1

R.T.: 6.939 min
Delta R.T.: 0.000 min
Response: 56806132
Conc: 500.00 ng/ml




#37 AR-1262-2

R.T.: 8.420 min
Delta R.T.: 0.000 min
Response: 137093050
Conc: 500.00 ng/ml



#37 AR-1262-2


R.T.: 7.198 min
Delta R.T.: 0.000 min
Response: 46169678
Conc: 500.00 ng/ml



#38 AR-1262-3


R.T.: 8.737 min
Delta R.T.: 0.000 min
Response: 91438860
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1262ICC500




#38 AR-1262-3

R.T.: 7.721 min
Delta R.T.: 0.000 min
Response: 37847231
Conc: 500.00 ng/ml




#39 AR-1262-4

R.T.: 8.822 min
Delta R.T.: 0.000 min
Response: 67105487
Conc: 500.00 ng/ml



#39 AR-1262-4


R.T.: 7.786 min
Delta R.T.: 0.000 min
Response: 62412938
Conc: 500.00 ng/ml



#40 AR-1262-5

R.T.: 9.478 min
Delta R.T.: 0.000 min
Response: 44080983
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1262ICC500



#40 AR-1262-5

R.T.: 8.286 min
Delta R.T.: 0.000 min
Response: 27746588
Conc: 500.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071412.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 16:44
 Operator : YP\AJ
 Sample : AR1268ICC1000
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC1000

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/23/2025
 Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 04:44:55 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 04:41:50 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.812	188.4E6	147.6E6	95.054	100.837
2) SA Decachlor...	10.235	8.850	248.7E6	149.9E6	93.152	93.528
Target Compounds						
41) L9 AR-1268-1	8.733	7.722	299.6E6	181.9E6	936.979	895.922m
42) L9 AR-1268-2	8.826	7.788	251.1E6	154.8E6	937.116	896.075
43) L9 AR-1268-3	9.058	7.993	211.9E6	128.5E6	940.656	898.014
44) L9 AR-1268-4	9.477	8.287	93231632	56964280	941.667	896.773
45) L9 AR-1268-5	9.894	8.588	590.1E6	367.0E6	950.117	952.869

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071412.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 16:44
Operator : YP\AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :

ECD_P

ClientSampleId :

AR1268ICC1000

Manual Integrations



APPROVED


Reviewed By :Yogesh Patel 04/23/2025

Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 23 04:44:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 04:41:50 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.514 min
Delta R.T.: -0.001 min
Response: 188397795
Conc: 95.05 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#1 Tetrachloro-m-xylene

R.T.: 3.812 min
Delta R.T.: -0.002 min
Response: 147647099
Conc: 100.84 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.235 min
Delta R.T.: 0.000 min
Response: 248662366
Conc: 93.15 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.850 min
Delta R.T.: -0.001 min
Response: 149905120
Conc: 93.53 ng/ml




#41 AR-1268-1

R.T.: 8.733 min
Delta R.T.: 0.000 min
Response: 299594266
Conc: 936.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#41 AR-1268-1

R.T.: 7.722 min
Delta R.T.: 0.000 min
Response: 181860271
Conc: 895.92 ng/ml




#42 AR-1268-2

R.T.: 8.826 min
Delta R.T.: 0.000 min
Response: 251109850
Conc: 937.12 ng/ml



#42 AR-1268-2

R.T.: 7.788 min
Delta R.T.: 0.000 min
Response: 154801411
Conc: 896.07 ng/ml




#43 AR-1268-3

R.T.: 9.058 min
Delta R.T.: 0.000 min
Response: 211915080
Conc: 940.66 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#43 AR-1268-3

R.T.: 7.993 min
Delta R.T.: 0.000 min
Response: 128491160
Conc: 898.01 ng/ml




#44 AR-1268-4

R.T.: 9.477 min
Delta R.T.: 0.000 min
Response: 93231632
Conc: 941.67 ng/ml



#44 AR-1268-4

R.T.: 8.287 min
Delta R.T.: 0.000 min
Response: 56964280
Conc: 896.77 ng/ml




#45 AR-1268-5

R.T.: 9.894 min
Delta R.T.: 0.000 min
Response: 590082071
Conc: 950.12 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025



#45 AR-1268-5

R.T.: 8.588 min
Delta R.T.: 0.000 min
Response: 367021219
Conc: 952.87 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071413.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 17:00
 Operator : YP\AJ
 Sample : AR1268ICC750
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC750

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/23/2025
 Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 04:45:40 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 04:41:50 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.813	144.9E6	108.4E6	73.091	74.065
2) SA Decachlor...	10.234	8.849	191.8E6	116.0E6	71.833	72.344
Target Compounds						
41) L9 AR-1268-1	8.733	7.721	231.8E6	141.3E6	725.026	696.327m
42) L9 AR-1268-2	8.826	7.787	194.3E6	119.6E6	725.052	692.534
43) L9 AR-1268-3	9.058	7.991	163.8E6	99816904	726.929	697.612
44) L9 AR-1268-4	9.476	8.285	72323269	44771879	730.486	704.831
45) L9 AR-1268-5	9.894	8.587	452.8E6	283.3E6	729.082	735.437

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071413.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 17:00
Operator : YP\AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1268ICC750


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 23 04:45:40 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 04:41:50 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.515 min
Delta R.T.: 0.000 min
Response: 144867571
Conc: 73.09 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#1 Tetrachloro-m-xylene

R.T.: 3.813 min
Delta R.T.: -0.001 min
Response: 108446461
Conc: 74.06 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.234 min
Delta R.T.: -0.001 min
Response: 191753222
Conc: 71.83 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.849 min
Delta R.T.: -0.002 min
Response: 115953007
Conc: 72.34 ng/ml




#41 AR-1268-1

R.T.: 8.733 min
Delta R.T.: 0.000 min
Response: 231823386
Conc: 725.03 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#41 AR-1268-1

R.T.: 7.721 min
Delta R.T.: -0.002 min
Response: 141345145
Conc: 696.33 ng/ml m




#42 AR-1268-2

R.T.: 8.826 min
Delta R.T.: 0.000 min
Response: 194285253
Conc: 725.05 ng/ml



#42 AR-1268-2

R.T.: 7.787 min
Delta R.T.: -0.001 min
Response: 119638783
Conc: 692.53 ng/ml




#43 AR-1268-3

R.T.: 9.058 min
Delta R.T.: 0.000 min
Response: 163765585
Conc: 726.93 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#43 AR-1268-3

R.T.: 7.991 min
Delta R.T.: -0.002 min
Response: 99816904
Conc: 697.61 ng/ml




#44 AR-1268-4

R.T.: 9.476 min
Delta R.T.: -0.002 min
Response: 72323269
Conc: 730.49 ng/ml



#44 AR-1268-4

R.T.: 8.285 min
Delta R.T.: -0.002 min
Response: 44771879
Conc: 704.83 ng/ml




#45 AR-1268-5

R.T.: 9.894 min
Delta R.T.: 0.000 min
Response: 452805185
Conc: 729.08 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025



#45 AR-1268-5

R.T.: 8.587 min
Delta R.T.: -0.002 min
Response: 283271897
Conc: 735.44 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071414.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 17:16
 Operator : YP\AJ
 Sample : AR1268ICC500
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/23/2025
 Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 04:46:05 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 04:41:50 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.814	99100413	73210719	50.000	50.000
2) SA Decachlor...	10.236	8.851	133.5E6	80139523	50.000	50.000
Target Compounds						
41) L9 AR-1268-1	8.733	7.722	159.9E6	101.7E6	500.000	501.224m
42) L9 AR-1268-2	8.826	7.788	134.0E6	86377504	500.000	500.000
43) L9 AR-1268-3	9.059	7.993	112.6E6	71541846	500.000	500.000
44) L9 AR-1268-4	9.478	8.287	49503520	31760699	500.000	500.000
45) L9 AR-1268-5	9.894	8.589	310.5E6	192.6E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071414.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 17:16
Operator : YP\AJ
Sample : AR1268ICC500
Misc :
ALS Vial : 28 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1268ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 23 04:46:05 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 04:41:50 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.515 min
Delta R.T.: 0.000 min
Response: 99100413
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#1 Tetrachloro-m-xylene

R.T.: 3.814 min
Delta R.T.: 0.000 min
Response: 73210719
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.236 min
Delta R.T.: 0.000 min
Response: 133470748
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.851 min
Delta R.T.: 0.000 min
Response: 80139523
Conc: 50.00 ng/ml




#41 AR-1268-1

R.T.: 8.733 min
Delta R.T.: 0.000 min
Response: 159872402
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#41 AR-1268-1

R.T.: 7.722 min
Delta R.T.: 0.000 min
Response: 101741890
Conc: 501.22 ng/ml




#42 AR-1268-2

R.T.: 8.826 min
Delta R.T.: 0.000 min
Response: 133980177
Conc: 500.00 ng/ml



#42 AR-1268-2

R.T.: 7.788 min
Delta R.T.: 0.000 min
Response: 86377504
Conc: 500.00 ng/ml




#43 AR-1268-3

R.T.: 9.059 min
Delta R.T.: 0.000 min
Response: 112642140
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#43 AR-1268-3

R.T.: 7.993 min
Delta R.T.: 0.000 min
Response: 71541846
Conc: 500.00 ng/ml




#44 AR-1268-4

R.T.: 9.478 min
Delta R.T.: 0.000 min
Response: 49503520
Conc: 500.00 ng/ml



#44 AR-1268-4

R.T.: 8.287 min
Delta R.T.: 0.000 min
Response: 31760699
Conc: 500.00 ng/ml




#45 AR-1268-5

R.T.: 9.894 min
Delta R.T.: 0.000 min
Response: 310531224
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025



#45 AR-1268-5

R.T.: 8.589 min
Delta R.T.: 0.000 min
Response: 192587534
Conc: 500.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071415.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 17:33
 Operator : YP\AJ
 Sample : AR1268ICC250
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC250

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/23/2025
 Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 04:47:10 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 04:41:50 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.517	3.813	50174786	36443440	25.315	24.889
2) SA Decachlor...	10.236	8.850	67366681	41318494	25.236	25.779
Target Compounds						
41) L9 AR-1268-1	8.735	7.721	82956522	51353583	259.446	252.990m
42) L9 AR-1268-2	8.828	7.788	68964227	43318102	257.367	250.749
43) L9 AR-1268-3	9.061	7.991	57876424	36193296	256.904	252.952
44) L9 AR-1268-4	9.480	8.286	24514996	16044448	247.609	252.583
45) L9 AR-1268-5	9.897	8.587	157.8E6	96869753	254.135	251.495

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071415.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 17:33
Operator : YP\AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1268ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 23 04:47:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 04:41:50 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.517 min
Delta R.T.: 0.002 min
Response: 50174786
Conc: 25.32 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#1 Tetrachloro-m-xylene

R.T.: 3.813 min
Delta R.T.: -0.001 min
Response: 36443440
Conc: 24.89 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.236 min
Delta R.T.: 0.000 min
Response: 67366681
Conc: 25.24 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.850 min
Delta R.T.: -0.002 min
Response: 41318494
Conc: 25.78 ng/ml




#41 AR-1268-1

R.T.: 8.735 min
Delta R.T.: 0.002 min
Response: 82956522
Conc: 259.45 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#41 AR-1268-1

R.T.: 7.721 min
Delta R.T.: -0.001 min
Response: 51353583
Conc: 252.99 ng/ml m




#42 AR-1268-2

R.T.: 8.828 min
Delta R.T.: 0.002 min
Response: 68964227
Conc: 257.37 ng/ml



#42 AR-1268-2

R.T.: 7.788 min
Delta R.T.: 0.000 min
Response: 43318102
Conc: 250.75 ng/ml




#43 AR-1268-3

R.T.: 9.061 min
Delta R.T.: 0.003 min
Response: 57876424
Conc: 256.90 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#43 AR-1268-3

R.T.: 7.991 min
Delta R.T.: -0.002 min
Response: 36193296
Conc: 252.95 ng/ml




#44 AR-1268-4

R.T.: 9.480 min
Delta R.T.: 0.002 min
Response: 24514996
Conc: 247.61 ng/ml



#44 AR-1268-4

R.T.: 8.286 min
Delta R.T.: -0.001 min
Response: 16044448
Conc: 252.58 ng/ml




#45 AR-1268-5

R.T.: 9.897 min
Delta R.T.: 0.003 min
Response: 157833758
Conc: 254.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025



#45 AR-1268-5

R.T.: 8.587 min
Delta R.T.: -0.002 min
Response: 96869753
Conc: 251.50 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071416.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 17:49
 Operator : YP\AJ
 Sample : AR1268ICC050
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/23/2025
 Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 04:47:36 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 04:41:50 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.813	7954830	6504410	4.014	4.442
2) SA Decachlor...	10.235	8.850	11316360	7324768	4.239	4.570
Target Compounds						
41) L9 AR-1268-1	8.731	7.721	14103190	10095232	44.108m	49.733m
42) L9 AR-1268-2	8.825	7.787	11406198	8528068	42.567m	49.365
43) L9 AR-1268-3	9.058	7.992	10823544	6773456	48.044	47.339
44) L9 AR-1268-4	9.476	8.287	4136815	2789666	41.783	43.917
45) L9 AR-1268-5	9.894	8.588	26835395	17143092	43.209	44.507

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071416.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 17:49
Operator : YP\AJ
Sample : AR1268ICC050
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :

ECD_P

ClientSampleId :

AR1268ICC050

Manual Integrations



APPROVED


Reviewed By :Yogesh Patel 04/23/2025

Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 23 04:47:36 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 04:41:50 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.514 min
Delta R.T.: -0.001 min
Response: 7954830
Conc: 4.01 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#1 Tetrachloro-m-xylene

R.T.: 3.813 min
Delta R.T.: -0.001 min
Response: 6504410
Conc: 4.44 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.235 min
Delta R.T.: 0.000 min
Response: 11316360
Conc: 4.24 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.850 min
Delta R.T.: 0.000 min
Response: 7324768
Conc: 4.57 ng/ml




#41 AR-1268-1

R.T.: 8.731 min
Delta R.T.: -0.001 min
Response: 14103190
Conc: 44.11 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#41 AR-1268-1

R.T.: 7.721 min
Delta R.T.: -0.001 min
Response: 10095232
Conc: 49.73 ng/ml




#42 AR-1268-2

R.T.: 8.825 min
Delta R.T.: -0.001 min
Response: 11406198
Conc: 42.57 ng/ml



#42 AR-1268-2

R.T.: 7.787 min
Delta R.T.: -0.001 min
Response: 8528068
Conc: 49.37 ng/ml




#43 AR-1268-3

R.T.: 9.058 min
Delta R.T.: 0.000 min
Response: 10823544
Conc: 48.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#43 AR-1268-3

R.T.: 7.992 min
Delta R.T.: 0.000 min
Response: 6773456
Conc: 47.34 ng/ml




#44 AR-1268-4

R.T.: 9.476 min
Delta R.T.: -0.002 min
Response: 4136815
Conc: 41.78 ng/ml



#44 AR-1268-4

R.T.: 8.287 min
Delta R.T.: 0.000 min
Response: 2789666
Conc: 43.92 ng/ml




#45 AR-1268-5

R.T.: 9.894 min
Delta R.T.: 0.000 min
Response: 26835395
Conc: 43.21 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025



#45 AR-1268-5

R.T.: 8.588 min
Delta R.T.: 0.000 min
Response: 17143092
Conc: 44.51 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071417.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 18:05
 Operator : YP\AJ
 Sample : PP042225ICV500
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP042225

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 19:27:41 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 17:08:14 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.812	98003244	70310326	49.431	49.914
2) SA Decachlor...	10.233	8.849	73424396	44681766	50.707	50.945
Target Compounds						
3) L1 AR-1016-1	5.668	4.898	33546404	27255669	500.321	509.132
4) L1 AR-1016-2	5.690	4.916	50377176	38783246	490.870	507.875
5) L1 AR-1016-3	5.752	5.093	30221252	21921068	490.230	519.328
6) L1 AR-1016-4	5.850	5.135	25444869	18087794	496.019	524.498
7) L1 AR-1016-5	6.142	5.349	24049539	22819712	498.556	513.411
31) L7 AR-1260-1	7.262	6.384	46974832	35744650	489.196	494.399
32) L7 AR-1260-2	7.515	6.573	70731391	43416292	494.358	496.405
33) L7 AR-1260-3	7.875	6.726	58007061	39241127	517.439	500.604
34) L7 AR-1260-4	8.098	7.198	56235072	31343464	498.252	496.643
35) L7 AR-1260-5	8.418	7.439	114.7E6	75416996	506.086	500.188



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071417.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 18:05
Operator : YP\AJ
Sample : PP042225ICV500
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP042225

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 19:27:41 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 17:08:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.514 min
Delta R.T.: -0.003 min
Response: 98003244
Conc: 49.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042225




#1 Tetrachloro-m-xylene

R.T.: 3.812 min
Delta R.T.: 0.000 min
Response: 70310326
Conc: 49.91 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.233 min
Delta R.T.: -0.004 min
Response: 73424396
Conc: 50.71 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.849 min
Delta R.T.: 0.000 min
Response: 44681766
Conc: 50.95 ng/ml

Response_

Signal: PP071417.D\ECD1A.ch

#3 AR-1016-1

ICAL Form



R.T.: 5.668 min
Delta R.T.: -0.003 min
Response: 33546404
Conc: 500.32 ng/ml


Instrument :
ECD_P
ClientSampleId :
ICVPP042225

Time

Response_

Signal: PP071417.D\ECD2B.ch

#3 AR-1016-1




R.T.: 4.898 min
Delta R.T.: 0.000 min
Response: 27255669
Conc: 509.13 ng/ml

Time

Response_

Signal: PP071417.D\ECD1A.ch

#4 AR-1016-2




R.T.: 5.690 min
Delta R.T.: -0.003 min
Response: 50377176
Conc: 490.87 ng/ml

Time

Response_


Signal: PP071417.D\ECD2B.ch

#4 AR-1016-2



R.T.: 4.916 min
Delta R.T.: 0.000 min
Response: 38783246
Conc: 507.88 ng/ml


Time



#5 AR-1016-3


R.T.: 5.752 min
Delta R.T.: -0.002 min
Response: 30221252
Conc: 490.23 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042225




#5 AR-1016-3

R.T.: 5.093 min
Delta R.T.: 0.000 min
Response: 21921068
Conc: 519.33 ng/ml




#6 AR-1016-4

R.T.: 5.850 min
Delta R.T.: -0.002 min
Response: 25444869
Conc: 496.02 ng/ml



#6 AR-1016-4


R.T.: 5.135 min
Delta R.T.: -0.001 min
Response: 18087794
Conc: 524.50 ng/ml



#7 AR-1016-5


R.T.: 6.142 min
Delta R.T.: -0.003 min
Response: 24049539
Conc: 498.56 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042225




#7 AR-1016-5

R.T.: 5.349 min
Delta R.T.: 0.000 min
Response: 22819712
Conc: 513.41 ng/ml




#31 AR-1260-1

R.T.: 7.262 min
Delta R.T.: -0.003 min
Response: 46974832
Conc: 489.20 ng/ml



#31 AR-1260-1

R.T.: 6.384 min
Delta R.T.: 0.000 min
Response: 35744650
Conc: 494.40 ng/ml



R.T.: 7.515 min
Delta R.T.: -0.003 min
Response: 70731391
Conc: 494.36 ng/ml


Instrument :
ECD_P
ClientSampleId :
ICVPP042225

Time

Response_

Signal: PP071417.D\ECD2B.ch

#32 AR-1260-2




R.T.: 6.573 min
Delta R.T.: 0.000 min
Response: 43416292
Conc: 496.41 ng/ml

Time

Response_

Signal: PP071417.D\ECD1A.ch

#33 AR-1260-3




R.T.: 7.875 min
Delta R.T.: -0.003 min
Response: 58007061
Conc: 517.44 ng/ml

Time


Response_

Signal: PP071417.D\ECD2B.ch

#33 AR-1260-3




R.T.: 6.726 min
Delta R.T.: -0.001 min
Response: 39241127
Conc: 500.60 ng/ml



#34 AR-1260-4


R.T.: 8.098 min
Delta R.T.: -0.003 min
Response: 56235072
Conc: 498.25 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042225




#34 AR-1260-4

R.T.: 7.198 min
Delta R.T.: 0.000 min
Response: 31343464
Conc: 496.64 ng/ml



#35 AR-1260-5

R.T.: 8.418 min
Delta R.T.: -0.003 min
Response: 114665502
Conc: 506.09 ng/ml



#35 AR-1260-5

R.T.: 7.439 min
Delta R.T.: 0.000 min
Response: 75416996
Conc: 500.19 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071418.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 18:38
 Operator : YP\AJ
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP042225AR1242

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 02:23:37 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 02:23:12 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.813	98322267	73653264	48.740	48.976
2) SA Decachlor...	10.234	8.849	73973652	44775920	49.685	49.917
Target Compounds						
16) L4 AR-1242-1	5.669	4.898	28608484	23663233	489.076	493.809
17) L4 AR-1242-2	5.690	4.917	43367430	33449196	480.598	489.769
18) L4 AR-1242-3	5.753	5.094	25764272	18726327	483.300	486.783
19) L4 AR-1242-4	5.850	5.178	21798483	18331184	484.040	488.202
20) L4 AR-1242-5	6.581	5.702	23472650	21942094	485.668	487.912



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071418.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 18:38
Operator : YP\AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1242

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 23 02:23:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 02:23:12 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.515 min
Delta R.T.: 0.000 min
Response: 98322267
Conc: 48.74 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1242




#1 Tetrachloro-m-xylene

R.T.: 3.813 min
Delta R.T.: 0.000 min
Response: 73653264
Conc: 48.98 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.234 min
Delta R.T.: -0.001 min
Response: 73973652
Conc: 49.69 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.849 min
Delta R.T.: 0.000 min
Response: 44775920
Conc: 49.92 ng/ml



#16 AR-1242-1


R.T.: 5.669 min
Delta R.T.: 0.000 min
Response: 28608484
Conc: 489.08 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1242




#16 AR-1242-1

R.T.: 4.898 min
Delta R.T.: 0.000 min
Response: 23663233
Conc: 493.81 ng/ml




#17 AR-1242-2

R.T.: 5.690 min
Delta R.T.: 0.000 min
Response: 43367430
Conc: 480.60 ng/ml



#17 AR-1242-2


R.T.: 4.917 min
Delta R.T.: 0.000 min
Response: 33449196
Conc: 489.77 ng/ml



#18 AR-1242-3


R.T.: 5.753 min
Delta R.T.: 0.000 min
Response: 25764272
Conc: 483.30 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1242




#18 AR-1242-3

R.T.: 5.094 min
Delta R.T.: 0.000 min
Response: 18726327
Conc: 486.78 ng/ml




#19 AR-1242-4

R.T.: 5.850 min
Delta R.T.: 0.000 min
Response: 21798483
Conc: 484.04 ng/ml



#19 AR-1242-4


R.T.: 5.178 min
Delta R.T.: 0.000 min
Response: 18331184
Conc: 488.20 ng/ml



#20 AR-1242-5

R.T.: 6.581 min
Delta R.T.: 0.000 min
Response: 23472650
Conc: 485.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1242



#20 AR-1242-5

R.T.: 5.702 min
Delta R.T.: 0.000 min
Response: 21942094
Conc: 487.91 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071419.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 19:27
 Operator : YP\AJ
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP042225AR1248

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 01 07:42:53 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.812	98804591	74071948	49.835	52.584
2) SA Decachlor...	10.234	8.848	74699675	45534064	51.588	51.917
Target Compounds						
21) L5 AR-1248-1	5.667	4.897	22002964	19168563	512.425	530.888
22) L5 AR-1248-2	5.938	5.135	30310191	26184570	501.555	534.789
23) L5 AR-1248-3	6.141	5.177	34304891	27346403	513.666	535.307
24) L5 AR-1248-4	6.541	5.349	43159659	31642146	510.096	521.654
25) L5 AR-1248-5	6.580	5.742	40310718	30029077	507.383	515.159



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071419.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 19:27
Operator : YP\AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1248

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 01 07:42:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.514 min
Delta R.T.: -0.003 min
Response: 98804591
Conc: 49.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1248




#1 Tetrachloro-m-xylene

R.T.: 3.812 min
Delta R.T.: 0.000 min
Response: 74071948
Conc: 52.58 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.234 min
Delta R.T.: -0.004 min
Response: 74699675
Conc: 51.59 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.848 min
Delta R.T.: -0.001 min
Response: 45534064
Conc: 51.92 ng/ml



#21 AR-1248-1


R.T.: 5.667 min
Delta R.T.: 0.000 min
Response: 22002964
Conc: 512.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1248




#21 AR-1248-1

R.T.: 4.897 min
Delta R.T.: 0.000 min
Response: 19168563
Conc: 530.89 ng/ml




#22 AR-1248-2

R.T.: 5.938 min
Delta R.T.: -0.002 min
Response: 30310191
Conc: 501.55 ng/ml



#22 AR-1248-2


R.T.: 5.135 min
Delta R.T.: -0.001 min
Response: 26184570
Conc: 534.79 ng/ml



#23 AR-1248-3


R.T.: 6.141 min
Delta R.T.: 0.000 min
Response: 34304891
Conc: 513.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1248




#23 AR-1248-3

R.T.: 5.177 min
Delta R.T.: 0.000 min
Response: 27346403
Conc: 535.31 ng/ml



#24 AR-1248-4

R.T.: 6.541 min
Delta R.T.: 0.000 min
Response: 43159659
Conc: 510.10 ng/ml



#24 AR-1248-4


R.T.: 5.349 min
Delta R.T.: 0.000 min
Response: 31642146
Conc: 521.65 ng/ml

Response_

Signal: PP071419.D\ECD1A.ch

#25 AR-1248-5

ICAL Form



R.T.: 6.580 min
Delta R.T.: 0.000 min
Response: 40310718
Conc: 507.38 ng/ml


Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1248

Time

Response_

Signal: PP071419.D\ECD2B.ch

#25 AR-1248-5



R.T.: 5.742 min
Delta R.T.: -0.001 min
Response: 30029077
Conc: 515.16 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071420.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 19:43
 Operator : YP\AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP042225AR1254

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 01:28:28 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 19:27:22 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.513	3.813	98111190	74345146	50.931	50.716
2) SA Decachlor...	10.232	8.849	73988546	45066273	50.788	50.509
Target Compounds						
26) L6 AR-1254-1	6.517	5.702	41728109	42206013	509.183	461.771
27) L6 AR-1254-2	6.733	5.850	64122746	35938664	504.601	456.583
28) L6 AR-1254-3	7.097	6.254	64574948	56055056	505.381	473.005
29) L6 AR-1254-4	7.379	6.482	58986526	36616929	501.956	474.524
30) L6 AR-1254-5	7.796	6.900	55244441	47442452	512.803	467.562



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071420.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 19:43
Operator : YP\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 23 01:28:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 19:27:22 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.513 min
Delta R.T.: 0.000 min
Response: 98111190
Conc: 50.93 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1254




#1 Tetrachloro-m-xylene

R.T.: 3.813 min
Delta R.T.: 0.000 min
Response: 74345146
Conc: 50.72 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.232 min
Delta R.T.: -0.004 min
Response: 73988546
Conc: 50.79 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.849 min
Delta R.T.: -0.001 min
Response: 45066273
Conc: 50.51 ng/ml



#26 AR-1254-1


R.T.: 6.517 min
Delta R.T.: -0.003 min
Response: 41728109
Conc: 509.18 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1254




#26 AR-1254-1

R.T.: 5.702 min
Delta R.T.: -0.001 min
Response: 42206013
Conc: 461.77 ng/ml




#27 AR-1254-2

R.T.: 6.733 min
Delta R.T.: -0.002 min
Response: 64122746
Conc: 504.60 ng/ml



#27 AR-1254-2


R.T.: 5.850 min
Delta R.T.: -0.001 min
Response: 35938664
Conc: 456.58 ng/ml



#28 AR-1254-3


R.T.: 7.097 min
Delta R.T.: -0.002 min
Response: 64574948
Conc: 505.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1254




#28 AR-1254-3

R.T.: 6.254 min
Delta R.T.: -0.001 min
Response: 56055056
Conc: 473.01 ng/ml




#29 AR-1254-4

R.T.: 7.379 min
Delta R.T.: -0.003 min
Response: 58986526
Conc: 501.96 ng/ml



#29 AR-1254-4


R.T.: 6.482 min
Delta R.T.: -0.002 min
Response: 36616929
Conc: 474.52 ng/ml



#30 AR-1254-5

R.T.: 7.796 min
Delta R.T.: -0.002 min
Response: 55244441
Conc: 512.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1254



#30 AR-1254-5

R.T.: 6.900 min
Delta R.T.: -0.002 min
Response: 47442452
Conc: 467.56 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071421.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 20:16
 Operator : YP\AJ
 Sample : AR1268ICV500
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP042225AR1268

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/23/2025
 Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 04:47:52 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 04:41:50 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.516	3.814	98428428	75098106	49.661	51.289
2) SA Decachlor...	10.234	8.850	131.8E6	79464849	49.365	49.579
Target Compounds						
41) L9 AR-1268-1	8.733	7.722	157.9E6	98709054	493.783	486.283m
42) L9 AR-1268-2	8.827	7.788	132.3E6	84218105	493.704	487.500
43) L9 AR-1268-3	9.058	7.992	112.3E6	69254815	498.379	484.016
44) L9 AR-1268-4	9.477	8.286	48904008	30724632	493.945	483.689
45) L9 AR-1268-5	9.894	8.588	309.0E6	189.8E6	497.455	492.818

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071421.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 20:16
Operator : YP\AJ
Sample : AR1268ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1268


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 23 04:47:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 04:41:50 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.516 min
Delta R.T.: 0.001 min
Response: 98428428
Conc: 49.66 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1268


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#1 Tetrachloro-m-xylene

R.T.: 3.814 min
Delta R.T.: 0.000 min
Response: 75098106
Conc: 51.29 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.234 min
Delta R.T.: -0.002 min
Response: 131775212
Conc: 49.36 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.850 min
Delta R.T.: -0.001 min
Response: 79464849
Conc: 49.58 ng/ml




#41 AR-1268-1

R.T.: 8.733 min
Delta R.T.: 0.000 min
Response: 157884565
Conc: 493.78 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1268


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#41 AR-1268-1

R.T.: 7.722 min
Delta R.T.: 0.000 min
Response: 98709054
Conc: 486.28 ng/ml




#42 AR-1268-2

R.T.: 8.827 min
Delta R.T.: 0.000 min
Response: 132293136
Conc: 493.70 ng/ml



#42 AR-1268-2

R.T.: 7.788 min
Delta R.T.: 0.000 min
Response: 84218105
Conc: 487.50 ng/ml




#43 AR-1268-3

R.T.: 9.058 min
Delta R.T.: 0.000 min
Response: 112276971
Conc: 498.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1268


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#43 AR-1268-3

R.T.: 7.992 min
Delta R.T.: 0.000 min
Response: 69254815
Conc: 484.02 ng/ml




#44 AR-1268-4

R.T.: 9.477 min
Delta R.T.: 0.000 min
Response: 48904008
Conc: 493.94 ng/ml



#44 AR-1268-4

R.T.: 8.286 min
Delta R.T.: 0.000 min
Response: 30724632
Conc: 483.69 ng/ml




#45 AR-1268-5

R.T.: 9.894 min
Delta R.T.: 0.000 min
Response: 308950692
Conc: 497.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1268

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025



#45 AR-1268-5

R.T.: 8.588 min
Delta R.T.: 0.000 min
Response: 189821225
Conc: 492.82 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2033 SAS No.: Q2033 SDG NO.: Q2033

Continuing Calib Date: 05/15/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 04:05 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.66	5.67	5.57	5.77	0.01
Aroclor-1016-2	(2)	5.68	5.69	5.59	5.79	0.01
Aroclor-1016-3	(3)	5.74	5.75	5.65	5.85	0.01
Aroclor-1016-4	(4)	5.84	5.85	5.75	5.95	0.01
Aroclor-1016-5	(5)	6.13	6.15	6.05	6.25	0.02
Aroclor-1260-1	(1)	7.25	7.27	7.17	7.37	0.02
Aroclor-1260-2	(2)	7.50	7.52	7.42	7.62	0.02
Aroclor-1260-3	(3)	7.86	7.88	7.78	7.98	0.02
Aroclor-1260-4	(4)	8.09	8.10	8.00	8.20	0.01
Aroclor-1260-5	(5)	8.41	8.42	8.32	8.52	0.02
Tetrachloro-m-xylene		4.50	4.52	4.42	4.62	0.02
Decachlorobiphenyl		10.21	10.24	10.14	10.34	0.03



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2033 SAS No.: Q2033 SDG NO.: Q2033

Continuing Calib Date: 05/15/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 04:05 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.88	4.90	4.80	5.00	0.02
Aroclor-1016-2	(2)	4.90	4.92	4.82	5.02	0.02
Aroclor-1016-3	(3)	5.08	5.09	4.99	5.19	0.01
Aroclor-1016-4	(4)	5.12	5.14	5.04	5.24	0.02
Aroclor-1016-5	(5)	5.34	5.35	5.25	5.45	0.01
Aroclor-1260-1	(1)	6.37	6.39	6.29	6.49	0.02
Aroclor-1260-2	(2)	6.56	6.57	6.47	6.67	0.01
Aroclor-1260-3	(3)	6.71	6.73	6.63	6.83	0.02
Aroclor-1260-4	(4)	7.18	7.20	7.10	7.30	0.02
Aroclor-1260-5	(5)	7.42	7.44	7.34	7.54	0.02
Tetrachloro-m-xylene		3.80	3.81	3.71	3.91	0.01
Decachlorobiphenyl		8.83	8.85	8.75	8.95	0.02



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2033 SAS No.: Q2033 SDG NO.: Q2033

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL01 Date Analyzed: 05/15/2025

Lab Sample No.: AR1660CCC500 Data File : PP072076.D Time Analyzed: 04:05

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.656	5.571	5.771	435.980	500.000	-12.8
Aroclor-1016-2	5.678	5.593	5.793	497.580	500.000	-0.5
Aroclor-1016-3	5.741	5.654	5.854	498.750	500.000	-0.3
Aroclor-1016-4	5.838	5.752	5.952	490.730	500.000	-1.9
Aroclor-1016-5	6.129	6.045	6.245	483.490	500.000	-3.3
Aroclor-1260-1	7.249	7.165	7.365	482.510	500.000	-3.5
Aroclor-1260-2	7.503	7.418	7.618	481.610	500.000	-3.7
Aroclor-1260-3	7.861	7.777	7.977	497.180	500.000	-0.6
Aroclor-1260-4	8.086	8.002	8.202	475.890	500.000	-4.8
Aroclor-1260-5	8.405	8.322	8.522	506.250	500.000	1.3
Decachlorobiphenyl	10.213	10.137	10.337	52.590	50.000	5.2
Tetrachloro-m-xylene	4.504	4.417	4.617	49.550	50.000	-0.9



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2033 SAS No.: Q2033 SDG NO.: Q2033

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL01 Date Analyzed: 05/15/2025

Lab Sample No.: AR1660CCC500 Data File : PP072076.D Time Analyzed: 04:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.884	4.798	4.998	550.030	500.000	10.0
Aroclor-1016-2	4.902	4.816	5.016	517.740	500.000	3.5
Aroclor-1016-3	5.079	4.993	5.193	530.110	500.000	6.0
Aroclor-1016-4	5.121	5.036	5.236	495.050	500.000	-1.0
Aroclor-1016-5	5.335	5.250	5.450	533.270	500.000	6.7
Aroclor-1260-1	6.368	6.285	6.485	547.160	500.000	9.4
Aroclor-1260-2	6.557	6.473	6.673	549.140	500.000	9.8
Aroclor-1260-3	6.710	6.627	6.827	544.890	500.000	9.0
Aroclor-1260-4	7.180	7.098	7.298	562.890	500.000	12.6
Aroclor-1260-5	7.422	7.339	7.539	557.850	500.000	11.6
Decachlorobiphenyl	8.826	8.749	8.949	56.210	50.000	12.4
Tetrachloro-m-xylene	3.801	3.712	3.912	52.980	50.000	6.0

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051425\
 Data File : PP072076.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 May 2025 04:05
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/15/2025
 Supervised By :mohammad ahmed 05/19/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 15 04:33:24 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.504	3.801	98232592	74632414	49.546	52.982m
2) SA Decachlor...	10.213	8.826	76147279	49297981	52.588	56.209m
Target Compounds						
3) L1 AR-1016-1	5.656	4.884	29232579	29445208	435.983m	550.033m#
4) L1 AR-1016-2	5.678	4.902	51065295	39536725	497.575m	517.742m
5) L1 AR-1016-3	5.741	5.079	30746257	22376312	498.746	530.114m
6) L1 AR-1016-4	5.838	5.121	25173526	17072167	490.730	495.047
7) L1 AR-1016-5	6.129	5.335	23322539	23702411	483.485m	533.270
31) L7 AR-1260-1	7.249	6.368	46333177	39559015	482.514	547.157
32) L7 AR-1260-2	7.503	6.557	68907153	48028262	481.608	549.137
33) L7 AR-1260-3	7.861	6.710	55735589	42712605	497.177	544.890
34) L7 AR-1260-4	8.086	7.180	53710725	35524594	475.885	562.894
35) L7 AR-1260-5	8.405	7.422	114.7E6	84111823	506.250	557.855

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051425\
Data File : PP072076.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 May 2025 04:05
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/15/2025
Supervised By :mohammad ahmed 05/19/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 15 04:33:24 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.504 min
 Delta R.T.: -0.013 min
 Response: 98232592
 Conc: 49.55 ng/ml

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500


Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/15/2025
 Supervised By :mohammad ahmed 05/19/2025




#1 Tetrachloro-m-xylene

R.T.: 3.801 min
 Delta R.T.: -0.011 min
 Response: 74632414
 Conc: 52.98 ng/ml m




#2 Decachlorobiphenyl

R.T.: 10.213 min
 Delta R.T.: -0.024 min
 Response: 76147279
 Conc: 52.59 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.826 min
 Delta R.T.: -0.023 min
 Response: 49297981
 Conc: 56.21 ng/ml m




#3 AR-1016-1

R.T.: 5.656 min
 Delta R.T.: -0.015 min
 Response: 29232579
 Conc: 435.98 ng/ml

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500


Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/15/2025
 Supervised By :mohammad ahmed 05/19/2025




#3 AR-1016-1

R.T.: 4.884 min
 Delta R.T.: -0.014 min
 Response: 29445208
 Conc: 550.03 ng/ml m




#4 AR-1016-2

R.T.: 5.678 min
 Delta R.T.: -0.015 min
 Response: 51065295
 Conc: 497.58 ng/ml m



#4 AR-1016-2

R.T.: 4.902 min
 Delta R.T.: -0.014 min
 Response: 39536725
 Conc: 517.74 ng/ml m




#5 AR-1016-3

R.T.: 5.741 min
Delta R.T.: -0.014 min
Response: 30746257
Conc: 498.75 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/15/2025
Supervised By :mohammad ahmed 05/19/2025




#5 AR-1016-3

R.T.: 5.079 min
Delta R.T.: -0.014 min
Response: 22376312
Conc: 530.11 ng/ml




#6 AR-1016-4

R.T.: 5.838 min
Delta R.T.: -0.014 min
Response: 25173526
Conc: 490.73 ng/ml



#6 AR-1016-4

R.T.: 5.121 min
Delta R.T.: -0.015 min
Response: 17072167
Conc: 495.05 ng/ml




#7 AR-1016-5

R.T.: 6.129 min
Delta R.T.: -0.016 min
Response: 23322539
Conc: 483.49 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/15/2025
Supervised By :mohammad ahmed 05/19/2025




#7 AR-1016-5

R.T.: 5.335 min
Delta R.T.: -0.014 min
Response: 23702411
Conc: 533.27 ng/ml




#31 AR-1260-1

R.T.: 7.249 min
Delta R.T.: -0.015 min
Response: 46333177
Conc: 482.51 ng/ml



#31 AR-1260-1

R.T.: 6.368 min
Delta R.T.: -0.017 min
Response: 39559015
Conc: 547.16 ng/ml




#32 AR-1260-2

R.T.: 7.503 min
Delta R.T.: -0.015 min
Response: 68907153
Conc: 481.61 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/15/2025
Supervised By :mohammad ahmed 05/19/2025




#32 AR-1260-2

R.T.: 6.557 min
Delta R.T.: -0.017 min
Response: 48028262
Conc: 549.14 ng/ml




#33 AR-1260-3

R.T.: 7.861 min
Delta R.T.: -0.016 min
Response: 55735589
Conc: 497.18 ng/ml



#33 AR-1260-3

R.T.: 6.710 min
Delta R.T.: -0.017 min
Response: 42712605
Conc: 544.89 ng/ml




#34 AR-1260-4

R.T.: 8.086 min
Delta R.T.: -0.016 min
Response: 53710725
Conc: 475.89 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/15/2025
Supervised By :mohammad ahmed 05/19/2025




#34 AR-1260-4

R.T.: 7.180 min
Delta R.T.: -0.018 min
Response: 35524594
Conc: 562.89 ng/ml



#35 AR-1260-5

R.T.: 8.405 min
Delta R.T.: -0.017 min
Response: 114702594
Conc: 506.25 ng/ml



#35 AR-1260-5

R.T.: 7.422 min
Delta R.T.: -0.017 min
Response: 84111823
Conc: 557.85 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2033 SAS No.: Q2033 SDG NO.: Q2033

Continuing Calib Date: 05/15/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 09:15 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.66	5.67	5.57	5.77	0.01
Aroclor-1016-2	(2)	5.68	5.69	5.59	5.79	0.01
Aroclor-1016-3	(3)	5.74	5.75	5.65	5.85	0.01
Aroclor-1016-4	(4)	5.84	5.85	5.75	5.95	0.01
Aroclor-1016-5	(5)	6.13	6.15	6.05	6.25	0.02
Aroclor-1260-1	(1)	7.25	7.27	7.17	7.37	0.02
Aroclor-1260-2	(2)	7.50	7.52	7.42	7.62	0.02
Aroclor-1260-3	(3)	7.86	7.88	7.78	7.98	0.02
Aroclor-1260-4	(4)	8.09	8.10	8.00	8.20	0.01
Aroclor-1260-5	(5)	8.40	8.42	8.32	8.52	0.02
Tetrachloro-m-xylene		4.50	4.52	4.42	4.62	0.02
Decachlorobiphenyl		10.21	10.24	10.14	10.34	0.03



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2033 SAS No.: Q2033 SDG NO.: Q2033

Continuing Calib Date: 05/15/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 09:15 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.88	4.90	4.80	5.00	0.02
Aroclor-1016-2	(2)	4.90	4.92	4.82	5.02	0.02
Aroclor-1016-3	(3)	5.08	5.09	4.99	5.19	0.01
Aroclor-1016-4	(4)	5.12	5.14	5.04	5.24	0.02
Aroclor-1016-5	(5)	5.34	5.35	5.25	5.45	0.01
Aroclor-1260-1	(1)	6.37	6.39	6.29	6.49	0.02
Aroclor-1260-2	(2)	6.56	6.57	6.47	6.67	0.01
Aroclor-1260-3	(3)	6.71	6.73	6.63	6.83	0.02
Aroclor-1260-4	(4)	7.18	7.20	7.10	7.30	0.02
Aroclor-1260-5	(5)	7.42	7.44	7.34	7.54	0.02
Tetrachloro-m-xylene		3.80	3.81	3.71	3.91	0.01
Decachlorobiphenyl		8.82	8.85	8.75	8.95	0.03



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2033 SAS No.: Q2033 SDG NO.: Q2033

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL02 Date Analyzed: 05/15/2025

Lab Sample No.: AR1660CCC500 Data File : PP072091.D Time Analyzed: 09:15

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.656	5.571	5.771	450.250	500.000	-10.0
Aroclor-1016-2	5.678	5.593	5.793	483.560	500.000	-3.3
Aroclor-1016-3	5.741	5.654	5.854	493.590	500.000	-1.3
Aroclor-1016-4	5.839	5.752	5.952	490.260	500.000	-1.9
Aroclor-1016-5	6.130	6.045	6.245	456.190	500.000	-8.8
Aroclor-1260-1	7.249	7.165	7.365	469.730	500.000	-6.1
Aroclor-1260-2	7.503	7.418	7.618	460.660	500.000	-7.9
Aroclor-1260-3	7.861	7.777	7.977	473.220	500.000	-5.4
Aroclor-1260-4	8.086	8.002	8.202	445.780	500.000	-10.8
Aroclor-1260-5	8.404	8.322	8.522	501.640	500.000	0.3
Decachlorobiphenyl	10.213	10.137	10.337	51.770	50.000	3.5
Tetrachloro-m-xylene	4.504	4.417	4.617	50.390	50.000	0.8



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2033 SAS No.: Q2033 SDG NO.: Q2033

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL02 Date Analyzed: 05/15/2025

Lab Sample No.: AR1660CCC500 Data File : PP072091.D Time Analyzed: 09:15

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.884	4.798	4.998	555.850	500.000	11.2
Aroclor-1016-2	4.902	4.816	5.016	530.340	500.000	6.1
Aroclor-1016-3	5.079	4.993	5.193	490.540	500.000	-1.9
Aroclor-1016-4	5.121	5.036	5.236	494.300	500.000	-1.1
Aroclor-1016-5	5.335	5.250	5.450	532.270	500.000	6.5
Aroclor-1260-1	6.368	6.285	6.485	539.150	500.000	7.8
Aroclor-1260-2	6.556	6.473	6.673	518.460	500.000	3.7
Aroclor-1260-3	6.709	6.627	6.827	506.420	500.000	1.3
Aroclor-1260-4	7.179	7.098	7.298	525.790	500.000	5.2
Aroclor-1260-5	7.421	7.339	7.539	510.340	500.000	2.1
Decachlorobiphenyl	8.823	8.749	8.949	54.320	50.000	8.6
Tetrachloro-m-xylene	3.801	3.712	3.912	52.660	50.000	5.3

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051425\
 Data File : PP072091.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 May 2025 09:15
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/16/2025
 Supervised By :mohammad ahmed 05/19/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 15 09:37:24 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.504	3.801	99895442	74180871	50.385	52.661m
2) SA Decachlor...	10.213	8.823	74957690	47644189	51.766	54.323m
Target Compounds						
3) L1 AR-1016-1	5.656	4.884	30188896	29756525	450.246m	555.848m
4) L1 AR-1016-2	5.678	4.902	49627287	40499076	483.563m	530.345m
5) L1 AR-1016-3	5.741	5.079	30428231	20705781	493.587	490.537
6) L1 AR-1016-4	5.839	5.121	25149441	17046520	490.260	494.304
7) L1 AR-1016-5	6.130	5.335	22005903	23657873	456.191	532.268
31) L7 AR-1260-1	7.249	6.368	45105693	38980393	469.731	539.154
32) L7 AR-1260-2	7.503	6.556	65909592	45345401	460.657	518.462
33) L7 AR-1260-3	7.861	6.709	53049663	39696650	473.218	506.415
34) L7 AR-1260-4	8.086	7.179	50313067	33182704	445.782	525.786m
35) L7 AR-1260-5	8.404	7.421	113.7E6	76948359	501.643	510.344

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051425\
Data File : PP072091.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 May 2025 09:15
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/16/2025
Supervised By :mohammad ahmed 05/19/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 15 09:37:24 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.504 min
 Delta R.T.: -0.013 min
 Response: 99895442
 Conc: 50.39 ng/ml

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500


Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/16/2025
 Supervised By :mohammad ahmed 05/19/2025




#1 Tetrachloro-m-xylene

R.T.: 3.801 min
 Delta R.T.: -0.011 min
 Response: 74180871
 Conc: 52.66 ng/ml m




#2 Decachlorobiphenyl

R.T.: 10.213 min
 Delta R.T.: -0.025 min
 Response: 74957690
 Conc: 51.77 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.823 min
 Delta R.T.: -0.026 min
 Response: 47644189
 Conc: 54.32 ng/ml m




#3 AR-1016-1

R.T.: 5.656 min
Delta R.T.: -0.015 min
Response: 30188896
Conc: 450.25 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/16/2025
Supervised By :mohammad ahmed 05/19/2025




#3 AR-1016-1

R.T.: 4.884 min
Delta R.T.: -0.014 min
Response: 29756525
Conc: 555.85 ng/ml




#4 AR-1016-2

R.T.: 5.678 min
Delta R.T.: -0.015 min
Response: 49627287
Conc: 483.56 ng/ml



#4 AR-1016-2

R.T.: 4.902 min
Delta R.T.: -0.014 min
Response: 40499076
Conc: 530.34 ng/ml




#5 AR-1016-3

R.T.: 5.741 min
Delta R.T.: -0.014 min
Response: 30428231
Conc: 493.59 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/16/2025
Supervised By :mohammad ahmed 05/19/2025




#5 AR-1016-3

R.T.: 5.079 min
Delta R.T.: -0.014 min
Response: 20705781
Conc: 490.54 ng/ml




#6 AR-1016-4

R.T.: 5.839 min
Delta R.T.: -0.013 min
Response: 25149441
Conc: 490.26 ng/ml



#6 AR-1016-4

R.T.: 5.121 min
Delta R.T.: -0.015 min
Response: 17046520
Conc: 494.30 ng/ml




#7 AR-1016-5

R.T.: 6.130 min
Delta R.T.: -0.015 min
Response: 22005903
Conc: 456.19 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/16/2025
Supervised By :mohammad ahmed 05/19/2025




#7 AR-1016-5

R.T.: 5.335 min
Delta R.T.: -0.015 min
Response: 23657873
Conc: 532.27 ng/ml




#31 AR-1260-1

R.T.: 7.249 min
Delta R.T.: -0.015 min
Response: 45105693
Conc: 469.73 ng/ml



#31 AR-1260-1

R.T.: 6.368 min
Delta R.T.: -0.017 min
Response: 38980393
Conc: 539.15 ng/ml




#32 AR-1260-2

R.T.: 7.503 min
Delta R.T.: -0.016 min
Response: 65909592
Conc: 460.66 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/16/2025
Supervised By :mohammad ahmed 05/19/2025




#32 AR-1260-2

R.T.: 6.556 min
Delta R.T.: -0.017 min
Response: 45345401
Conc: 518.46 ng/ml




#33 AR-1260-3

R.T.: 7.861 min
Delta R.T.: -0.016 min
Response: 53049663
Conc: 473.22 ng/ml



#33 AR-1260-3

R.T.: 6.709 min
Delta R.T.: -0.018 min
Response: 39696650
Conc: 506.42 ng/ml




#34 AR-1260-4

R.T.: 8.086 min
Delta R.T.: -0.016 min
Response: 50313067
Conc: 445.78 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/16/2025
Supervised By :mohammad ahmed 05/19/2025




#34 AR-1260-4

R.T.: 7.179 min
Delta R.T.: -0.019 min
Response: 33182704
Conc: 525.79 ng/ml



#35 AR-1260-5

R.T.: 8.404 min
Delta R.T.: -0.017 min
Response: 113658833
Conc: 501.64 ng/ml



#35 AR-1260-5

R.T.: 7.421 min
Delta R.T.: -0.018 min
Response: 76948359
Conc: 510.34 ng/ml

Analytical Sequence

Client: ENTACT

SDG No.: Q2033

Project: 540 Degraw St, Brooklyn, NY - E9309

Instrument ID: ECD_P

GC Column: ZB-MR1

ID: 0.32 (mm)

Inst. Calib. Date(s): 04/22/2025

04/22/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	04/22/2025	10:13	PP071388.D	10.23	4.51
AR1660ICC1000	AR1660ICC1000	04/22/2025	10:29	PP071389.D	10.23	4.51
AR1660ICC750	AR1660ICC750	04/22/2025	10:45	PP071390.D	10.23	4.51
AR1660ICC500	AR1660ICC500	04/22/2025	11:02	PP071391.D	10.24	4.52
AR1660ICC250	AR1660ICC250	04/22/2025	11:18	PP071392.D	10.23	4.51
AR1660ICC050	AR1660ICC050	04/22/2025	11:34	PP071393.D	10.24	4.52
AR1221ICC500	AR1221ICC500	04/22/2025	11:51	PP071394.D	10.23	4.51
AR1232ICC500	AR1232ICC500	04/22/2025	12:07	PP071395.D	10.23	4.52
AR1242ICC1000	AR1242ICC1000	04/22/2025	12:23	PP071396.D	10.23	4.51
AR1242ICC750	AR1242ICC750	04/22/2025	12:39	PP071397.D	10.24	4.52
AR1242ICC500	AR1242ICC500	04/22/2025	12:56	PP071398.D	10.24	4.51
AR1242ICC250	AR1242ICC250	04/22/2025	13:12	PP071399.D	10.23	4.51
AR1242ICC050	AR1242ICC050	04/22/2025	13:28	PP071400.D	10.24	4.52
AR1248ICC1000	AR1248ICC1000	04/22/2025	13:45	PP071401.D	10.24	4.52
AR1248ICC750	AR1248ICC750	04/22/2025	14:01	PP071402.D	10.24	4.52
AR1248ICC500	AR1248ICC500	04/22/2025	14:17	PP071403.D	10.24	4.51
AR1248ICC250	AR1248ICC250	04/22/2025	14:33	PP071404.D	10.24	4.51
AR1248ICC050	AR1248ICC050	04/22/2025	14:50	PP071405.D	10.24	4.51
AR1254ICC1000	AR1254ICC1000	04/22/2025	15:06	PP071406.D	10.24	4.52
AR1254ICC750	AR1254ICC750	04/22/2025	15:22	PP071407.D	10.24	4.52
AR1254ICC500	AR1254ICC500	04/22/2025	15:38	PP071408.D	10.24	4.52
AR1254ICC250	AR1254ICC250	04/22/2025	15:55	PP071409.D	10.23	4.51
AR1254ICC050	AR1254ICC050	04/22/2025	16:11	PP071410.D	10.24	4.51
AR1262ICC500	AR1262ICC500	04/22/2025	16:27	PP071411.D	10.24	4.52
AR1268ICC1000	AR1268ICC1000	04/22/2025	16:44	PP071412.D	10.24	4.51
AR1268ICC750	AR1268ICC750	04/22/2025	17:00	PP071413.D	10.23	4.52
AR1268ICC500	AR1268ICC500	04/22/2025	17:16	PP071414.D	10.24	4.52
AR1268ICC250	AR1268ICC250	04/22/2025	17:33	PP071415.D	10.24	4.52
AR1268ICC050	AR1268ICC050	04/22/2025	17:49	PP071416.D	10.24	4.51
AR1660CCC500	AR1660CCC500	05/15/2025	04:05	PP072076.D	10.21	4.50
IBLK	IBLK	05/15/2025	05:27	PP072080.D	10.21	4.50
PB168005BL	PB168005BL	05/15/2025	05:43	PP072081.D	10.21	4.50
PB168005BS	PB168005BS	05/15/2025	05:59	PP072082.D	10.22	4.51
PB168005BSD	PB168005BSD	05/15/2025	06:16	PP072083.D	10.22	4.51
TW-WTS-08	Q2033-01	05/15/2025	08:10	PP072090.D	10.21	4.51
AR1660CCC500	AR1660CCC500	05/15/2025	09:15	PP072091.D	10.21	4.50
IBLK	IBLK	05/15/2025	10:22	PP072095.D	10.22	4.51

Analytical Sequence

Client: ENTACT

SDG No.: Q2033

Project: 540 Degraw St, Brooklyn, NY - E9309

Instrument ID: ECD_P

GC Column: ZB-MR2

ID: 0.32 (mm)

Inst. Calib. Date(s): 04/22/2025

04/22/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	04/22/2025	10:13	PP071388.D	8.85	3.81
AR1660ICC1000	AR1660ICC1000	04/22/2025	10:29	PP071389.D	8.85	3.81
AR1660ICC750	AR1660ICC750	04/22/2025	10:45	PP071390.D	8.85	3.81
AR1660ICC500	AR1660ICC500	04/22/2025	11:02	PP071391.D	8.85	3.81
AR1660ICC250	AR1660ICC250	04/22/2025	11:18	PP071392.D	8.85	3.81
AR1660ICC050	AR1660ICC050	04/22/2025	11:34	PP071393.D	8.85	3.81
AR1221ICC500	AR1221ICC500	04/22/2025	11:51	PP071394.D	8.85	3.81
AR1232ICC500	AR1232ICC500	04/22/2025	12:07	PP071395.D	8.85	3.81
AR1242ICC1000	AR1242ICC1000	04/22/2025	12:23	PP071396.D	8.85	3.81
AR1242ICC750	AR1242ICC750	04/22/2025	12:39	PP071397.D	8.85	3.81
AR1242ICC500	AR1242ICC500	04/22/2025	12:56	PP071398.D	8.85	3.81
AR1242ICC250	AR1242ICC250	04/22/2025	13:12	PP071399.D	8.85	3.81
AR1242ICC050	AR1242ICC050	04/22/2025	13:28	PP071400.D	8.85	3.81
AR1248ICC1000	AR1248ICC1000	04/22/2025	13:45	PP071401.D	8.85	3.81
AR1248ICC750	AR1248ICC750	04/22/2025	14:01	PP071402.D	8.85	3.81
AR1248ICC500	AR1248ICC500	04/22/2025	14:17	PP071403.D	8.85	3.81
AR1248ICC250	AR1248ICC250	04/22/2025	14:33	PP071404.D	8.85	3.81
AR1248ICC050	AR1248ICC050	04/22/2025	14:50	PP071405.D	8.85	3.81
AR1254ICC1000	AR1254ICC1000	04/22/2025	15:06	PP071406.D	8.85	3.81
AR1254ICC750	AR1254ICC750	04/22/2025	15:22	PP071407.D	8.85	3.81
AR1254ICC500	AR1254ICC500	04/22/2025	15:38	PP071408.D	8.85	3.81
AR1254ICC250	AR1254ICC250	04/22/2025	15:55	PP071409.D	8.85	3.81
AR1254ICC050	AR1254ICC050	04/22/2025	16:11	PP071410.D	8.85	3.81
AR1262ICC500	AR1262ICC500	04/22/2025	16:27	PP071411.D	8.85	3.81
AR1268ICC1000	AR1268ICC1000	04/22/2025	16:44	PP071412.D	8.85	3.81
AR1268ICC750	AR1268ICC750	04/22/2025	17:00	PP071413.D	8.85	3.81
AR1268ICC500	AR1268ICC500	04/22/2025	17:16	PP071414.D	8.85	3.81
AR1268ICC250	AR1268ICC250	04/22/2025	17:33	PP071415.D	8.85	3.81
AR1268ICC050	AR1268ICC050	04/22/2025	17:49	PP071416.D	8.85	3.81
AR1660CCC500	AR1660CCC500	05/15/2025	04:05	PP072076.D	8.83	3.80
IBLK	IBLK	05/15/2025	05:27	PP072080.D	8.83	3.80
PB168005BL	PB168005BL	05/15/2025	05:43	PP072081.D	8.82	3.80
PB168005BS	PB168005BS	05/15/2025	05:59	PP072082.D	8.82	3.80
PB168005BSD	PB168005BSD	05/15/2025	06:16	PP072083.D	8.83	3.80
TW-WTS-08	Q2033-01	05/15/2025	08:10	PP072090.D	8.83	3.80
AR1660CCC500	AR1660CCC500	05/15/2025	09:15	PP072091.D	8.82	3.80
IBLK	IBLK	05/15/2025	10:22	PP072095.D	8.83	3.80

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB168005BS

Contract: ENTA05

Lab Code: CHEM Case No.: Q2033 SAS No.: Q2033 SDG NO.: Q2033

Lab Sample ID: PB168005BS Date(s) Analyzed: 05/15/2025 05/15/2025

Instrument ID (1): ECD_P Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP072082.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.66	5.61	5.71	4.02	4.60	6.32
COLUMN 1	2	5.682	5.632	5.732	4.59		
	3	5.743	5.693	5.793	4.57		
	4	5.841	5.791	5.891	4.57		
	5	6.131	6.081	6.181	5.44		
	1	4.884	4.834	4.934	4.88		
COLUMN 2	2	4.902	4.852	4.952	4.80	4.90	
	3	5.079	5.029	5.129	4.97		
	4	5.121	5.071	5.171	5.05		
	5	5.335	5.285	5.385	4.74		
Aroclor-1260	1	7.251	7.201	7.301	4.57	4.30	13.04
COLUMN 1	2	7.506	7.456	7.556	4.62		
	3	7.863	7.813	7.913	4.02		
	4	8.087	8.037	8.137	4.19		
	5	8.407	8.357	8.457	4.29		
	1	6.367	6.317	6.417	5.06	4.90	
COLUMN 2	2	6.556	6.506	6.606	5.06		
	3	6.708	6.658	6.758	5.11		
	4	7.179	7.129	7.229	4.57		
	5	7.422	7.372	7.472	4.55		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB168005BSD

Contract: ENTA05

Lab Code: CHEM

Case No.: Q2033

SAS No.: Q2033

SDG NO.: Q2033

Lab Sample ID: PB168005BSD

Date(s) Analyzed: 05/15/2025

05/15/2025

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP072083.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.657	5.607	5.707	5.25	4.50	2.2
COLUMN 1	2	5.679	5.629	5.729	4.24		
	3	5.742	5.692	5.792	4.46		
	4	5.84	5.79	5.89	4.44		
	5	6.132	6.082	6.182	4.21		
	1	4.885	4.835	4.935	4.70		
COLUMN 2	2	4.903	4.853	4.953	4.83	4.60	
	3	5.081	5.031	5.131	4.45		
	4	5.122	5.072	5.172	4.58		
	5	5.336	5.286	5.386	4.57		
Aroclor-1260	1	7.25	7.2	7.3	4.50	4.30	10.99
COLUMN 1	2	7.504	7.454	7.554	4.59		
	3	7.862	7.812	7.912	4.02		
	4	8.086	8.036	8.136	4.16		
	5	8.406	8.356	8.456	4.27		
	1	6.369	6.319	6.419	4.97		
COLUMN 2	2	6.557	6.507	6.607	4.96	4.80	
	3	6.709	6.659	6.759	4.98		
	4	7.18	7.13	7.23	4.59		
	5	7.422	7.372	7.472	4.55		



QC SAMPLE DATA

Report of Analysis

Client:	ENTACT	Date Collected:	
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	
Client Sample ID:	PB168005BL	SDG No.:	Q2033
Lab Sample ID:	PB168005BL	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP072081.D	1	05/14/25 09:08	05/15/25 05:43	PB168005

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.0		30 (16) - 150 (158)	105%	SPK: 20
2051-24-3	Decachlorobiphenyl	21.9		30 (10) - 150 (173)	109%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051425\
 Data File : PP072081.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 May 2025 05:43
 Operator : YP\AJ
 Sample : PB168005BL
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PB168005BL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/15/2025
 Supervised By :mohammad ahmed 05/19/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 15 06:27:46 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.504	3.800	37379671	29532202	18.854	20.965m
2) SA Decachlor...	10.211	8.822	28444765	19159434	19.644	21.845m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051425\
Data File : PP072081.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 May 2025 05:43
Operator : YP\AJ
Sample : PB168005BL
Misc :
ALS Vial : 37 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
PB168005BL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/15/2025
Supervised By :mohammad ahmed 05/19/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 15 06:27:46 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm





#1 Tetrachloro-m-xylene

R.T.: 4.504 min
Delta R.T.: -0.013 min
Response: 37379671
Conc: 18.85 ng/ml

Instrument :

ECD_P

ClientSampleId :


PB168005BL

Manual Integrations

APPROVED


Reviewed By :Yogesh Patel 05/15/2025

Supervised By :mohammad ahmed 05/19/2025




#1 Tetrachloro-m-xylene

R.T.: 3.800 min
Delta R.T.: -0.012 min
Response: 29532202
Conc: 20.97 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.211 min
Delta R.T.: -0.026 min
Response: 28444765
Conc: 19.64 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.822 min
Delta R.T.: -0.027 min
Response: 19159434
Conc: 21.85 ng/ml m

Report of Analysis

Client:	ENTACT	Date Collected:	04/22/25
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	04/22/25
Client Sample ID:	PIBLK-PP071388.D	SDG No.:	Q2033
Lab Sample ID:	I.BLK-PP071388.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071388.D	1		04/22/25	PP042225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	17.1		70 (60) - 130 (140)	85%	SPK: 20
2051-24-3	Decachlorobiphenyl	17.6		70 (60) - 130 (140)	88%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071388.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 10:13
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 19:33:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 17:08:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.512	3.811	33865777	24238749	17.081	17.207
2) SA Decachlor...	10.233	8.847	25533987	15462668	17.634	17.630

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071388.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 10:13
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 19:33:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 17:08:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.512 min
Delta R.T.: -0.005 min
Response: 33865777
Conc: 17.08 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.811 min
Delta R.T.: -0.001 min
Response: 24238749
Conc: 17.21 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.233 min
Delta R.T.: -0.005 min
Response: 25533987
Conc: 17.63 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.847 min
Delta R.T.: -0.002 min
Response: 15462668
Conc: 17.63 ng/ml

Report of Analysis

Client:	ENTACT	Date Collected:	05/15/25
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	05/15/25
Client Sample ID:	PIBLK-PP072080.D	SDG No.:	Q2033
Lab Sample ID:	I.BLK-PP072080.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP072080.D	1		05/15/25	pp051425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	16.7		70 (60) - 130 (140)	83%	SPK: 20
2051-24-3	Decachlorobiphenyl	18.4		70 (60) - 130 (140)	92%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051425\
 Data File : PP072080.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 May 2025 05:27
 Operator : YP\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 I.BLK

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/15/2025
 Supervised By :mohammad ahmed 05/19/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 15 06:27:28 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.504	3.801	33097214	27368255	16.694	19.429m
2) SA Decachlor...	10.212	8.825	26591585	18322660	18.364	20.891m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051425\
Data File : PP072080.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 May 2025 05:27
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
I.BLK


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/15/2025
Supervised By :mohammad ahmed 05/19/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 15 06:27:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.504 min
Delta R.T.: -0.013 min
Response: 33097214
Conc: 16.69 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/15/2025
Supervised By :mohammad ahmed 05/19/2025




#1 Tetrachloro-m-xylene

R.T.: 3.801 min
Delta R.T.: -0.011 min
Response: 27368255
Conc: 19.43 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.212 min
Delta R.T.: -0.025 min
Response: 26591585
Conc: 18.36 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.825 min
Delta R.T.: -0.024 min
Response: 18322660
Conc: 20.89 ng/ml

Report of Analysis

Client:	ENTACT	Date Collected:	05/15/25
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	05/15/25
Client Sample ID:	PIBLK-PP072095.D	SDG No.:	Q2033
Lab Sample ID:	I.BLK-PP072095.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP072095.D	1		05/15/25	PP051425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	16.9		70 (60) - 130 (140)	85%	SPK: 20
2051-24-3	Decachlorobiphenyl	17.9		70 (60) - 130 (140)	90%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051425\
 Data File : PP072095.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 May 2025 10:22
 Operator : YP\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 I.BLK

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/16/2025
 Supervised By :mohammad ahmed 05/19/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 15 12:09:43 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.507	3.801	33527146	27037046	16.910	19.194m
2) SA Decachlor...	10.216	8.825	25977191	18413860	17.940	20.995m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051425\
Data File : PP072095.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 May 2025 10:22
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
I.BLK


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/16/2025
Supervised By :mohammad ahmed 05/19/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 15 12:09:43 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.507 min
Delta R.T.: -0.010 min
Response: 33527146
Conc: 16.91 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/16/2025
Supervised By :mohammad ahmed 05/19/2025




#1 Tetrachloro-m-xylene

R.T.: 3.801 min
Delta R.T.: -0.011 min
Response: 27037046
Conc: 19.19 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.216 min
Delta R.T.: -0.021 min
Response: 25977191
Conc: 17.94 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.825 min
Delta R.T.: -0.024 min
Response: 18413860
Conc: 21.00 ng/ml

Report of Analysis

Client:	ENTACT	Date Collected:	
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	
Client Sample ID:	PB168005BS	SDG No.:	Q2033
Lab Sample ID:	PB168005BS	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP072082.D	1	05/14/25 09:08	05/15/25 05:59	PB168005

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	4.90		0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
11096-82-5	Aroclor-1260	4.90		0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.2		30 (16) - 150 (158)	106%	SPK: 20
2051-24-3	Decachlorobiphenyl	23.0		30 (10) - 150 (173)	115%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051425\
 Data File : PP072082.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 May 2025 05:59
 Operator : YP\AJ
 Sample : PB168005BS
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :

ECD_P

ClientSampleId :

PB168005BS

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 05/15/2025

Supervised By :mohammad ahmed 05/19/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 15 06:28:06 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.507	3.801	38411540	29860823	19.374	21.198m
2) SA Decachlor...	10.215	8.824	29952930	20134857	20.686	22.957m
Target Compounds						
3) L1 AR-1016-1	5.660	4.884	26968884	26111318	402.222	487.756m
4) L1 AR-1016-2	5.682	4.902	47104369	36665498	458.980	480.143m
5) L1 AR-1016-3	5.743	5.079	28182901	20964774	457.165	496.673m
6) L1 AR-1016-4	5.841	5.121	23458186	17426477	457.291	505.321m
7) L1 AR-1016-5	6.131	5.335	26232671	21072523	543.813m	474.101
31) L7 AR-1260-1	7.251	6.367	43922701	36559677	457.411	505.672
32) L7 AR-1260-2	7.506	6.556	66103165	44274625	462.010	506.219
33) L7 AR-1260-3	7.863	6.708	45027940	40038765	401.662	510.779 #
34) L7 AR-1260-4	8.087	7.179	47265182	28832740	418.777	456.860m
35) L7 AR-1260-5	8.407	7.422	97226918	68587149	429.119	454.890

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051425\
Data File : PP072082.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 May 2025 05:59
Operator : YP\AJ
Sample : PB168005BS
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :

ECD_P

ClientSampleId :

PB168005BS

Manual Integrations



APPROVED


Reviewed By :Yogesh Patel 05/15/2025

Supervised By :mohammad ahmed 05/19/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 15 06:28:06 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.507 min
Delta R.T.: -0.010 min
Response: 38411540
Conc: 19.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB168005BS


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/15/2025
Supervised By :mohammad ahmed 05/19/2025




#1 Tetrachloro-m-xylene

R.T.: 3.801 min
Delta R.T.: -0.011 min
Response: 29860823
Conc: 21.20 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.215 min
Delta R.T.: -0.022 min
Response: 29952930
Conc: 20.69 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.824 min
Delta R.T.: -0.025 min
Response: 20134857
Conc: 22.96 ng/ml




#3 AR-1016-1

R.T.: 5.660 min
Delta R.T.: -0.011 min
Response: 26968884
Conc: 402.22 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB168005BS


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/15/2025
Supervised By :mohammad ahmed 05/19/2025




#3 AR-1016-1

R.T.: 4.884 min
Delta R.T.: -0.014 min
Response: 26111318
Conc: 487.76 ng/ml




#4 AR-1016-2

R.T.: 5.682 min
Delta R.T.: -0.011 min
Response: 47104369
Conc: 458.98 ng/ml



#4 AR-1016-2

R.T.: 4.902 min
Delta R.T.: -0.014 min
Response: 36665498
Conc: 480.14 ng/ml




#5 AR-1016-3

R.T.: 5.743 min
Delta R.T.: -0.011 min
Response: 28182901
Conc: 457.17 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB168005BS


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/15/2025
Supervised By :mohammad ahmed 05/19/2025




#5 AR-1016-3

R.T.: 5.079 min
Delta R.T.: -0.014 min
Response: 20964774
Conc: 496.67 ng/ml




#6 AR-1016-4

R.T.: 5.841 min
Delta R.T.: -0.011 min
Response: 23458186
Conc: 457.29 ng/ml



#6 AR-1016-4

R.T.: 5.121 min
Delta R.T.: -0.015 min
Response: 17426477
Conc: 505.32 ng/ml




#7 AR-1016-5

R.T.: 6.131 min
Delta R.T.: -0.014 min
Response: 26232671
Conc: 543.81 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB168005BS


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/15/2025
Supervised By :mohammad ahmed 05/19/2025




#7 AR-1016-5

R.T.: 5.335 min
Delta R.T.: -0.015 min
Response: 21072523
Conc: 474.10 ng/ml




#31 AR-1260-1

R.T.: 7.251 min
Delta R.T.: -0.013 min
Response: 43922701
Conc: 457.41 ng/ml



#31 AR-1260-1

R.T.: 6.367 min
Delta R.T.: -0.017 min
Response: 36559677
Conc: 505.67 ng/ml




#32 AR-1260-2

R.T.: 7.506 min
Delta R.T.: -0.013 min
Response: 66103165
Conc: 462.01 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB168005BS


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/15/2025
Supervised By :mohammad ahmed 05/19/2025




#32 AR-1260-2

R.T.: 6.556 min
Delta R.T.: -0.017 min
Response: 44274625
Conc: 506.22 ng/ml




#33 AR-1260-3

R.T.: 7.863 min
Delta R.T.: -0.014 min
Response: 45027940
Conc: 401.66 ng/ml



#33 AR-1260-3

R.T.: 6.708 min
Delta R.T.: -0.018 min
Response: 40038765
Conc: 510.78 ng/ml




#34 AR-1260-4

R.T.: 8.087 min
Delta R.T.: -0.014 min
Response: 47265182
Conc: 418.78 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB168005BS


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/15/2025
Supervised By :mohammad ahmed 05/19/2025




#34 AR-1260-4

R.T.: 7.179 min
Delta R.T.: -0.019 min
Response: 28832740
Conc: 456.86 ng/ml



#35 AR-1260-5

R.T.: 8.407 min
Delta R.T.: -0.015 min
Response: 97226918
Conc: 429.12 ng/ml



#35 AR-1260-5

R.T.: 7.422 min
Delta R.T.: -0.017 min
Response: 68587149
Conc: 454.89 ng/ml

Report of Analysis

Client:	ENTACT	Date Collected:	
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	
Client Sample ID:	PB168005BSD	SDG No.:	Q2033
Lab Sample ID:	PB168005BSD	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP072083.D	1	05/14/25 09:08	05/15/25 06:16	PB168005

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	4.60		0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
11096-82-5	Aroclor-1260	4.80		0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.0		30 (16) - 150 (158)	105%	SPK: 20
2051-24-3	Decachlorobiphenyl	23.4		30 (10) - 150 (173)	117%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051425\
 Data File : PP072083.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 May 2025 06:16
 Operator : YP\AJ
 Sample : PB168005BSD
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PB168005BSD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/15/2025
 Supervised By :mohammad ahmed 05/19/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 15 07:23:37 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.506	3.802	38050237	29575690	19.192	20.996m
2) SA Decachlor...	10.215	8.825	29729540	20512678	20.531	23.388m
Target Compounds						
3) L1 AR-1016-1	5.657	4.885	35218176	25148079	525.254m	469.763m
4) L1 AR-1016-2	5.679	4.903	43487248	36872403	423.735m	482.852m
5) L1 AR-1016-3	5.742	5.081	27520111	18804333	446.414	445.490
6) L1 AR-1016-4	5.840	5.122	22797728	15801220	444.416	458.193
7) L1 AR-1016-5	6.132	5.336	20302896	20307730	420.887	456.895
31) L7 AR-1260-1	7.250	6.369	43247525	35938154	450.380	497.075
32) L7 AR-1260-2	7.504	6.557	65641459	43422555	458.783	496.477
33) L7 AR-1260-3	7.862	6.709	45034859	39031709	401.724	497.932
34) L7 AR-1260-4	8.086	7.180	46918966	28983463	415.709	459.249
35) L7 AR-1260-5	8.406	7.422	96679569	68591762	426.703	454.921

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051425\
Data File : PP072083.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 May 2025 06:16
Operator : YP\AJ
Sample : PB168005BSD
Misc :
ALS Vial : 39 Sample Multiplier: 1

Instrument :

ECD_P

ClientSampleId :

PB168005BSD

Manual Integrations



APPROVED


Reviewed By :Yogesh Patel 05/15/2025

Supervised By :mohammad ahmed 05/19/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 15 07:23:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.506 min
Delta R.T.: -0.011 min
Response: 38050237
Conc: 19.19 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB168005BSD


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/15/2025
Supervised By :mohammad ahmed 05/19/2025




#1 Tetrachloro-m-xylene

R.T.: 3.802 min
Delta R.T.: -0.010 min
Response: 29575690
Conc: 21.00 ng/ml m




#2 Decachlorobiphenyl

R.T.: 10.215 min
Delta R.T.: -0.023 min
Response: 29729540
Conc: 20.53 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.825 min
Delta R.T.: -0.024 min
Response: 20512678
Conc: 23.39 ng/ml m




#3 AR-1016-1

R.T.: 5.657 min
Delta R.T.: -0.014 min
Response: 35218176
Conc: 525.25 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB168005BSD


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/15/2025
Supervised By :mohammad ahmed 05/19/2025




#3 AR-1016-1

R.T.: 4.885 min
Delta R.T.: -0.013 min
Response: 25148079
Conc: 469.76 ng/ml




#4 AR-1016-2

R.T.: 5.679 min
Delta R.T.: -0.014 min
Response: 43487248
Conc: 423.74 ng/ml



#4 AR-1016-2

R.T.: 4.903 min
Delta R.T.: -0.014 min
Response: 36872403
Conc: 482.85 ng/ml




#5 AR-1016-3

R.T.: 5.742 min
Delta R.T.: -0.012 min
Response: 27520111
Conc: 446.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB168005BSD


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/15/2025
Supervised By :mohammad ahmed 05/19/2025




#5 AR-1016-3

R.T.: 5.081 min
Delta R.T.: -0.012 min
Response: 18804333
Conc: 445.49 ng/ml




#6 AR-1016-4

R.T.: 5.840 min
Delta R.T.: -0.012 min
Response: 22797728
Conc: 444.42 ng/ml



#6 AR-1016-4

R.T.: 5.122 min
Delta R.T.: -0.014 min
Response: 15801220
Conc: 458.19 ng/ml




#7 AR-1016-5

R.T.: 6.132 min
Delta R.T.: -0.013 min
Response: 20302896
Conc: 420.89 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB168005BSD


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/15/2025
Supervised By :mohammad ahmed 05/19/2025




#7 AR-1016-5

R.T.: 5.336 min
Delta R.T.: -0.014 min
Response: 20307730
Conc: 456.89 ng/ml




#31 AR-1260-1

R.T.: 7.250 min
Delta R.T.: -0.015 min
Response: 43247525
Conc: 450.38 ng/ml



#31 AR-1260-1

R.T.: 6.369 min
Delta R.T.: -0.016 min
Response: 35938154
Conc: 497.08 ng/ml




#32 AR-1260-2

R.T.: 7.504 min
Delta R.T.: -0.014 min
Response: 65641459
Conc: 458.78 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB168005BSD


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/15/2025
Supervised By :mohammad ahmed 05/19/2025




#32 AR-1260-2

R.T.: 6.557 min
Delta R.T.: -0.016 min
Response: 43422555
Conc: 496.48 ng/ml




#33 AR-1260-3

R.T.: 7.862 min
Delta R.T.: -0.015 min
Response: 45034859
Conc: 401.72 ng/ml



#33 AR-1260-3

R.T.: 6.709 min
Delta R.T.: -0.017 min
Response: 39031709
Conc: 497.93 ng/ml




#34 AR-1260-4

R.T.: 8.086 min
Delta R.T.: -0.016 min
Response: 46918966
Conc: 415.71 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB168005BSD


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/15/2025
Supervised By :mohammad ahmed 05/19/2025




#34 AR-1260-4

R.T.: 7.180 min
Delta R.T.: -0.018 min
Response: 28983463
Conc: 459.25 ng/ml



#35 AR-1260-5

R.T.: 8.406 min
Delta R.T.: -0.016 min
Response: 96679569
Conc: 426.70 ng/ml



#35 AR-1260-5

R.T.: 7.422 min
Delta R.T.: -0.017 min
Response: 68591762
Conc: 454.92 ng/ml

Manual Integration Report

Sequence:	PP042225	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660ICC050	PP071393.D	AR-1260-3	yogesh	4/23/2025 7:32:38 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1660ICC050	PP071393.D	AR-1260-4	yogesh	4/23/2025 7:32:38 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1660ICC050	PP071393.D	AR-1260-5	yogesh	4/23/2025 7:32:38 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1254ICC750	PP071407.D	AR-1254-5	yogesh	4/23/2025 7:32:40 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1254ICC050	PP071410.D	Tetrachloro-m-xylene	yogesh	4/23/2025 7:32:41 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1268ICC1000	PP071412.D	AR-1268-1 #2	yogesh	4/23/2025 7:32:43 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1268ICC750	PP071413.D	AR-1268-1 #2	yogesh	4/23/2025 7:32:44 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1268ICC500	PP071414.D	AR-1268-1 #2	yogesh	4/23/2025 7:32:46 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1268ICC250	PP071415.D	AR-1268-1 #2	yogesh	4/23/2025 7:32:48 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1268ICC050	PP071416.D	AR-1268-1	yogesh	4/23/2025 7:32:50 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1268ICC050	PP071416.D	AR-1268-1 #2	yogesh	4/23/2025 7:32:50 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1268ICC050	PP071416.D	AR-1268-2	yogesh	4/23/2025 7:32:50 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1268ICV500	PP071421.D	AR-1268-1 #2	yogesh	4/23/2025 7:32:52 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:

PP042225

Instrument

ECD_p

Sample ID

File ID

Parameter

Review By

Review On

Supervised
By

Supervised On

Reason

Manual Integration Report

Sequence:	pp051425	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660CCC500	PP072033.D	AR-1016-1	yogesh	5/15/2025 7:57:58 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1660CCC500	PP072033.D	AR-1016-1 #2	yogesh	5/15/2025 7:57:58 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1660CCC500	PP072033.D	AR-1016-2	yogesh	5/15/2025 7:57:58 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1660CCC500	PP072033.D	AR-1016-2 #2	yogesh	5/15/2025 7:57:58 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1660CCC500	PP072033.D	AR-1260-4 #2	yogesh	5/15/2025 7:57:58 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1660CCC500	PP072033.D	Decachlorobiphenyl #2	yogesh	5/15/2025 7:57:58 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1242CCC500	PP072034.D	AR-1242-1 #2	yogesh	5/15/2025 7:58:00 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1242CCC500	PP072034.D	AR-1242-2 #2	yogesh	5/15/2025 7:58:00 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1242CCC500	PP072034.D	AR-1242-3 #2	yogesh	5/15/2025 7:58:00 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1242CCC500	PP072034.D	Decachlorobiphenyl #2	yogesh	5/15/2025 7:58:00 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1242CCC500	PP072034.D	Tetrachloro-m-xylene #2	yogesh	5/15/2025 7:58:00 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1248CCC500	PP072035.D	AR-1248-1 #2	yogesh	5/15/2025 7:58:03 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1248CCC500	PP072035.D	Decachlorobiphenyl	yogesh	5/15/2025 7:58:03 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software

Manual Integration Report

Sequence:	pp051425	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1248CCC500	PP072035.D	Decachlorobiphenyl #2	yogesh	5/15/2025 7:58:03 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1248CCC500	PP072035.D	Tetrachloro-m-xylene #2	yogesh	5/15/2025 7:58:03 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1254CCC500	PP072036.D	AR-1254-4 #2	yogesh	5/15/2025 7:58:53 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1254CCC500	PP072036.D	Decachlorobiphenyl #2	yogesh	5/15/2025 7:58:53 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1254CCC500	PP072036.D	Tetrachloro-m-xylene #2	yogesh	5/15/2025 7:58:53 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
I.BLK	PP072037.D	Decachlorobiphenyl #2	yogesh	5/15/2025 7:58:55 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1660CCC500	PP072048.D	AR-1016-1	yogesh	5/15/2025 7:59:24 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1660CCC500	PP072048.D	AR-1016-1 #2	yogesh	5/15/2025 7:59:24 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1660CCC500	PP072048.D	AR-1016-2	yogesh	5/15/2025 7:59:24 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1660CCC500	PP072048.D	AR-1016-2 #2	yogesh	5/15/2025 7:59:24 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1660CCC500	PP072048.D	AR-1016-3 #2	yogesh	5/15/2025 7:59:24 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1660CCC500	PP072048.D	AR-1260-4 #2	yogesh	5/15/2025 7:59:24 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1660CCC500	PP072048.D	Decachlorobiphenyl	yogesh	5/15/2025 7:59:24 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software

Manual Integration Report

Sequence:

pp051425

Instrument

ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660CCC500	PP072048.D	Decachlorobiphenyl #2	yogesh	5/15/2025 7:59:24 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1660CCC500	PP072048.D	Tetrachloro-m-xylene #2	yogesh	5/15/2025 7:59:24 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1242CCC500	PP072049.D	AR-1242-1 #2	yogesh	5/15/2025 7:59:26 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1242CCC500	PP072049.D	AR-1242-2 #2	yogesh	5/15/2025 7:59:26 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1242CCC500	PP072049.D	AR-1242-3 #2	yogesh	5/15/2025 7:59:26 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1242CCC500	PP072049.D	Decachlorobiphenyl #2	yogesh	5/15/2025 7:59:26 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1242CCC500	PP072049.D	Tetrachloro-m-xylene #2	yogesh	5/15/2025 7:59:26 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1248CCC500	PP072050.D	AR-1248-1	yogesh	5/15/2025 7:59:28 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1248CCC500	PP072050.D	AR-1248-1 #2	yogesh	5/15/2025 7:59:28 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1248CCC500	PP072050.D	AR-1248-3	yogesh	5/15/2025 7:59:28 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1248CCC500	PP072050.D	Decachlorobiphenyl	yogesh	5/15/2025 7:59:28 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1248CCC500	PP072050.D	Decachlorobiphenyl #2	yogesh	5/15/2025 7:59:28 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1254CCC500	PP072051.D	AR-1254-4 #2	yogesh	5/15/2025 7:59:30 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software

Manual Integration Report

Sequence:	pp051425	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1254CCC500	PP072051.D	Decachlorobiphenyl #2	yogesh	5/15/2025 7:59:30 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1254CCC500	PP072051.D	Tetrachloro-m-xylene #2	yogesh	5/15/2025 7:59:30 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
I.BLK	PP072052.D	Decachlorobiphenyl #2	yogesh	5/15/2025 7:59:32 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1660CCC500	PP072063.D	AR-1016-1	yogesh	5/15/2025 8:00:30 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1660CCC500	PP072063.D	AR-1016-1 #2	yogesh	5/15/2025 8:00:30 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1660CCC500	PP072063.D	AR-1016-2 #2	yogesh	5/15/2025 8:00:30 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1660CCC500	PP072063.D	Tetrachloro-m-xylene #2	yogesh	5/15/2025 8:00:30 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1242CCC500	PP072064.D	AR-1242-1 #2	yogesh	5/15/2025 8:00:32 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1242CCC500	PP072064.D	AR-1242-2 #2	yogesh	5/15/2025 8:00:32 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1242CCC500	PP072064.D	Decachlorobiphenyl	yogesh	5/15/2025 8:00:32 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1242CCC500	PP072064.D	Decachlorobiphenyl #2	yogesh	5/15/2025 8:00:32 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1242CCC500	PP072064.D	Tetrachloro-m-xylene #2	yogesh	5/15/2025 8:00:32 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1248CCC500	PP072065.D	AR-1248-1 #2	yogesh	5/15/2025 8:00:34 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software

Manual Integration Report

Sequence:	pp051425	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1248CCC500	PP072065.D	AR-1248-3	yogesh	5/15/2025 8:00:34 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1248CCC500	PP072065.D	Decachlorobiphenyl #2	yogesh	5/15/2025 8:00:34 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1248CCC500	PP072065.D	Tetrachloro-m-xylene #2	yogesh	5/15/2025 8:00:34 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1254CCC500	PP072066.D	Decachlorobiphenyl #2	yogesh	5/15/2025 8:00:36 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1254CCC500	PP072066.D	Tetrachloro-m-xylene #2	yogesh	5/15/2025 8:00:36 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
I.BLK	PP072067.D	Decachlorobiphenyl #2	yogesh	5/15/2025 8:00:38 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
I.BLK	PP072067.D	Tetrachloro-m-xylene #2	yogesh	5/15/2025 8:00:38 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1660CCC500	PP072076.D	AR-1016-1	yogesh	5/15/2025 8:00:55 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1660CCC500	PP072076.D	AR-1016-1 #2	yogesh	5/15/2025 8:00:55 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1660CCC500	PP072076.D	AR-1016-2	yogesh	5/15/2025 8:00:55 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1660CCC500	PP072076.D	AR-1016-2 #2	yogesh	5/15/2025 8:00:55 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1660CCC500	PP072076.D	AR-1016-3 #2	yogesh	5/15/2025 8:00:55 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1660CCC500	PP072076.D	AR-1016-5	yogesh	5/15/2025 8:00:55 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software

Manual Integration Report

Sequence:	pp051425	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660CCC500	PP072076.D	Decachlorobiphenyl #2	yogesh	5/15/2025 8:00:55 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1660CCC500	PP072076.D	Tetrachloro-m-xylene #2	yogesh	5/15/2025 8:00:55 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1242CCC500	PP072077.D	AR-1242-1 #2	yogesh	5/15/2025 8:00:56 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1242CCC500	PP072077.D	AR-1242-2 #2	yogesh	5/15/2025 8:00:56 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1242CCC500	PP072077.D	Decachlorobiphenyl #2	yogesh	5/15/2025 8:00:56 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1242CCC500	PP072077.D	Tetrachloro-m-xylene #2	yogesh	5/15/2025 8:00:56 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1248CCC500	PP072078.D	AR-1248-1 #2	yogesh	5/15/2025 8:00:58 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1248CCC500	PP072078.D	Decachlorobiphenyl #2	yogesh	5/15/2025 8:00:58 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1248CCC500	PP072078.D	Tetrachloro-m-xylene #2	yogesh	5/15/2025 8:00:58 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1254CCC500	PP072079.D	Decachlorobiphenyl	yogesh	5/15/2025 8:01:00 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1254CCC500	PP072079.D	Decachlorobiphenyl #2	yogesh	5/15/2025 8:01:00 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1254CCC500	PP072079.D	Tetrachloro-m-xylene #2	yogesh	5/15/2025 8:01:00 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
I.BLK	PP072080.D	Decachlorobiphenyl #2	yogesh	5/15/2025 8:01:02 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software

Manual Integration Report

Sequence:	pp051425	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
I.BLK	PP072080.D	Tetrachloro-m-xylene #2	yogesh	5/15/2025 8:01:02 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
PB168005BL	PP072081.D	Decachlorobiphenyl #2	yogesh	5/15/2025 8:01:04 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
PB168005BL	PP072081.D	Tetrachloro-m-xylene #2	yogesh	5/15/2025 8:01:04 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
PB168005BS	PP072082.D	AR-1016-1 #2	yogesh	5/15/2025 8:01:06 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
PB168005BS	PP072082.D	AR-1016-2 #2	yogesh	5/15/2025 8:01:06 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
PB168005BS	PP072082.D	AR-1016-3 #2	yogesh	5/15/2025 8:01:06 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
PB168005BS	PP072082.D	AR-1016-4 #2	yogesh	5/15/2025 8:01:06 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
PB168005BS	PP072082.D	AR-1016-5	yogesh	5/15/2025 8:01:06 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
PB168005BS	PP072082.D	AR-1260-4 #2	yogesh	5/15/2025 8:01:06 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
PB168005BS	PP072082.D	Decachlorobiphenyl #2	yogesh	5/15/2025 8:01:06 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
PB168005BS	PP072082.D	Tetrachloro-m-xylene #2	yogesh	5/15/2025 8:01:06 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
PB168005BSD	PP072083.D	AR-1016-1	yogesh	5/15/2025 8:01:08 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
PB168005BSD	PP072083.D	AR-1016-1 #2	yogesh	5/15/2025 8:01:08 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software

Manual Integration Report

Sequence:	pp051425	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PB168005BSD	PP072083.D	AR-1016-2	yogesh	5/15/2025 8:01:08 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
PB168005BSD	PP072083.D	AR-1016-2 #2	yogesh	5/15/2025 8:01:08 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
PB168005BSD	PP072083.D	Decachlorobiphenyl #2	yogesh	5/15/2025 8:01:08 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
PB168005BSD	PP072083.D	Tetrachloro-m-xylene #2	yogesh	5/15/2025 8:01:08 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1660CCC500	PP072091.D	AR-1016-1	yogesh	5/16/2025 8:09:39 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1660CCC500	PP072091.D	AR-1016-1 #2	yogesh	5/16/2025 8:09:39 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1660CCC500	PP072091.D	AR-1016-2	yogesh	5/16/2025 8:09:39 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1660CCC500	PP072091.D	AR-1016-2 #2	yogesh	5/16/2025 8:09:39 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1660CCC500	PP072091.D	AR-1260-4 #2	yogesh	5/16/2025 8:09:39 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1660CCC500	PP072091.D	Decachlorobiphenyl #2	yogesh	5/16/2025 8:09:39 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1660CCC500	PP072091.D	Tetrachloro-m-xylene #2	yogesh	5/16/2025 8:09:39 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1242CCC500	PP072092.D	AR-1242-1 #2	yogesh	5/16/2025 8:09:41 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1242CCC500	PP072092.D	AR-1242-2 #2	yogesh	5/16/2025 8:09:41 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software

Manual Integration Report

Sequence:	pp051425	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1242CCC500	PP072092.D	Decachlorobiphenyl #2	yogesh	5/16/2025 8:09:41 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1242CCC500	PP072092.D	Tetrachloro-m-xylene #2	yogesh	5/16/2025 8:09:41 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1248CCC500	PP072093.D	AR-1248-5 #2	yogesh	5/16/2025 8:09:43 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1248CCC500	PP072093.D	Decachlorobiphenyl #2	yogesh	5/16/2025 8:09:43 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1248CCC500	PP072093.D	Tetrachloro-m-xylene #2	yogesh	5/16/2025 8:09:43 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1254CCC500	PP072094.D	AR-1254-4 #2	yogesh	5/16/2025 8:09:45 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1254CCC500	PP072094.D	Decachlorobiphenyl #2	yogesh	5/16/2025 8:09:45 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
AR1254CCC500	PP072094.D	Tetrachloro-m-xylene #2	yogesh	5/16/2025 8:09:45 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
I.BLK	PP072095.D	Decachlorobiphenyl #2	yogesh	5/16/2025 8:09:47 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software
I.BLK	PP072095.D	Tetrachloro-m-xylene #2	yogesh	5/16/2025 8:09:47 AM	mohammad	5/20/2025 2:50:52	Peak Integrated by Software

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP042225

Review By	yogesh	Review On	4/22/2025 1:34:56 PM
Supervise By	mohammad	Supervise On	4/24/2025 6:33:42 AM
SubDirectory	PP042225	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP071387.D	22 Apr 2025 09:56	YP\AJ	Ok
2	I.BLK	PP071388.D	22 Apr 2025 10:13	YP\AJ	Ok
3	AR1660ICC1000	PP071389.D	22 Apr 2025 10:29	YP\AJ	Ok
4	AR1660ICC750	PP071390.D	22 Apr 2025 10:45	YP\AJ	Ok
5	AR1660ICC500	PP071391.D	22 Apr 2025 11:02	YP\AJ	Ok
6	AR1660ICC250	PP071392.D	22 Apr 2025 11:18	YP\AJ	Ok
7	AR1660ICC050	PP071393.D	22 Apr 2025 11:34	YP\AJ	Ok,M
8	AR1221ICC500	PP071394.D	22 Apr 2025 11:51	YP\AJ	Ok
9	AR1232ICC500	PP071395.D	22 Apr 2025 12:07	YP\AJ	Ok
10	AR1242ICC1000	PP071396.D	22 Apr 2025 12:23	YP\AJ	Ok
11	AR1242ICC750	PP071397.D	22 Apr 2025 12:39	YP\AJ	Ok
12	AR1242ICC500	PP071398.D	22 Apr 2025 12:56	YP\AJ	Ok
13	AR1242ICC250	PP071399.D	22 Apr 2025 13:12	YP\AJ	Ok
14	AR1242ICC050	PP071400.D	22 Apr 2025 13:28	YP\AJ	Ok
15	AR1248ICC1000	PP071401.D	22 Apr 2025 13:45	YP\AJ	Ok
16	AR1248ICC750	PP071402.D	22 Apr 2025 14:01	YP\AJ	Ok
17	AR1248ICC500	PP071403.D	22 Apr 2025 14:17	YP\AJ	Ok
18	AR1248ICC250	PP071404.D	22 Apr 2025 14:33	YP\AJ	Ok
19	AR1248ICC050	PP071405.D	22 Apr 2025 14:50	YP\AJ	Ok
20	AR1254ICC1000	PP071406.D	22 Apr 2025 15:06	YP\AJ	Ok
21	AR1254ICC750	PP071407.D	22 Apr 2025 15:22	YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP042225

Review By	yogesh	Review On	4/22/2025 1:34:56 PM
Supervise By	mohammad	Supervise On	4/24/2025 6:33:42 AM
SubDirectory	PP042225	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	AR1254ICC500	PP071408.D	22 Apr 2025 15:38	YP\AJ	Ok
23	AR1254ICC250	PP071409.D	22 Apr 2025 15:55	YP\AJ	Ok
24	AR1254ICC050	PP071410.D	22 Apr 2025 16:11	YP\AJ	Ok,M
25	AR1262ICC500	PP071411.D	22 Apr 2025 16:27	YP\AJ	Ok
26	AR1268ICC1000	PP071412.D	22 Apr 2025 16:44	YP\AJ	Ok,M
27	AR1268ICC750	PP071413.D	22 Apr 2025 17:00	YP\AJ	Ok,M
28	AR1268ICC500	PP071414.D	22 Apr 2025 17:16	YP\AJ	Ok,M
29	AR1268ICC250	PP071415.D	22 Apr 2025 17:33	YP\AJ	Ok,M
30	AR1268ICC050	PP071416.D	22 Apr 2025 17:49	YP\AJ	Ok,M
31	PP042225ICV500	PP071417.D	22 Apr 2025 18:05	YP\AJ	Ok
32	AR1242ICV500	PP071418.D	22 Apr 2025 18:38	YP\AJ	Ok
33	AR1248ICV500	PP071419.D	22 Apr 2025 19:27	YP\AJ	Ok
34	AR1254ICV500	PP071420.D	22 Apr 2025 19:43	YP\AJ	Ok
35	AR1268ICV500	PP071421.D	22 Apr 2025 20:16	YP\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP051425

Review By	yogesh	Review On	5/14/2025 10:30:08 AM
Supervise By	mohammad	Supervise On	5/20/2025 2:50:52 AM
SubDirectory	PP051425	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP072032.D	14 May 2025 08:51	YP\AJ	Ok
2	AR1660CCC500	PP072033.D	14 May 2025 09:07	YP\AJ	Ok,M
3	AR1242CCC500	PP072034.D	14 May 2025 09:25	YP\AJ	Ok,M
4	AR1248CCC500	PP072035.D	14 May 2025 09:41	YP\AJ	Ok,M
5	AR1254CCC500	PP072036.D	14 May 2025 09:58	YP\AJ	Ok,M
6	I.BLK	PP072037.D	14 May 2025 10:14	YP\AJ	Ok,M
7	PB167997BL	PP072038.D	14 May 2025 13:39	YP\AJ	Ok,M
8	PB167997BS	PP072039.D	14 May 2025 13:55	YP\AJ	Ok,M
9	Q2010-01	PP072040.D	14 May 2025 14:12	YP\AJ	Ok,M
10	Q2010-02	PP072041.D	14 May 2025 14:28	YP\AJ	Ok,M
11	Q2010-03	PP072042.D	14 May 2025 14:44	YP\AJ	Ok,M
12	Q2010-04	PP072043.D	14 May 2025 15:01	YP\AJ	Ok,M
13	Q2019-01	PP072044.D	14 May 2025 15:17	YP\AJ	Ok,M
14	Q2019-01MS	PP072045.D	14 May 2025 15:34	YP\AJ	Ok,M
15	Q2019-01MSD	PP072046.D	14 May 2025 15:50	YP\AJ	Ok,M
16	Q2020-01	PP072047.D	14 May 2025 16:06	YP\AJ	Ok,M
17	AR1660CCC500	PP072048.D	14 May 2025 17:12	YP\AJ	Ok,M
18	AR1242CCC500	PP072049.D	14 May 2025 17:44	YP\AJ	Ok,M
19	AR1248CCC500	PP072050.D	14 May 2025 18:01	YP\AJ	Ok,M
20	AR1254CCC500	PP072051.D	14 May 2025 18:17	YP\AJ	Ok,M
21	I.BLK	PP072052.D	14 May 2025 18:33	YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP051425

Review By	yogesh	Review On	5/14/2025 10:30:08 AM
Supervise By	mohammad	Supervise On	5/20/2025 2:50:52 AM
SubDirectory	PP051425	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q2005-01	PP072053.D	14 May 2025 18:50	YP\AJ	Ok
23	Q1990-01	PP072054.D	14 May 2025 19:06	YP\AJ	Dilution
24	Q1990-02	PP072055.D	14 May 2025 19:22	YP\AJ	Ok,M
25	Q2041-01	PP072056.D	14 May 2025 19:39	YP\AJ	Ok,M
26	Q2041-02	PP072057.D	14 May 2025 19:55	YP\AJ	Ok,M
27	Q2041-03	PP072058.D	14 May 2025 20:11	YP\AJ	Ok,M
28	Q2041-04	PP072059.D	14 May 2025 20:28	YP\AJ	Not Ok
29	Q2041-05	PP072060.D	14 May 2025 20:44	YP\AJ	Ok,M
30	Q2041-06	PP072061.D	14 May 2025 21:00	YP\AJ	Ok
31	Q2041-07	PP072062.D	14 May 2025 21:17	YP\AJ	Ok,M
32	AR1660CCC500	PP072063.D	14 May 2025 22:55	YP\AJ	Ok,M
33	AR1242CCC500	PP072064.D	14 May 2025 23:27	YP\AJ	Ok,M
34	AR1248CCC500	PP072065.D	14 May 2025 23:44	YP\AJ	Ok,M
35	AR1254CCC500	PP072066.D	15 May 2025 00:00	YP\AJ	Ok,M
36	I.BLK	PP072067.D	15 May 2025 00:16	YP\AJ	Ok,M
37	Q2041-08	PP072068.D	15 May 2025 00:33	YP\AJ	Ok
38	Q2042-01	PP072069.D	15 May 2025 00:49	YP\AJ	Ok,M
39	Q2042-02	PP072070.D	15 May 2025 01:05	YP\AJ	Ok
40	Q2042-04	PP072071.D	15 May 2025 01:22	YP\AJ	Ok,M
41	Q2042-05	PP072072.D	15 May 2025 01:38	YP\AJ	Ok,M
42	Q2022-01	PP072073.D	15 May 2025 01:55	YP\AJ	Ok,M
43	Q2022-02	PP072074.D	15 May 2025 02:11	YP\AJ	Ok,M
44	Q2022-04	PP072075.D	15 May 2025 02:27	YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP051425

Review By	yogesh	Review On	5/14/2025 10:30:08 AM
Supervise By	mohammad	Supervise On	5/20/2025 2:50:52 AM
SubDirectory	PP051425	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

45	AR1660CCC500	PP072076.D	15 May 2025 04:05	YP\AJ	Ok,M
46	AR1242CCC500	PP072077.D	15 May 2025 04:38	YP\AJ	Ok,M
47	AR1248CCC500	PP072078.D	15 May 2025 04:54	YP\AJ	Ok,M
48	AR1254CCC500	PP072079.D	15 May 2025 05:11	YP\AJ	Ok,M
49	I.BLK	PP072080.D	15 May 2025 05:27	YP\AJ	Ok,M
50	PB168005BL	PP072081.D	15 May 2025 05:43	YP\AJ	Ok,M
51	PB168005BS	PP072082.D	15 May 2025 05:59	YP\AJ	Ok,M
52	PB168005BSD	PP072083.D	15 May 2025 06:16	YP\AJ	Ok,M
53	PB168009BL	PP072084.D	15 May 2025 06:32	YP\AJ	Ok,M
54	PB168009BS	PP072085.D	15 May 2025 06:48	YP\AJ	Ok,M
55	Q2023-01	PP072086.D	15 May 2025 07:05	YP\AJ	Ok,M
56	Q2024-01	PP072087.D	15 May 2025 07:21	YP\AJ	Ok,M
57	Q2027-03	PP072088.D	15 May 2025 07:37	YP\AJ	Ok,M
58	Q2027-04	PP072089.D	15 May 2025 07:54	YP\AJ	Ok,M
59	Q2033-01	PP072090.D	15 May 2025 08:10	YP\AJ	Ok,NR
60	AR1660CCC500	PP072091.D	15 May 2025 09:15	YP\AJ	Ok,M
61	AR1242CCC500	PP072092.D	15 May 2025 09:32	YP\AJ	Ok,M
62	AR1248CCC500	PP072093.D	15 May 2025 09:48	YP\AJ	Ok,M
63	AR1254CCC500	PP072094.D	15 May 2025 10:05	YP\AJ	Ok,M
64	I.BLK	PP072095.D	15 May 2025 10:22	YP\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP042225

Review By	yogesh	Review On	4/22/2025 1:34:56 PM
Supervise By	mohammad	Supervise On	4/24/2025 6:33:42 AM
SubDirectory	PP042225	HP Acquire Method	HP Processing Method PP042225

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP071387.D	22 Apr 2025 09:56		YP\AJ	Ok
2	I.BLK	I.BLK	PP071388.D	22 Apr 2025 10:13		YP\AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PP071389.D	22 Apr 2025 10:29		YP\AJ	Ok
4	AR1660ICC750	AR1660ICC750	PP071390.D	22 Apr 2025 10:45		YP\AJ	Ok
5	AR1660ICC500	AR1660ICC500	PP071391.D	22 Apr 2025 11:02		YP\AJ	Ok
6	AR1660ICC250	AR1660ICC250	PP071392.D	22 Apr 2025 11:18		YP\AJ	Ok
7	AR1660ICC050	AR1660ICC050	PP071393.D	22 Apr 2025 11:34		YP\AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PP071394.D	22 Apr 2025 11:51		YP\AJ	Ok
9	AR1232ICC500	AR1232ICC500	PP071395.D	22 Apr 2025 12:07		YP\AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PP071396.D	22 Apr 2025 12:23		YP\AJ	Ok
11	AR1242ICC750	AR1242ICC750	PP071397.D	22 Apr 2025 12:39		YP\AJ	Ok
12	AR1242ICC500	AR1242ICC500	PP071398.D	22 Apr 2025 12:56		YP\AJ	Ok
13	AR1242ICC250	AR1242ICC250	PP071399.D	22 Apr 2025 13:12		YP\AJ	Ok
14	AR1242ICC050	AR1242ICC050	PP071400.D	22 Apr 2025 13:28		YP\AJ	Ok
15	AR1248ICC1000	AR1248ICC1000	PP071401.D	22 Apr 2025 13:45		YP\AJ	Ok
16	AR1248ICC750	AR1248ICC750	PP071402.D	22 Apr 2025 14:01		YP\AJ	Ok
17	AR1248ICC500	AR1248ICC500	PP071403.D	22 Apr 2025 14:17		YP\AJ	Ok
18	AR1248ICC250	AR1248ICC250	PP071404.D	22 Apr 2025 14:33		YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP042225

Review By	yogesh	Review On	4/22/2025 1:34:56 PM
Supervise By	mohammad	Supervise On	4/24/2025 6:33:42 AM
SubDirectory	PP042225	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248ICC050	AR1248ICC050	PP071405.D	22 Apr 2025 14:50		YPIAJ	Ok
20	AR1254ICC1000	AR1254ICC1000	PP071406.D	22 Apr 2025 15:06		YPIAJ	Ok
21	AR1254ICC750	AR1254ICC750	PP071407.D	22 Apr 2025 15:22		YPIAJ	Ok,M
22	AR1254ICC500	AR1254ICC500	PP071408.D	22 Apr 2025 15:38		YPIAJ	Ok
23	AR1254ICC250	AR1254ICC250	PP071409.D	22 Apr 2025 15:55		YPIAJ	Ok
24	AR1254ICC050	AR1254ICC050	PP071410.D	22 Apr 2025 16:11		YPIAJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PP071411.D	22 Apr 2025 16:27		YPIAJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PP071412.D	22 Apr 2025 16:44		YPIAJ	Ok,M
27	AR1268ICC750	AR1268ICC750	PP071413.D	22 Apr 2025 17:00		YPIAJ	Ok,M
28	AR1268ICC500	AR1268ICC500	PP071414.D	22 Apr 2025 17:16		YPIAJ	Ok,M
29	AR1268ICC250	AR1268ICC250	PP071415.D	22 Apr 2025 17:33		YPIAJ	Ok,M
30	AR1268ICC050	AR1268ICC050	PP071416.D	22 Apr 2025 17:49		YPIAJ	Ok,M
31	PP042225ICV500	ICVPP042225	PP071417.D	22 Apr 2025 18:05		YPIAJ	Ok
32	AR1242ICV500	ICVPP042225AR1242	PP071418.D	22 Apr 2025 18:38		YPIAJ	Ok
33	AR1248ICV500	ICVPP042225AR1248	PP071419.D	22 Apr 2025 19:27		YPIAJ	Ok
34	AR1254ICV500	ICVPP042225AR1254	PP071420.D	22 Apr 2025 19:43		YPIAJ	Ok
35	AR1268ICV500	ICVPP042225AR1268	PP071421.D	22 Apr 2025 20:16		YPIAJ	Ok,M

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP051425

Review By	yogesh	Review On	5/14/2025 10:30:08 AM
Supervise By	mohammad	Supervise On	5/20/2025 2:50:52 AM
SubDirectory	PP051425	HP Acquire Method	HP Processing Method PP042225

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP072032.D	14 May 2025 08:51		YP\AJ	Ok
2	AR1660CCC500	AR1660CCC500	PP072033.D	14 May 2025 09:07		YP\AJ	Ok,M
3	AR1242CCC500	AR1242CCC500	PP072034.D	14 May 2025 09:25	DCB high in 2nd column	YP\AJ	Ok,M
4	AR1248CCC500	AR1248CCC500	PP072035.D	14 May 2025 09:41	DCB high in 2nd column	YP\AJ	Ok,M
5	AR1254CCC500	AR1254CCC500	PP072036.D	14 May 2025 09:58	DCB high in 2nd column	YP\AJ	Ok,M
6	I.BLK	I.BLK	PP072037.D	14 May 2025 10:14		YP\AJ	Ok,M
7	PB167997BL	PB167997BL	PP072038.D	14 May 2025 13:39		YP\AJ	Ok,M
8	PB167997BS	PB167997BS	PP072039.D	14 May 2025 13:55		YP\AJ	Ok,M
9	Q2010-01	TP-10	PP072040.D	14 May 2025 14:12		YP\AJ	Ok,M
10	Q2010-02	TP-13	PP072041.D	14 May 2025 14:28		YP\AJ	Ok,M
11	Q2010-03	TP-14	PP072042.D	14 May 2025 14:44		YP\AJ	Ok,M
12	Q2010-04	TP-17	PP072043.D	14 May 2025 15:01		YP\AJ	Ok,M
13	Q2019-01	MH-K	PP072044.D	14 May 2025 15:17		YP\AJ	Ok,M
14	Q2019-01MS	MH-KMS	PP072045.D	14 May 2025 15:34		YP\AJ	Ok,M
15	Q2019-01MSD	MH-KMSD	PP072046.D	14 May 2025 15:50		YP\AJ	Ok,M
16	Q2020-01	TP-A	PP072047.D	14 May 2025 16:06	AR1254 Hit	YP\AJ	Ok,M
17	AR1660CCC500	AR1660CCC500	PP072048.D	14 May 2025 17:12		YP\AJ	Ok,M
18	AR1242CCC500	AR1242CCC500	PP072049.D	14 May 2025 17:44		YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP051425

Review By	yogesh	Review On	5/14/2025 10:30:08 AM
Supervise By	mohammad	Supervise On	5/20/2025 2:50:52 AM
SubDirectory	PP051425	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248CCC500	AR1248CCC500	PP072050.D	14 May 2025 18:01		YPIAJ	Ok,M
20	AR1254CCC500	AR1254CCC500	PP072051.D	14 May 2025 18:17		YPIAJ	Ok,M
21	I.BLK	I.BLK	PP072052.D	14 May 2025 18:33		YPIAJ	Ok,M
22	Q2005-01	252806	PP072053.D	14 May 2025 18:50	DCB low in 1st column	YPIAJ	Ok
23	Q1990-01	43025-A	PP072054.D	14 May 2025 19:06	AR1254 + 1242 Hit , need dilution	YPIAJ	Dilution
24	Q1990-02	43025-B	PP072055.D	14 May 2025 19:22	AR1254 + 1242 Hit (Need to confirm)	YPIAJ	Ok,M
25	Q2041-01	HIH359A-1-1	PP072056.D	14 May 2025 19:39	AR1254 Hit	YPIAJ	Ok,M
26	Q2041-02	HIH359A-1-2	PP072057.D	14 May 2025 19:55	AR1254 Hit	YPIAJ	Ok,M
27	Q2041-03	HIA9895-1-1	PP072058.D	14 May 2025 20:11		YPIAJ	Ok,M
28	Q2041-04	HIA9895-1-2	PP072059.D	14 May 2025 20:28	Surrogates not detected	YPIAJ	Not Ok
29	Q2041-05	250-1-1	PP072060.D	14 May 2025 20:44		YPIAJ	Ok,M
30	Q2041-06	250-1-2	PP072061.D	14 May 2025 21:00		YPIAJ	Ok
31	Q2041-07	250-2-1	PP072062.D	14 May 2025 21:17		YPIAJ	Ok,M
32	AR1660CCC500	AR1660CCC500	PP072063.D	14 May 2025 22:55		YPIAJ	Ok,M
33	AR1242CCC500	AR1242CCC500	PP072064.D	14 May 2025 23:27		YPIAJ	Ok,M
34	AR1248CCC500	AR1248CCC500	PP072065.D	14 May 2025 23:44		YPIAJ	Ok,M
35	AR1254CCC500	AR1254CCC500	PP072066.D	15 May 2025 00:00		YPIAJ	Ok,M
36	I.BLK	I.BLK	PP072067.D	15 May 2025 00:16		YPIAJ	Ok,M
37	Q2041-08	250-2-2	PP072068.D	15 May 2025 00:33	DCB high in 1st column	YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP051425

Review By	yogesh	Review On	5/14/2025 10:30:08 AM
Supervise By	mohammad	Supervise On	5/20/2025 2:50:52 AM
SubDirectory	PP051425	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

38	Q2042-01	HEI570L-1-1	PP072069.D	15 May 2025 00:49	AR1254 Hit	YPIAJ	Ok,M
39	Q2042-02	HEI570L-1-2	PP072070.D	15 May 2025 01:05		YPIAJ	Ok
40	Q2042-04	HEI570L-2-1	PP072071.D	15 May 2025 01:22	AR1254 Hit	YPIAJ	Ok,M
41	Q2042-05	HEI570L-2-2	PP072072.D	15 May 2025 01:38		YPIAJ	Ok,M
42	Q2022-01	336	PP072073.D	15 May 2025 01:55		YPIAJ	Ok,M
43	Q2022-02	COMP-1	PP072074.D	15 May 2025 02:11		YPIAJ	Ok,M
44	Q2022-04	COMP-2	PP072075.D	15 May 2025 02:27		YPIAJ	Ok,M
45	AR1660CCC500	AR1660CCC500	PP072076.D	15 May 2025 04:05		YPIAJ	Ok,M
46	AR1242CCC500	AR1242CCC500	PP072077.D	15 May 2025 04:38		YPIAJ	Ok,M
47	AR1248CCC500	AR1248CCC500	PP072078.D	15 May 2025 04:54		YPIAJ	Ok,M
48	AR1254CCC500	AR1254CCC500	PP072079.D	15 May 2025 05:11		YPIAJ	Ok,M
49	I.BLK	I.BLK	PP072080.D	15 May 2025 05:27		YPIAJ	Ok,M
50	PB168005BL	PB168005BL	PP072081.D	15 May 2025 05:43		YPIAJ	Ok,M
51	PB168005BS	PB168005BS	PP072082.D	15 May 2025 05:59		YPIAJ	Ok,M
52	PB168005BSD	PB168005BSD	PP072083.D	15 May 2025 06:16		YPIAJ	Ok,M
53	PB168009BL	PB168009BL	PP072084.D	15 May 2025 06:32		YPIAJ	Ok,M
54	PB168009BS	PB168009BS	PP072085.D	15 May 2025 06:48		YPIAJ	Ok,M
55	Q2023-01	HEATER-PAD	PP072086.D	15 May 2025 07:05		YPIAJ	Ok,M
56	Q2024-01	PL-02-051325	PP072087.D	15 May 2025 07:21		YPIAJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP051425

Review By	yogesh	Review On	5/14/2025 10:30:08 AM
Supervise By	mohammad	Supervise On	5/20/2025 2:50:52 AM
SubDirectory	PP051425	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

57	Q2027-03	B27-SOIL-SAMPLE	PP072088.D	15 May 2025 07:37		YPIAJ	Ok,M
58	Q2027-04	B28-SOIL-SAMPLE	PP072089.D	15 May 2025 07:54	AR1260 + 1254 Hit	YPIAJ	Ok,M
59	Q2033-01	TW-WTS-08	PP072090.D	15 May 2025 08:10		YPIAJ	Ok,NR
60	AR1660CCC500	AR1660CCC500	PP072091.D	15 May 2025 09:15		YPIAJ	Ok,M
61	AR1242CCC500	AR1242CCC500	PP072092.D	15 May 2025 09:32		YPIAJ	Ok,M
62	AR1248CCC500	AR1248CCC500	PP072093.D	15 May 2025 09:48		YPIAJ	Ok,M
63	AR1254CCC500	AR1254CCC500	PP072094.D	15 May 2025 10:05		YPIAJ	Ok,M
64	I.BLK	I.BLK	PP072095.D	15 May 2025 10:22		YPIAJ	Ok,M

M : Manual Integration

SOP ID:	M3510C,3580A-Extraction PCB-14		
Clean Up SOP #:	Acid Cleanup	Extraction Start Date :	05/14/2025
Matrix :	Water	Extraction Start Time :	09:08
Weigh By:	N/A	Extraction By:	RS
Balance check:	N/A	Extraction End Date :	05/14/2025
Balance ID:	N/A	Filter By:	RS
pH Strip Lot#:	E3880	Extraction End Time :	15:15
		Concentration By:	EH
		Supervisor By :	RUPESH
Extraction Method:	☒ Separatory Funnel ☐ Continous Liquid/Liquid ☐ Sonication ☐ Waste Dilution ☐ Soxhlet		

Standard Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	5000 PPB	PP24461
Surrogate	1.0ML	200 PPB	PP24460
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Methylene Chloride	N/A	E3930
Baked Na2SO4	N/A	EP2607
Hexane	N/A	E3933
H2SO4 1:1	N/A	EP2610
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

40 ML Vial lot# 03-40BTS723, Q2032 & Q2033 added at 10:36 AM. pH Adjusted with 1:1 H2SO4.


KD Bath ID:	WATER BATH-1,2	Envap ID:	NEVAP-02
KD Bath Temperature:	60 °C	Envap Temperature:	40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
5/14/25	RS (Ext Lab)	Y-PPESTIPUS
15:20	Preparation Group	Analysis Group

Analytical Method: M3510C,3580A-Extraction PCB-14

Concentration Date: 05/09/2025

Sample ID	Client Sample ID	Test	g / mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB168005BL	ABLK005	PCB	1000	6	RUPESH	ritesh	10			SEP-11
PB168005BS	ALCS005	PCB	1000	6	RUPESH	ritesh	10			12
PB168005BS D	ALCSD005	PCB	1000	6	RUPESH	ritesh	10			13
Q2005-01	252806	PCB	990	6	RUPESH	ritesh	10	P		14
Q2032-11	FB-05132025	PCB	980	6	RUPESH	ritesh	10	I		15
Q2033-01	TW-WTS-08	PCB	970	12	RUPESH	ritesh	10	G		16



RS
5/11/4

1680015 957
1680015 957
05-14-25

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q2005P WorkList ID : 189503 Department : Extraction Date : 05-14-2025 09:01:28

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2005-01	252806	Water	Herbicide	Cool 4 deg C	PSEG03	L41	05/09/2025	8151A
Q2005-01	252806	Water	PCB	Cool 4 deg C	PSEG03	L41	05/09/2025	8082A
Q2005-01	252806	Water	Pesticide-TCL	Cool 4 deg C	PSEG03	L41	05/09/2025	8081B

Date/Time 5/14/25 9:01
Raw Sample Received by: RS (Ext-96)
Raw Sample Relinquished by: CP

Date/Time 5/14/25 9:30
Raw Sample Received by: CP
Raw Sample Relinquished by: RS (Ext-96)

16805
500391

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q2032 WorkList ID : 189514 Department : Extraction Date : 05-14-2025 10:36:49

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2032-11	FB-05132025	Water	Herbicide	Cool 4 deg C	CAMP02	L41	05/13/2025	8151A
Q2032-11	FB-05132025	Water	PCB	Cool 4 deg C	CAMP02	L41	05/13/2025	8082A
Q2032-11	FB-05132025	Water	Pesticide-TCL	Cool 4 deg C	CAMP02	L41	05/13/2025	8081B
Q2033-01	TW-WTS-08	Water	PCB	Cool 4 deg C	ENTA05	L41	05/12/2025	8082A

Date/Time 5/14/25 10:36
Raw Sample Received by: RS (Ext-66)
Raw Sample Relinquished by: JO (SM)

Date/Time 5/14/25 11:00
Raw Sample Received by: JO (SM)
Raw Sample Relinquished by: RS (Ext-66)

Prep Standard - Chemical Standard Summary

Order ID : Q2033

Test : PCB

Prepbatch ID : PB168005,

Sequence ID/Qc Batch ID: pp051425,

Standard ID :

EP2607,EP2610,PP24329,PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369,PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387,PP24460,PP24461,

Chemical ID :

E3551,E3804,E3877,E3917,E3930,E3933,M6157,P11522,P12699,P12702,P12931,P12936,P12949,P12955,P12957,P13355,P13356,P13373,P13381,P13589,P13591,P13697,P13702,P13830,P13878,P13883,W3112,W3177,



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3923	Baked Sodium Sulfate	EP2607	04/25/2025	07/01/2025	RUPESHKUMAR SHAH	Extraction_SCALE_2 (EX-SC-2)	None	Riteshkumar Patel 04/25/2025
<u>FROM</u> 4000.00000gram of E3551 = Final Quantity: 4000.000 gram								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
314	1.1 H2SO4 SOLN	EP2610	05/07/2025	11/07/2025	RUPESHKUMAR SHAH	Extraction_SCALE_2 (EX-SC-2)	None	Riteshkumar Patel 05/07/2025
<u>FROM</u>	1000.00000ml of M6157 + 1000.00000ml of W3112 = Final Quantity: 2000.000 ml							

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
84	Pest/PCB Surrogate Stock 20 PPM	PP24329	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P13356 + 9.00000ml of W3177 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
202	AR1660 1000/100 ppb working solution 1st source	PP24330	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13697 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
203	AR1660 750 PPB STD	PP24331	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24330 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
204	AR1660 500 PPB STD	PP24332	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24330 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
205	AR1660 250 PPB STD	PP24333	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24330 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
206	AR1660 50 PPB STD	PP24334	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24332 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
213	AR1221 1000 PPB WORKING SOLUTION	PP24335	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13702 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1079	AR1221 750 PPB STD	PP24336	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24335 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
222	AR1221 500 PPB STD	PP24337	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24335 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1080	AR1221 250 PPB STD	PP24338	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24335 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1081	AR1221 50 PPB STD	PP24339	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24337 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
214	AR1232 1000 PPB WORKING SOLUTION	PP24340	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13878 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1063	AR1232 750 PPB STD	PP24341	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24340 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
223	AR1232 500 PPB STD	PP24342	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24340 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1064	AR1232 250 PPB STD	PP24343	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24340 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1065	AR1232 50 PPB STD	PP24344	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24342 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
215	AR1242 1000 PPB WORKING STD	PP24345	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P12931 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1067	AR1242 750 PPB STD	PP24346	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.75000ml of PP24345 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
224	AR1242 500 PPB STD	PP24347	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24345 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1068	AR1242 250 PPB STD	PP24348	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24345 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1069	AR1242 50 PPB STD	PP24349	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24347 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
216	AR1248 1000 PPB WORKING STD	PP24350	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P12936 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1075	AR1248 750 PPB STD	PP24351	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24350 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
225	AR1248 500 PPB STD	PP24352	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24350 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1076	AR1248 250 PPB STD	PP24353	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24350 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1077	AR1248 50 PPB STD	PP24354	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24352 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
217	AR1254 1000 PPB WORKING STD	PP24355	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13830 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1071	AR1254 750 PPB STD	PP24356	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24355 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
226	AR1254 500 PPB STD	PP24357	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24355 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1072	AR1254 250 PPB STD	PP24358	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24355 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1073	AR1254 50 PPB STD	PP24359	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24357 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1529	AR1262 1000 PPB Working Solution	PP24360	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13883 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3753	AR1262 750 PPB STD	PP24361	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24360 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1530	AR1262 500 PPB STD	PP24362	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24360 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3754	AR1262 250 PPB STD	PP24363	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24360 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3755	AR1262 50 PPB STD	PP24364	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24362 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1532	AR1268 1000 PPB Working Solution	PP24365	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13381 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3820	AR1268 750 PPB STD	PP24366	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24365 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1533	AR1268 500 PPB STD	PP24367	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24365 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3821	AR1268 250 PPB STD	PP24368	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24365 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3822	AR1268 50 PPB STD	PP24369	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24367 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
404	AR1660 100 PPM Stock Solution 2nd Source	PP24370	03/18/2025	09/18/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P12949 + 9.00000ml of E3804 = Final Quantity: 10.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
405	AR1660 1000/100 PPB ICV STD	PP24371	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 98.50000ml of W3177 + 0.50000ml of PP24329 + 1.00000ml of PP24370 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
406	AR1660 500 PPB ICV	PP24372	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24371 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3789	AR1221 1000 PPB WORKING SOL.2ND SOURCE(AGILENT)	PP24373	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P13373 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1886	AR1221 500 PPB ICV	PP24374	03/18/2025	08/12/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of E3877 + 0.50000ml of W3177 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1887	AR1232 1000 PPB Working Sol. 2nd Source	PP24375	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 1.00000ml of P12699 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1888	AR1232 500 PPB ICV	PP24376	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24375 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1889	AR1242 1000 PPB Working Sol. 2nd Source	PP24377	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P13589 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1891	AR1242 500 PPB ICV	PP24378	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24377 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1890	AR1248 1000 PPB Working Sol. 2nd Source	PP24379	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P13591 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1892	AR1248 500 PPB ICV	PP24380	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24379 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1893	AR1254 1000 PPB Working Sol. 2nd Source	PP24381	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P12957 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1894	AR1254 500 PPB ICV	PP24382	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24381 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3757	AR1262 1000 PPB Working Solution second source	PP24384	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P12702 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3758	AR1262 500 PPB STD ICV	PP24385	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24384 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3817	AR1268 1000 ppb Working Soln. 2nd source	PP24386	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P11522 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3823	AR1268 500 PPB STD ICV	PP24387	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24386 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
465	200 PPB Pest/PCB Surrogate Spike	PP24460	04/11/2025	10/03/2025	Abdul Mirza	None	None	Yogesh Patel
								04/16/2025

FROM 1.00000ml of P13355 + 999.00000ml of E3917 = Final Quantity: 1000.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3857	5000 PPB PCB SPIKE SOLUTION 2ND SOURCE	PP24461	04/11/2025	10/03/2025	Abdul Mirza	None	None	Yogesh Patel
								04/16/2025

FROM 0.50000ml of P12955 + 99.50000ml of E3917 = Final Quantity: 100.000 ml

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1	313201	07/01/2025	01/03/2024 / Rajesh	07/20/2023 / Rajesh	E3551

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	9005-05 / Acetone Ultra (cs/4x4L)	24E0761004	11/05/2025	10/01/2024 / Rajesh	09/25/2024 / Rajesh	E3804

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	243570	08/12/2025	02/12/2025 / Rajesh	02/12/2025 / Rajesh	E3877

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H2762008	10/03/2025	04/03/2025 / Rajesh	03/31/2025 / Rajesh	E3917

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	25A0262002	02/20/2026	05/02/2025 / RUPESH	03/09/2025 / RUPESH	E3930

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362005	11/05/2025	05/05/2025 / RUPESH	04/23/2025 / RUPESH	E3933

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9673-33 / Sulfuric Acid, Instra-Analyzed (cs/6c2.5L)	24i1262013	11/07/2025	05/07/2025 / RUPESH	02/18/2025 / Mohan	M6157

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-382-1 / Aroclor 1268	0006587800	09/18/2025	03/18/2025 / yogesh	02/21/2022 / Ankita	P11522

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc	91867 / Aroclor 1232 100 ug/mL	020823	09/18/2025	03/18/2025 / yogesh	08/07/2023 / Ankita	P12699

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc	x9166 / Aroclor 1262 100 ug/mL	060523	09/18/2025	03/18/2025 / yogesh	08/07/2023 / Ankita	P12702

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32009 / PCB Mix, Aroclor 1242, 1000ug/mL, Hexane, 1mL/ampul	a0203672	09/18/2025	03/18/2025 / yogesh	12/07/2023 / Ankita	P12931

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32010 / PCB Mix, Aroclor 1248, 1000ug/mL, Hexane, 1mL/ampul	a0202803	09/18/2025	03/18/2025 / yogesh	12/07/2023 / Ankita	P12936

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	20064 / Aroclor 1016/1260	022023	09/18/2025	03/18/2025 / yogesh	12/20/2023 / Yogesh	P12949

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	20064 / Aroclor 1016/1260	022023	10/11/2025	04/11/2025 / Abdul	12/20/2023 / Yogesh	P12955

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ Arochlor 1254	121823	04/03/2025	10/03/2024 / Ankita	12/20/2023 / Yogesh	P12957

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0206810	10/11/2025	04/11/2025 / Abdul	04/22/2024 / Abdul	P13355

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0206810	09/18/2025	03/18/2025 / yogesh	04/22/2024 / Abdul	P13356

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-292-1 / Aroclor 1221	0006783205	09/18/2025	03/18/2025 / yogesh	05/02/2024 / Ankita	P13373

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32410 / PCB Stock Solution, Aroclor 1268 Std, 1mL, Hexane	A0207475	09/18/2025	03/18/2025 / yogesh	05/03/2024 / Abdul	P13381

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-312-1 / Aroclor 1242	0006665550	09/18/2025	03/18/2025 / yogesh	10/14/2024 / Ankita	P13589

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-342-1 / Aroclor 1248	0006726317	09/18/2025	03/18/2025 / yogesh	10/14/2024 / Ankita	P13591

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32039 / PCB Mix, Aroclor 1016/1260, 1000ug/mL, hexane, 1mL/ampul	A0210629	09/18/2025	03/18/2025 / yogesh	10/17/2024 / yogesh	P13697

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32007 / PCB Mix, Aroclor 1221, 1000ug/mL, Hexane, 1mL/ampul	A0215270	09/18/2025	03/18/2025 / yogesh	10/17/2024 / yogesh	P13702

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32011 / PCB Mix, Aroclor 1254, 1000ug/mL, Hexane, 1mL/ampul	A0217391	09/18/2025	03/18/2025 / yogesh	12/09/2024 / Ankita	P13830

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32008 / PCB Mix, Aroclor 1232, 1000ug/mL, Hexane, 1mL/ampul	A0219655	09/18/2025	03/18/2025 / yogesh	01/23/2025 / Ankita	P13878

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32409 / PCB Stock Solution, Aroclor 1262 Std, 1mL, Hexane	A0220950	09/18/2025	03/18/2025 / yogesh	01/23/2025 / Ankita	P13883

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / lwona	07/03/2024 / lwona	W3112

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	24G1962003	08/22/2025	02/03/2025 / jignesh	01/31/2025 / jignesh	W3177



**PRODUCTOS
QUÍMICOS
MONTERREY, S.A. DE C.V.**

MIRADOR 201, COL. MIRADOR
MONTERREY, N.L. MEXICO
CP 64070
TEL +52 81 13 52 57 57
www.pqm.com.mx

CERTIFICATE OF ANALYSIS

PRODUCT :	SODIUM SULFATE CRYSTALS ANHYDROUS		
QUALITY :	ACS (CODE RMB3375)	FORMULA :	Na ₂ SO ₄
SPECIFICATION NUMBER :	6399	RELEASE DATE:	ABR/21/2023
LOT NUMBER :	313201		

TEST	SPECIFICATIONS	LOT VALUES
Assay (Na ₂ SO ₄)	Min. 99.0%	99.7 %
pH of a 5% solution at 25°C	5.2 - 9.2	6.1
Insoluble matter	Max. 0.01%	0.005 %
Loss on ignition	Max. 0.5%	0.1 %
Chloride (Cl)	Max. 0.001%	<0.001 %
Nitrogen compounds (as N)	Max. 5 ppm	<5 ppm
Phosphate (PO ₄)	Max. 0.001%	<0.001 %
Heavy metals (as Pb)	Max. 5 ppm	<5 ppm
Iron (Fe)	Max. 0.001%	<0.001 %
Calcium (Ca)	Max. 0.01%	0.002 %
Magnesium (Mg)	Max. 0.005%	0.001 %
Potassium (K)	Max. 0.008%	0.003 %
Extraction-concentration suitability	Passes test	Passes test
Appearance	Passes test	Passes test
Identification	Passes test	Passes test
Solubility and foreign matter	Passes test	Passes test
Retained on US Standard No. 10 sieve	Max. 1%	0.1 %
Retained on US Standard No. 60 sieve	Min. 94%	97.3 %
Through US Standard No. 60 sieve	Max. 5%	2.5 %
Through US Standard No. 100 sieve	Max. 10%	0.1 %

COMMENTS

QC: PhC Irma Belmares

If you need further details, please call our factory or contact our local distributor.

Recd. by R3 on 7/24/23 E 3551

RC-02-01, Ed. 3

Acetone
CMOS

avantor™



Material No.: 9005-05
Batch No.: 24E0761004
Manufactured Date: 2024-05-02
Retest Date: 2029-05-01
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	≥ 99.5 %	99.8 %
Color (APHA)	≤ 10	< 5
Residue after Evaporation	≤ 5 ppm	< 1 ppm
Titration Acid (μeq/g)	≤ 0.3	0.1
Titration Base (μeq/g)	≤ 0.5	0.1
Water (H ₂ O)	≤ 0.5 %	0.1 %
Solubility in H ₂ O	Passes Test	Passes Test
Chloride (Cl)	≤ 0.2 ppm	< 0.2 ppm
Phosphate (PO ₄)	≤ 0.05 ppm	< 0.05 ppm
Trace Impurities – Aluminum (Al)	≤ 50.0 ppb	< 5.0 ppb
Arsenic and Antimony (as As)	≤ 5.0 ppb	< 5.0 ppb
Trace Impurities – Barium (Ba)	≤ 20.0 ppb	< 1.0 ppb
Trace Impurities – Beryllium (Be)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Bismuth (Bi)	≤ 20.0 ppb	< 10.0 ppb
Trace Impurities – Boron (B)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Cadmium (Cd)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Calcium (Ca)	≤ 25.0 ppb	3.6 ppb
Trace Impurities – Chromium (Cr)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Cobalt (Co)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Copper (Cu)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Gallium (Ga)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Germanium (Ge)	≤ 10.0 ppb	< 10.0 ppb
Trace Impurities – Gold (Au)	≤ 20 ppb	< 5 ppb
Trace Impurities – Iron (Fe)	≤ 20.0 ppb	< 1.0 ppb
Trace Impurities – Lead (Pb)	≤ 10.0 ppb	< 10.0 ppb
Trace Impurities – Lithium (Li)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Magnesium (Mg)	≤ 20 ppb	< 1 ppb
Trace Impurities – Manganese (Mn)	≤ 10.0 ppb	< 1.0 ppb

Recd by RP on 9/25/24

>>> Continued on page 2 >>>

E 3804

Acetone
CMOS

 **avantor™**



Material No.: 9005-05
Batch No.: 24E0761004

Test	Specification	Result
Trace Impurities – Molybdenum (Mo)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Nickel (Ni)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Niobium (Nb)	≤ 50.0 ppb	< 1.0 ppb
Trace Impurities – Potassium (K)	≤ 10.0 ppb	< 10.0 ppb
Trace Impurities – Silicon (Si)	≤ 50 ppb	< 10 ppb
Trace Impurities – Silver (Ag)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Sodium (Na)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Strontium (Sr)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Tantalum (Ta)	≤ 50.0 ppb	< 5.0 ppb
Trace Impurities – Thallium (Tl)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Tin (Sn)	≤ 20.0 ppb	< 10.0 ppb
Trace Impurities – Titanium (Ti)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Vanadium (V)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Zinc (Zn)	≤ 20.0 ppb	7.9 ppb
Trace Impurities – Zirconium (Zr)	≤ 10.0 ppb	< 1.0 ppb
Particle Count – 0.5 µm and greater (Rion KS42AF)	≤ 100 par/ml	8 par/ml
Particle Count – 1.0 µm and greater (Rion KS42AF)	≤ 8 par/ml	2 par/ml

>>> Continued on page 3 >>>

Acetone
CMOS

 **avantor™**



Material No.: 9005-05
Batch No.: 24E0761004

Test	Specification	Result

For Microelectronic Use

Country of Origin: USA
Packaging Site: Paris Mfg Ctr & DC



Michelle Bales
Sr. Manager, Quality Assurance

Certificate of Analysis

1 Reagent Lane
 Fair Lawn, NJ 07410
 201.796.7100 tel
 201.796.1329 fax

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System
 Standard ISO9001:2015 by SAI Global Certificate Number CERT – 0120633

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the final formulator and end user to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

Catalog Number	H303	Quality Test / Release Date	11/07/2024
Lot Number	243570		
Description	HEXANES - OPTIMA		
Country of Origin	United States	Suggested Retest Date	Nov/2029
Chemical Origin	Organic - non animal		
BSE/TSE Comment	No animal products are used as starting raw material ingredients, or used in processing, including lubricants, processing aids, or any other material that might migrate to the finished product.		

N/A			
Result Name	Units	Specifications	Test Value
APPEARANCE		REPORT	Clear, colorless liquid
ASSAY (N-HEXANE)	%	>= 60	69
ASSAY (SUM C6 HYDROCARBONS)	%	>= 99.9	>99.9
COLOR	APHA	<= 5	<5
DENSITY AT 25 DEGREES C	GM/ML	Inclusive Between 0.653 - 0.673	0.669
EVAPORATION RESIDUE	ppm	<= 1	<1
FLUORESCENCE BACKGROUND	ppb	<= 1	<1
IDENTIFICATION	PASS/FAIL	= PASS TEST	PASS TEST
OPTICAL ABS AT 195 NM	ABS. UNITS	<= 1	0.74
OPTICAL ABS AT 210 NM	ABS. UNITS	<= 0.25	0.17
OPTICAL ABS AT 220 NM	ABS. UNITS	<= 0.07	0.05
OPTICAL ABS AT 254 NM	ABS. UNITS	<= 0.005	0.001
PESTICIDE RESIDUE ANALYSIS	NG/L	<= 10	<10
REFRACTIVE INDEX @ 25 DEG C		Inclusive Between 1.375 - 1.385	1.379
SUITABILITY FOR GC/MS		= PASS TEST	PASS TEST
SULFUR COMPOUNDS	%	<= 0.005	<0.005
THIOPHENE	PASS/FAIL	= PASS TEST	PASS TEST
WATER (H2O)	%	<= 0.01	<0.01
WATER-SOLUBLE TITRABLE ACID	MEQ/G	<= 0.0003	0.0001

Recd. by RP on 2/12/25

Harout Sahagian E3877

Harout Sahagian - Quality Control Manager - Fair Lawn

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of this catalog number listed above.

If there are any questions with this certificate, please call at (800) 227-6701.

*Based on suggested storage condition.

Acetone

BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor™



Material No.: 9254-03

Batch No.: 24H2762008

Manufactured Date: 2024-04-18

Expiration Date: 2027-04-18

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	100.0 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.0 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (μeq/g)	<= 0.3	0.2
Titration Base (μeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	<0.1 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use

MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC

Recd by RP on 03/31/25

E3917

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)



Material No.: 9266-A4
Batch No.: 25A0262002
Manufactured Date: 2024-11-21
Expiration Date: 2026-02-20
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	4
Assay (CH ₂ Cl ₂) (by GC, exclusive of preservative, corrected for water)	$\geq 99.8\%$	99.9%
Color (APHA)	≤ 10	10
Residue after Evaporation	≤ 1.0 ppm	0.8 ppm
Titration Acid (μ eq/g)	≤ 0.3	< 0.1
Chloride (Cl)	≤ 10 ppm	< 5 ppm
Water (by KF, coulometric)	$\leq 0.02\%$	$< 0.01\%$

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3930

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA, 19087, U.S.A. Phone 610.386.1700

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis

 **avantors**TM



Material No.: 9262-03

Batch No.: 25C0362005

Manufactured Date: 2025-01-29

Expiration Date: 2026-04-30

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	6
ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL)	≤ 5	5
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5 \%$	100.0 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95 \%$	100 %
Color (APHA)	≤ 10	10
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05 \%$	$< 0.01 \%$

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3933



Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

Sulfuric Acid
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis
Low Selenium

avantor™



M6157
MS

Material No.: 9673-33

Batch No.: 24I1262013

Manufactured Date: 2024-08-07

Retest Date: 2029-08-06

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
ACS – Assay (H ₂ SO ₄)	95.0 – 98.0 %	96.2 %
Appearance	Passes Test	Passes Test
ACS – Color (APHA)	<= 10	5
ACS – Residue after Ignition	<= 3 ppm	<1 ppm
ACS – Substances Reducing Permanganate(as SO ₂)	<= 2 ppm	<2 ppm
Ammonium (NH ₄)	<= 1 ppm	<1 ppm
Chloride (Cl)	<= 0.1 ppm	<0.1 ppm
Nitrate (NO ₃)	<= 0.2 ppm	0.1 ppm
Phosphate (PO ₄)	<= 0.5 ppm	<0.1 ppm
Trace Impurities – Aluminum (Al)	<= 30.0 ppb	<5.0 ppb
Arsenic & Antimony (as As)	<= 4.0 ppb	<2.0 ppb
Trace Impurities – Boron (B)	<= 10.0 ppb	<5.0 ppb
Trace Impurities – Cadmium (Cd)	<= 2.0 ppb	<1.0 ppb
Trace Impurities – Chromium (Cr)	<= 6.0 ppb	<1.0 ppb
Trace Impurities – Cobalt (Co)	<= 0.5 ppb	<0.3 ppb
Trace Impurities – Copper (Cu)	<= 1.0 ppb	<1.0 ppb
Trace Impurities – Gold (Au)	<= 10.0 ppb	<5.0 ppb
Heavy Metals (as Pb)	<= 500.0 ppb	<100.0 ppb
Trace Impurities – Iron (Fe)	<= 50.0 ppb	<1.0 ppb
Trace Impurities – Lead (Pb)	<= 0.5 ppb	<0.5 ppb
Trace Impurities – Magnesium (Mg)	<= 7.0 ppb	<1.0 ppb
Trace Impurities – Manganese (Mn)	<= 1.0 ppb	<1.0 ppb
Trace Impurities – Mercury (Hg)	<= 0.5 ppb	<0.1 ppb
Trace Impurities – Nickel (Ni)	<= 2.0 ppb	<0.3 ppb
Trace Impurities – Potassium (K)	<= 500.0 ppb	<10.0 ppb
Trace Impurities – Selenium (Se)	<= 50.0 ppb	7.2 ppb
Trace Impurities – Silicon (Si)	<= 100.0 ppb	12.8 ppb
Trace Impurities – Silver (Ag)	<= 1.0 ppb	<1.0 ppb

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

Sulfuric Acid

BAKER INSTRA-ANALYZED® Reagent

For Trace Metal Analysis

Low Selenium

 **avantors**™



Material No.: 9673-33

Batch No.: 2411262013

Test	Specification	Result
Trace Impurities – Sodium (Na)	≤ 500.0 ppb	< 5.0 ppb
Trace Impurities – Strontium (Sr)	≤ 5.0 ppb	< 1.0 ppb
Trace Impurities – Tin (Sn)	≤ 5.0 ppb	1.1 ppb
Trace Impurities – Zinc (Zn)	≤ 5.0 ppb	< 1.0 ppb

For Laboratory, Research, or Manufacturing Use

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC



Jamie Croak

Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC



Certificate of Analysis

P11518
↓
P11522
AJ
02/21/22

Product Name: Aroclor 1268 Standard

Product Number: PP-382-1

Lot Issue Date: 09-Feb-2021

Lot Number: 0006587800

Expiration Date: 31-Mar-2029

Description:

This analytical reference material (RM) was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	Concentration ± Uncertainty
Aroclor 1268	011100-14-4	RM00937	100.0 ± 0.5 µg/mL

Matrix: isooctane (2,2,4-trimethylpentane)

Storage Conditions: Store at Room Temperature (15° to 30°C).

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This RM was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Intended Use:

This RM is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Hazards:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this RM.


Expiration of Certification:

The certification of this RM is valid until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:


Monica Bourgeois
QMS Representative



ISO 17034 Cert
No. AR-1936

RM was produced in accordance with TUV USA Inc registered ISO 9001 Quality Management System. Cert # 56 100 18560026
Page: 1 of 1

www.agilent.com/quality/
CSD-QA-015.1



ISO 17025 Cert
No. AT-1937



CERTIFIED WEIGHT REPORT

Part Number: **91867**
 Lot Number: **020823**
 Description: **WP 037 - Aroclor 1232**
PCB Technical Mixture
 Expiration Date: **020833**
 Recommended Storage: **Ambient (20 °C)**
 Nominal Concentration ($\mu\text{g/mL}$): **100**
 NIST Test ID#: **6UTB**

Solvent(
Aceton


Weight(s) shown below were combined and diluted to (mL): **100.0**
 5E-05 Balance Uncertainty
 0.057 Flask Uncertainty

Compound	RM#	Lot Number	Nominal Conc ($\mu\text{g/mL}$)	Purity (%)	Uncertainty Purity	Target Weight (g)
1. Aroclor 1232	17	45-6A	100	100	0.5	0.01000

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measure Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments

GC3-M1 Analysis by Melissa Stonier
 Column ID SPB-608 30 meter X 0.53mm X 5 μm film thickness
 Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
 Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
 Oven Profile: Temp 1 = 150°C (Time 1 = 4 min), Temp 2 = 290°C (Time 2 = 13.5 min)
 Rate = 8°C/min, Total run time = 35 min
 Injector temp. = 200°C, FID Temp. = 300°C. FID Signal = Edaq Channel 1
 Standard injection = 1.5 μL , Range=3






110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No.: 32009 **Lot No.:** A0203672

Description: Aroclor® 1242 Standard

Aroclor® 1242 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size: 2 mL **Pkg Amt:** > 1 mL

Expiration Date: January 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

p12928

→

p12932

AJ

12/27/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty* (95% C.L.; K=2)
1	Aroclor 1242	53469-21-9	01141	—%	1,004.7 µg/mL	+/- 55.7515

Solvent: Hexane

CAS # 110-54-3

Purity 99%

* Expanded Uncertainty displayed in same units as Grav. Conc.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

0.2ul



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Russ Boothamer

Russ Boothamer - Operations Technician I

Date Mixed: 26-Oct-2023

Balance Serial #

B442140311

Jennifer Polino

Jennifer Polino - Operations Tech III - ARM QC

Date Passed: 06-Nov-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FW 80397




110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No.: 32010 **Lot No.:** A0202803

Description: Aroclor® 1248 Standard

Aroclor® 1248 Standard 1,000µg/mL, Hexane, 1mL/ampul

Container Size: 2 mL **Pkg Amt:** > 1 mL

Expiration Date: January 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

P1219697
P1219697
A1
12104123

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1248	12672-29-6	13897600	---%	1,001.7 µg/mL	+/- 55.5850

Solvent: Hexane

CAS # 110-54-3

Purity 99%

* Expanded Uncertainty displayed in same units as Grav. Conc.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

0.2ul



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Laith Clemente - Operations Technician I

Date Mixed: 03-Oct-2023

Balance Serial #

1128360905

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 09-Oct-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FW 80397



Certified Reference Material CRM

CERTIFIED WEIGHT REPORT

Part Number:
20064

Lot Number:
022023

Description:
CLP PCB'S - Aroclor Mix
Aroclors 1016 & 1260

Expiration Date:
022033

Recommended Storage:
Ambient (20 °C)

Nominal Concentration (µg/mL):
1000

NIST Test ID#:
6UTB

Solvent(s):
Hexane

Lot#
273615

Formulated By:
Benson Chan

Reviewed By:
Pedro L. Rentas

DATE
022023

DATE
022023

5E-05 Balance Uncertainty
0.010 Flask Uncertainty

Nominal Conc (µg/mL):
200.0

Weight(s) shown below were combined and diluted to (mL):

Part Number:
20064

Lot Number:
022023

Description:
CLP PCB'S - Aroclor Mix
Aroclors 1016 & 1260

Expiration Date:
022033

Recommended Storage:
Ambient (20 °C)

Nominal Concentration (µg/mL):
1000

NIST Test ID#:
6UTB

Solvent(s):
Hexane

Lot#
273615

Formulated By:
Benson Chan

Reviewed By:
Pedro L. Rentas

DATE
022023

DATE
022023

5E-05 Balance Uncertainty
0.010 Flask Uncertainty


Nominal Conc (µg/mL):
200.0

Weight(s) shown below were combined and diluted to (mL):

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Purity Uncertainty	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	CAS#	OSHA PEL (TWA)	LD50
1. Aroclor 1016	15	020491JC	1000	100	0.2	0.20004	0.20060	1002.8	4.0	12674-11-2	N/A	N/A
2. Aroclor 1260	21	020491JC	1000	100	0.2	0.20004	0.20081	1003.9	4.0	11086-82-5	0.5mg/m3	orl-rat 1315mg/kg

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments
GC3-M1 Analysis by Melissa Stortier
Column ID SPB-608 30 meter X 0.53mm X5µm film thickness
Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
Oven Profile: Temp 1 = 150°C (Time 1 = 4 min), Temp 2 = 280°C (Time 2 = 13.5 min)
Rate = 8°C/min, Total run time = 35 min
Injector temp. = 200°C, FID Temp. = 300°C, FID Signal = Etdaq Channel 1
Standard Injection = 1.5µL, Range=3





Certified Reference Material CRM

CERTIFIED WEIGHT REPORT

Part Number:
20064

Lot Number:
022023

Description:
CLP PCB'S - Aroclor Mix
Aroclors 1016 & 1260

Expiration Date:
022033

Recommended Storage:
Ambient (20 °C)

Nominal Concentration (µg/mL):
1000

NIST Test ID#:
6UTB

Solvent(s):
Hexane

Lot#
273615

Formulated By:
Benson Chan

Reviewed By:
Pedro L. Rentas

DATE
022023

DATE
022023

5E-05 Balance Uncertainty
0.010 Flask Uncertainty

Nominal Conc (µg/mL):
200.0

Weight(s) shown below were combined and diluted to (mL):

Part Number:
20064

Lot Number:
022023

Description:
CLP PCB'S - Aroclor Mix
Aroclors 1016 & 1260

Expiration Date:
022033

Recommended Storage:
Ambient (20 °C)

Nominal Concentration (µg/mL):
1000

NIST Test ID#:
6UTB

Solvent(s):
Hexane

Lot#
273615

Formulated By:
Benson Chan

Reviewed By:
Pedro L. Rentas

DATE
022023

DATE
022023

5E-05 Balance Uncertainty
0.010 Flask Uncertainty


Nominal Conc (µg/mL):
200.0

Weight(s) shown below were combined and diluted to (mL):

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Purity Uncertainty	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	CAS#	OSHA PEL (TWA)	LD50
1. Aroclor 1016	15	020491JC	1000	100	0.2	0.20004	0.20060	1002.8	4.0	12674-11-2	N/A	N/A
2. Aroclor 1260	21	020491JC	1000	100	0.2	0.20004	0.20081	1003.9	4.0	11096-82-5	0.5mg/m3	orl-rat 1315mg/kg

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments
GC3-M1 Analysis by Melissa Stortier
Column ID SPB-608 30 meter X 0.53mm X5µm film thickness
Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
Oven Profile: Temp 1 = 150°C (Time 1 = 4 min), Temp 2 = 280°C (Time 2 = 13.5 min)
Rate = 8°C/min, Total run time = 35 min
Injector temp. = 200°C, FID Temp. = 300°C, FID Signal = Etdaq Channel 1
Standard Injection = 1.5µL, Range=3





CERTIFIED WEIGHT REPORT

Part Number: 99139
Lot Number: 121823
Description: Aroclor 1254

Expiration Date: 121833
Recommended Storage: Ambient (20 °C)
Nominal Concentration (µg/mL): 100
NIST Test ID#: 6UTB

Volume(s) shown below were combined and diluted to (mL): 20.0
Note: Aroclor 1254 is a mix of isomers.

Solvent(s): Iso-octane
Lot#: 82227

Formulated By: Anthony Mahoney	121823	DATE
Reviewed By: Pedro L. Rentas	121823	DATE

P12956 / Y.P.
12/18/23
P12957


Note: Aroclor 1254 is a mix of isomers.

Compound		Part Number	Lot Number	Dilution Factor	Initial Vol. (mL)	Initial Uncertainty	Pipette (mL)	Conc.(µg/mL)	Initial Conc.(µg/mL)	Final Conc.(µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	CAS#	OSHA PEL (TWA)	LD50
1. Aroclor 1254		79100	121823	0.10	2.00	0.017	0.017	1003.3	100.1	11097-69-1	1.8	0.5mg/m3 (skin)	or-rat 1295mg/kg	

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
* All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments
GC3-M1 Analysis by Melissa Stonier
Column ID SPB-608 30 meter X 0.53mm X 5µm film thickness
Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
Oven Profile: Temp 1 = 150 °C (Time 1 = 4 min), Temp 2 = 280 °C (Time 2 = 13.5 min)
Rate = 8 °C/min, Total run time = 35 min
Injector temp. = 200 °C, FID Temp. = 300 °C, FID Signal = EDAQ Channel 1
Standard injection = 1.5µL, Range=3

Peak No.	Name	FID RT (min.)
1	Total Aroclor 1254	18.12





110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0206810

Description : Pesticide Surrogate Mix
Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P13348
↓
P13357
10
DAUF
04/25/2024

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.3 µg/mL	+/- 11.1143
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30638	99%	200.6 µg/mL	+/- 11.1298

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:


10 ml/min.

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 22-Jan-2024


Balance Serial # 1128360905


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P 13348
↓
P 13357
10


SAUF
04/25/2025



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0206810

Description : Pesticide Surrogate Mix
Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P13348
↓
P13357
10
DAUF
04/25/2024

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.3 µg/mL	+/- 11.1143
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30638	99%	200.6 µg/mL	+/- 11.1298

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:


10 ml/min.

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 22-Jan-2024


Balance Serial # 1128360905


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P 13348
↓
P 13357
10


SAUF
04/25/2025



ISO 17034

Reference Material Certificate

Product Information Sheet

Product Name: Aroclor 1221 Standard
Product Number: PP-292-1
Storage Conditions: Store at Room Temperature (15° to 30°C).

Lot Number: 0006783205
Lot Issue Date: 20-Feb-2024
Expiration Date: 31-Mar-2032

Component Name	Concentration	Uncertainty	CAS#	Analyte Lot
Aroclor 1221	100.3 ±	0.5 µg/mL	011104-28-2	NT01017

Matrix: isooctane (2,2,4-trimethylpentane)

Description:

This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above.

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This analytical reference standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:

This analytical reference standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Expiration of Certification:

The certification of this analytical reference standard is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

P13342
AJ
05106124
P13343

Page: 1 of 2

CSD-QA-015.2

ISO 17025
Cert No. AT-1937


250 Smith Street North Kingstown, Rhode Island 02852 www.agilent.com/quality



Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:


Monica Bourgeois
QMS Representative



ISO 17034
Cert No. AR-1936

RM was produced in accordance with the TUV/SUD registered ISO
9001:2015 Quality Management System. Cert# 951215321

Page: 2 of 2

www.agilent.com/quality/
CSD-QA-015.2

ISO 17025
Cert No. AT-1937

250 Smith Street North Kingstown, Rhode Island 02852 www.agilent.com/quality



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32410 **Lot No.:** A0207475

Description : Aroclor® 1268 Standard

Aroclor® 1268 Standard 1,000 µg/mL, 1mL/ampul, Hexane

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : May 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1268	11100-14-4	10947000	---%	1,000.0 µg/mL	+/- 55.4925

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P 13380
↓
P 13381
↓
SAZU
05/6/2024

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

Split ratio 500:1

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maye
Michael Maye - Operations Tech I

Date Mixed: 06-Feb-2024

Balance Serial # B442140311

Dylan Murphy
Dylan Murphy - Operations Technician I


Date Passed: 09-Feb-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P13380 } (2)
↓
P13381
1

[Signature]
05/6/2024

ISO 17034



Agilent

Trusted Answers

Reference Material Certificate

Product Name: Aroclor 1242 Standard

Lot Number: 0006665550

Product Number: PP-312-1

Lot Issue Date: 08-Feb-2022

Storage Conditions: Store at Room Temperature (15° to 30°C).

Expiration Date: 31-Jan-2027

Component Name	CERTIFIED VALUES		CAS#	Analyte Lot
	Concentration	Expanded Uncertainty		
Aroclor 1242	100.4	± 0.5 µg/mL	053469-21-9	NT01020

Matrix: isooctane (2,2,4-trimethylpentane)

Description:

This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above.

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This analytical reference standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:

This analytical reference standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Expiration of Certification:

The certification of this analytical reference standard is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

P13589
↓
P13590AJ
10/14/24

ISO 17034



Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:

Monica Bourgeois
QMS Representative



ISO 17034 Cert
No. AR-1936

RM was produced in accordance with the TUV/SUD registered ISO 9001:2015
Quality Management System. Cert# 951215321

Page: 2 of 2

www.agilent.com/quality/

CSD-QA-015.1



ISO 17025
Cert No. AT-

Reference Material Certificate
Product Information Sheet

Product Name: Aroclor 1248 Standard

Lot Number: 0006726317

Product Number: PP-342-1

Lot Issue Date: 27-Jan-2023

Storage Conditions: Store at Room Temperature (15° to 30°C).

Expiration Date: 28-Feb-2031

Component Name	Concentration	Uncertainty	CAS#	Analyte Lot
Aroclor 1248	100.3 ±	0.5 µg/mL	012672-29-6	NT01582

Matrix: isooctane (2,2,4-trimethylpentane)

Description:

This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material (RM) standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above. Purity values are taken from approved vendor raw material certificates.

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This analytical reference (RM) standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:

This analytical reference (RM) standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Expiration of Certification:

The certification of this analytical reference standard (RM) is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

P13591
↓
P13592

AS
10/14/2024

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:



Monica Bourgeois
QMS Representative



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32039 **Lot No.:** A0210629

Description : Aroclor® 1016/1260 Mix
Aroclor® 1016/1260 Mix 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

P13697
↓
P13701 } Y.P.
10/19/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1016	12674-11-2	07	----%	1,005.3 µg/mL	+/- 55.7809
2	Aroclor 1260	11096-82-5	1320657	----%	1,000.0 µg/mL	+/- 55.4850

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:


10 ml/min.

Inj. Vol

0.2µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 22-Apr-2024

Balance Serial # B442140311


Dillan Murphy - Operations Technician I

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32007 **Lot No.:** A0215270

Description : Aroclor® 1221 Standard

Aroclor® 1221 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

P13902
P13703 } Y.P.
10/17/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1221	11104-28-2	14969200	----%	1,005.0 µg/mL	+/- 55.7700

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maye
Michael Maye - Operations Tech I

Date Mixed: 16-Aug-2024

Balance Serial # 1128360905

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 20-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32011 **Lot No.:** A0217391

Description : Aroclor® 1254 Standard

Aroclor® 1254 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : January 31, 2031 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1254	11097-69-1	124-191-B	----%	1,004.7 µg/mL	+/- 55.7515

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P13830
↓
P13832
AJ
12/09/24

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maye
Michael Maye - Operations Tech I

Date Mixed: 02-Oct-2024

Balance Serial # C322230531

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 07-Oct-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32008

Lot No.: A0219655

Description : Aroclor® 1232 Standard

Aroclor® 1232 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : March 31, 2031

Storage: 25°C nominal

Handling:

This product contains PCBs.

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1232	11141-16-5	15665-01	---%	1,007.0 µg/mL	+/- 55.8810

Solvent: Hexane

CAS # 110-54-3

Purity 99%

* Expanded Uncertainty displayed in same units as Grav. Conc.

P13878



AJ
01/28/25

P13880

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maye
Michael Maye - Operations Tech I

Date Mixed: 02-Dec-2024 **Balance Serial #** C322230531

Brittany Federhiko
Brittany Federhiko - Operations Tech I

Date Passed: 05-Dec-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397




110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No.: 32409 **Lot No.:** A0220950

Description: Aroclor® 1262 Standard

Aroclor® 1262 Standard 1,000 µg/mL, 1mL/ampul, Hexane

Container Size: 2 mL **Pkg Amt:** > 1 mL

Expiration Date: April 30, 2031 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1262	37324-23-5	10849100	----%	1,002.0 µg/mL	+/- 55.6035

Solvent: Hexane

CAS # 110-54-3
Purity 99%

* Expanded Uncertainty displayed in same units as Grav. Conc.

P13882

↓

AJ
01/28/25

P13883

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Tom Suckal - Mix Technician

Date Mixed: 09-Jan-2025 Balance Serial # C322230531



Britiany Federenko - Operations Tech I

Date Passed: 14-Jan-2025

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis



Material No.: 9262-03
Batch No.: 24G1962003
Manufactured Date: 2024-05-23
Expiration Date: 2025-08-22
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	3
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	1
ECD-Sensitive Impurities (as Ethylene Dibromide) - Single Impurity Peak (ng/mL)	≤ 5	1
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5 \%$	99.7 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95 \%$	98 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05 \%$	< 0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Croak
Director Quality Operations, Bioscience Production



SHIPPING DOCUMENTS



284 Sheffield Street, Mountainside, NJ 07092
(908) 789-8900 Fax: (908) 788-9222
www.chemtech.net

CHAIN OF CUSTODY RECORD

Alliance Project Number:

Q2033

COC Number:

Page 1 of 1

CLIENT INFORMATION		PROJECT INFORMATION		BILLING INFORMATION																					
COMPANY: ENTACT, LLC		PROJECT NAME: 540 Degraw St Brooklyn, NY		BILL TO: ENTACT, LLC PO# E9309																					
ADDRESS: 150 Bay Street, Suite 806		PROJECT #: E9309 LOCATION: Brooklyn, NY		ADDRESS: 999 Oakmont Plaza Drive, Suite 300																					
CITY: Jersey City STATE: NJ ZIP: 07302		PROJECT-MANAGER: Austin Farmerie		CITY: Westmont STATE: IL ZIP: 60559																					
ATTENTION: Austin Farmerie		E-MAIL: afarmerie@entact.com		ATTENTION: Wendy Murray PHONE: 800-936-8228																					
PHONE: 412-716-1366 FAX:		PHONE: 412-716-1366 FAX:																							
DATA TURNAROUND INFORMATION		DATA DELIVERABLE INFORMATION		ANALYSIS																					
FAX: 5 DAYS* HARD COPY: 5 DAYS* EDD: 5 DAYS* * TO BE APPROVED BY ALLIANCE STANDARD TURNAROUND TIME IS 10 BUSINESS DAYS		☐ RESULTS ONLY ☐ USEPA C1 ☐ RESULTS + QC ☐ New York B" ☐ New Jersey REDUCED ☐ New York State ASP "A" ☐ New Jersey CLP ☐ Other _____ ☐ EDD Format		<table border="1"><tr><td>SVOC-TCL</td><td>Flash Point</td><td>PCB</td><td>BOD5</td><td>TSS</td><td>VOC-TCLVOC-A-10</td><td>Metals ICP-TAL</td><td></td><td></td><td></td></tr><tr><td>1</td><td>2</td><td>3</td><td>4</td><td>5</td><td>6</td><td>7</td><td>8</td><td>9</td><td></td></tr></table>		SVOC-TCL	Flash Point	PCB	BOD5	TSS	VOC-TCLVOC-A-10	Metals ICP-TAL				1	2	3	4	5	6	7	8	9	
SVOC-TCL	Flash Point	PCB	BOD5	TSS	VOC-TCLVOC-A-10	Metals ICP-TAL																			
1	2	3	4	5	6	7	8	9																	
				PRESERVATIVES		COMMENTS																			
CHEMTECH SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE	SAMPLE COLLECTION	# of Bottles																				
			COMP	GRAB	DATE	TIME	1	2	3	4	5	6	7	8	9										
1.	TW-WTS-08	Surface Water		X	5/12	15:00	X	X	X	X	X	X	X			ph 1.0									
2.																									
3.																									
4.																									
5.																									
6.																									
7.																									
8.																									
9.																									
10.																									
SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY																									
RELINQUISHED BY SAMPLER		DATE/TIME		RECEIVED BY		Conditions of bottles or coolers at receipt:		☐ Compliant ☐ Non Compliant ☐ Cooler Temp		3.12		Ice in Cooler?: yes													
1. Austin Farmerie		05/12/25 15:00		1. [Signature]		Comments: 1547																			
RELINQUISHED BY		DATE/TIME		RECEIVED BY																					
2. [Signature]				2. [Signature]																					
RELINQUISHED BY		DATE/TIME		RECEIVED FOR LAB BY																					
3. [Signature]		5-13-2025		3. [Signature]																					
						Page _____ of _____		SHIPPED VIA: CLIENT: ☐ Hand Delivered ☐ Overnight				ALLIANCE: ☐ Picked Up ☐ Overnight		Shipment Complete ☐ YES ☐ NO											
WHITE - ALLIANCE COPY FOR RETURN TO CLIENT YELLOW - ALLIANCE COPY PINK - SAMPLER COPY																									

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q2033	ENTA05	Order Date : 5/13/2025 4:02:00 PM	Project Mgr :
Client Name : ENTACT		Project Name : 540 Degraw St, Brooklyn, N	Report Type : Level 1
Client Contact : Jarod Stanfield		Receive DateTime : 5/13/2025 12:00:00 AM	EDD Type : Excel NJ
Invoice Name : ENTACT		Purchase Order : 18:00	Hard Copy Date :
Invoice Contact : Jarod Stanfield			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q2033-01	TW-WTS-08	Water	05/12/2025	15:00	VOCMS Group4		8260-Low	5 Bus. Days	

Relinquished By :

Date / Time :


5/14/25 0940

Received By :

Date / Time :


05/14/25 9:40 

Storage Area : VOA Refridgerator Room