

## NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

## GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJE	CT NUMBER:	bp052025			
SequenceID :	bp052025		NA	NO	YES
1. Chromatograms La	beled/Compounds Ide	entified. (Field samples and Method Blanks)			✓
2. GC/MS Tuning Specifications. DFTPP Meet Criteria Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)					<b>_</b>
3. GC/MS Tuning Fre series	equency - Performed e	very 24 hours for 600 series and 12 hours for 8000			_ <b>√</b>
<ol> <li>GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series</li> </ol>					✓
5. GC/MS Calibration Met:					✓
a. Initial calibratio If not met, list those c		ecoveries which fall outside the acceptable range.			<u> </u>
b. Continuous Calibration(CCC) Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.					
	-	n the SSTDCCC (BP024703.D and BP024720.D) if a h a properly passing CCC. Certain compounds are bia		of these	
6. Blank Contaminati		✓			
a. B/N Fraction					

d. Acid Fraction

7. Surrogate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges.	<u>✓</u>	
a. B/N Fraction		
d. Acid Fraction		
<ul> <li>8. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria</li> <li>If not met, list those compounds and their recoveries which fall outside the acceptable range.</li> <li>a. B/N Fraction Only the base surrogates are biased low in the Q2071-13MS/MSD proving matrix interference. Hence no corrective action</li> </ul>	✓ _	
<ul> <li>a. B/N Fraction</li> <li>d. Acid Fraction</li> </ul>	<u>in is required.</u>	
9. Internal Standard Area/Retention Time Shift Meet Criteria Comments:		✓
10. Extraction Holding Time Met If not met, list number of days exceeded for each sample:		✓
11. Analysis Holding Time Met     If not met, list number of days exceeded for each sample:	<u>✓ _</u>	

# ADDITIONAL COMMENTS:

1,4-Dioxane is marginally biased low in the PB168048BS/BSD but there is no hit in any of the associated samples. The data will be used for hardcopies.



### NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

## GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER:

SequenceID :

bp052025

bp052025

NA NO YES